

EXHIBIT A

The diagram illustrates a dialysis machine system. A central dialyzer (20) is connected to a patient (12) via a circuit. The circuit includes a fluid pressure sensor (68), an air in line detector (60), and various valves (V1-V6). A supply bag (14) provides fluid to the system. A vacuum pump (46) maintains negative pressure. A control unit (30) with a display (40) and video controller (38) manages the system. Other components include a linear encoder piston pump (44), a diaphragm piston pump (22), and various sensors like temperature (T1-T4), pressure (P1-P2), and flow (F1-F2).

US 7,815,595 B2

Page 2

U.S. PATENT DOCUMENTS					
3,626,670 A	12/1971	Pecker	4,479,760 A	10/1984	Bilstad et al.
3,656,873 A	4/1972	Schiff	4,479,761 A	10/1984	Bilstad et al.
3,667,612 A	6/1972	Leonard	4,479,762 A	10/1984	Bilstad et al.
3,689,204 A	9/1972	Prisk	4,488,961 A	12/1984	Spencer
3,703,959 A	11/1972	Raymond	4,498,900 A	2/1985	Buoncristiani
3,707,967 A	1/1973	Kitrilakis et al.	4,504,038 A	3/1985	King
3,709,222 A	1/1973	DeVries	4,530,759 A	7/1985	Schal
3,727,612 A	4/1973	Sayers et al.	4,532,414 A	7/1985	Shah et al.
3,730,183 A	5/1973	Goldsmith et al.	4,552,552 A	11/1985	Polaschegg et al.
3,792,643 A	2/1974	Scheafer	4,559,036 A	12/1985	Wunsch
3,799,873 A	3/1974	Brown	4,559,044 A	12/1985	Robinson et al.
3,809,241 A	5/1974	Alvine	4,560,472 A	12/1985	Granmzow et al.
3,820,661 A	6/1974	Pages	4,574,876 A	3/1986	Aid
3,823,724 A	7/1974	Davis	4,579,642 A	4/1986	Niiyama et al.
3,825,493 A	7/1974	Brown et al.	4,584,061 A	4/1986	Shelton
3,839,200 A	10/1974	Gigou	4,586,920 A	5/1986	Peabody
3,882,899 A	5/1975	Ginsberg	4,613,327 A	9/1986	Tegrarian et al.
3,884,808 A	5/1975	Scott	4,618,343 A	10/1986	Polaschegg
3,902,490 A	9/1975	Jacobsen	RE32,303 E	12/1986	Lasker et al.
3,903,478 A	9/1975	Stuart	4,634,430 A	1/1987	Polaschegg
3,926,797 A	12/1975	Gigou	4,639,245 A	1/1987	Pastrone et al.
3,929,314 A	12/1975	Stratynski	4,642,098 A	2/1987	Lundquist
3,939,069 A	2/1976	Granger	4,648,810 A	3/1987	Schippers et al.
3,955,901 A	5/1976	Hamilton	4,655,753 A	4/1987	Bellotti et al.
3,967,809 A	7/1976	Skantar	4,657,490 A	4/1987	Abbott
3,976,574 A	8/1976	White	4,661,246 A	4/1987	Ash
3,979,284 A	9/1976	Granger	4,663,006 A	5/1987	Yao et al.
4,071,040 A	1/1978	Moriarty	4,678,460 A	7/1987	Rosner
4,079,007 A	3/1978	Hutchisson	4,694,848 A	9/1987	Jorgensen et al.
4,083,777 A	4/1978	Hutchisson	4,696,671 A	9/1987	Epstein et al.
4,086,653 A	4/1978	Gernes	4,703,773 A	11/1987	Hansen et al.
4,096,859 A	6/1978	Agarwal et al.	4,717,117 A	1/1988	Cook
4,118,314 A	10/1978	Yoshida	4,718,890 A	1/1988	Peabody
4,126,132 A	11/1978	Portner et al.	4,735,609 A	4/1988	Comeau et al.
4,142,524 A	3/1979	Jassawalla et al.	4,742,870 A	5/1988	Darone et al.
4,158,530 A	6/1979	Bernstein	4,747,822 A	5/1988	Peabody
4,180,460 A	12/1979	Calari	4,747,828 A	5/1988	Tseo
4,181,245 A	1/1980	Garrett et al.	4,765,907 A	8/1988	Scott
4,190,047 A	2/1980	Jacobsen et al.	4,769,151 A	9/1988	Shouldice
4,191,646 A	3/1980	Larsson et al.	4,778,356 A	10/1988	Hicks
4,192,748 A	3/1980	Hyden	4,778,451 A	10/1988	Kamen
4,194,536 A	3/1980	Stine et al.	4,784,576 A	11/1988	Bloom et al.
4,199,307 A	4/1980	Jassawalla	4,794,942 A	1/1989	Yasuda et al.
4,212,738 A	7/1980	Henne	4,798,580 A	1/1989	DeMeo et al.
4,235,231 A	11/1980	Schindler et al.	4,804,360 A	2/1989	Kamen
4,236,880 A	12/1980	Archibald	4,804,474 A	2/1989	Blum
4,240,408 A	12/1980	Schael	4,808,161 A	2/1989	Kamen
4,252,115 A	2/1981	Schael	4,816,019 A	3/1989	Kamen
4,252,651 A	2/1981	Soderstrom	4,818,186 A	4/1989	Pastrone et al.
4,265,601 A	5/1981	Mandoian	4,818,190 A	4/1989	Pelmulder et al.
4,269,708 A	5/1981	Bonomini et al.	4,823,552 A	4/1989	Ezell et al.
4,273,121 A	6/1981	Jassawalla	4,826,482 A	5/1989	Kamen
4,276,175 A	6/1981	Bower	4,828,545 A	5/1989	Epstein et al.
4,277,226 A	7/1981	Archibald	4,830,586 A	5/1989	Herter et al.
4,293,762 A	10/1981	Ogawa	4,842,584 A	6/1989	Pastrone
4,303,376 A *	12/1981	Siekmann 417/360	4,844,074 A	7/1989	Kurucz
4,310,141 A	1/1982	Tamura	4,847,470 A	7/1989	Bakke
4,338,190 A	7/1982	Kraus et al.	4,848,722 A	7/1989	Webster
4,351,333 A	9/1982	Lazarus et al.	4,850,805 A	7/1989	Madsen et al.
4,375,346 A	3/1983	Kraus et al.	4,852,851 A	8/1989	Webster
4,381,003 A	4/1983	Buoncristiani	4,858,883 A	8/1989	Webster
4,381,005 A	4/1983	Bujan	4,859,319 A	8/1989	Borsari
4,391,600 A	7/1983	Archibald	4,865,584 A	9/1989	Epstein et al.
4,396,382 A	8/1983	Goldhaber	4,872,813 A	10/1989	Gorton et al.
4,413,988 A	11/1983	Handt et al.	4,886,432 A	12/1989	Kimberlin
4,430,048 A	2/1984	Fritsch	4,889,812 A	12/1989	Guinn et al.
4,456,218 A	6/1984	Kawabata et al.	4,906,816 A	3/1990	Van Leerdam
4,464,563 A	8/1984	Jewett	4,918,019 A	4/1990	Guinn
4,468,222 A	8/1984	Lundquist	4,923,598 A	5/1990	Schal
4,475,900 A	10/1984	Popovich et al.	4,925,152 A	5/1990	Huber
4,476,005 A	10/1984	Tokinaga et al.	4,927,411 A	5/1990	Pastrone et al.
			4,941,519 A	7/1990	Sestak et al.
			4,942,735 A	7/1990	Mushika et al.

US 7,815,595 B2

Page 3

4,956,996 A	9/1990	Morris	5,556,263 A	9/1996	Jacobsen et al.
4,976,162 A	12/1990	Kamen	5,563,584 A	10/1996	Rader et al.
4,976,683 A	12/1990	Gauthier et al.	5,580,460 A	12/1996	Polaschegg
4,990,054 A	2/1991	Janocko	5,582,732 A	12/1996	Mao et al.
5,000,664 A	3/1991	Lawless et al.	5,586,868 A	12/1996	Lawless et al.
5,002,471 A	3/1991	Perlov	5,588,816 A	12/1996	Abbott et al.
5,004,459 A	4/1991	Peabody et al.	5,591,344 A	1/1997	Kenley et al.
5,006,050 A	4/1991	Cooke et al.	5,601,420 A	2/1997	Warner et al.
5,006,997 A	4/1991	Reich	5,603,354 A	2/1997	Jacobsen et al.
5,019,140 A	5/1991	Bowser et al.	5,609,572 A *	3/1997	Lang 604/22
5,037,385 A	8/1991	O'Byrne	5,620,312 A	4/1997	Hyman et al.
5,043,201 A	8/1991	Cote	5,628,908 A	5/1997	Kamen et al.
5,053,003 A	10/1991	Dadson et al.	5,630,935 A	5/1997	Treu
5,062,774 A	11/1991	Kramer et al.	5,632,606 A	5/1997	Jacobsen et al.
5,073,167 A	12/1991	Carr et al.	5,634,896 A	6/1997	Bryant et al.
5,088,515 A	2/1992	Kamen	5,636,653 A	6/1997	Titus
5,098,262 A	3/1992	Wecker et al.	5,641,405 A	6/1997	Keshaviah et al.
5,106,366 A	4/1992	Steppe	5,643,201 A	7/1997	Peabody et al.
5,108,367 A	4/1992	Epstein et al.	5,645,734 A	7/1997	Kenley et al.
5,116,203 A	5/1992	Natwick et al.	5,669,764 A	9/1997	Behringer et al.
5,125,069 A	6/1992	O'Boyle	5,676,530 A	10/1997	Nazarifar
5,127,907 A	7/1992	Courtne et al.	5,683,381 A	11/1997	Carr et al.
5,135,485 A	8/1992	Cohen et al.	5,685,988 A	11/1997	Malchesky
5,141,492 A	8/1992	Dadson et al.	5,690,160 A	11/1997	Sutton et al.
5,141,493 A	8/1992	Jacobsen et al.	5,690,614 A	11/1997	Carr et al.
5,163,900 A	11/1992	Wortrich	5,698,083 A	12/1997	Glass
5,178,182 A	1/1993	Kamen	5,718,569 A	2/1998	Holst
5,180,896 A	1/1993	Gibby et al.	5,720,313 A	2/1998	Grobbe
5,194,157 A	3/1993	Ghezzi et al.	5,722,947 A	3/1998	Jeppsson et al.
5,215,312 A	6/1993	Knappe et al.	5,724,478 A	3/1998	Thweatt
5,240,614 A	8/1993	Ofsthun et al.	5,729,653 A	3/1998	Magliochetti et al.
5,245,693 A	9/1993	Ford et al.	5,758,563 A	6/1998	Robinson
5,247,434 A	9/1993	Peterson et al.	5,790,752 A	8/1998	Anglin et al.
5,252,044 A	10/1993	Raines et al.	5,792,367 A	8/1998	Mattisson et al.
5,277,820 A	1/1994	Ash	5,807,075 A	9/1998	Jacobsen et al.
5,284,470 A	2/1994	Beltz	5,814,004 A	9/1998	Tamari
5,292,306 A	3/1994	Wynkoop et al.	5,816,779 A	10/1998	Lawless et al.
5,295,964 A	3/1994	Gauthier	5,836,908 A	11/1998	Beden et al.
5,302,093 A	4/1994	Owens et al.	5,843,474 A	12/1998	Williams
5,316,452 A	5/1994	Bogen et al.	5,858,186 A	1/1999	Glass
5,324,422 A	6/1994	Colleran et al.	5,858,238 A	1/1999	McRea et al.
5,332,372 A	7/1994	Reynolds	5,871,566 A	2/1999	Rutz
5,334,139 A	8/1994	Jeppsson et al.	5,875,282 A	2/1999	Jordan et al.
5,338,293 A	8/1994	Jeppsson et al.	5,919,369 A	7/1999	Ash
5,344,292 A	9/1994	Rabenau et al.	5,921,951 A	7/1999	Morris
5,350,357 A *	9/1994	Kamen et al. 604/29	5,924,975 A	7/1999	Goldowsky
5,352,364 A	10/1994	Kruger et al.	5,928,196 A	7/1999	Johnson et al.
5,364,342 A	11/1994	Beuchat et al.	5,931,647 A	8/1999	Jacobsen et al.
5,370,674 A	12/1994	Farrell	5,938,634 A	8/1999	Packard
5,378,126 A	1/1995	Abrahamson et al.	5,944,495 A	8/1999	Jacobsen et al.
5,381,510 A	1/1995	Ford et al.	5,944,684 A	8/1999	Roberts et al.
5,385,540 A	1/1995	Abbott et al.	5,945,831 A	8/1999	Sargent et al.
5,389,243 A	2/1995	Kaplan	5,951,863 A	9/1999	Kruger et al.
5,408,576 A	4/1995	Bishop	5,960,160 A	9/1999	Clark et al.
5,409,355 A	4/1995	Brooke	5,965,433 A	10/1999	Gardetto et al.
5,415,528 A	5/1995	Ogden et al.	5,980,481 A	11/1999	Gorsuch
5,420,962 A	5/1995	Bakke	5,989,238 A	11/1999	Ginsburg
5,421,208 A	6/1995	Packard et al.	5,989,423 A	11/1999	Kamen et al.
5,421,823 A	6/1995	Kamen et al.	6,007,310 A	12/1999	Jacobsen et al.
5,429,485 A	7/1995	Dodge	6,017,194 A	1/2000	North, Jr.
5,431,626 A	7/1995	Bryant et al.	6,017,198 A	1/2000	Traylor et al.
5,438,510 A	8/1995	Bryant et al.	6,030,359 A	2/2000	Nowosielski
5,439,587 A	8/1995	Stankowski et al.	6,036,668 A	3/2000	Mathis
5,441,606 A	8/1995	Schlesinger et al.	6,041,801 A	3/2000	Gray et al.
5,458,468 A	10/1995	Ye et al.	6,065,941 A	5/2000	Gray et al.
5,474,683 A	12/1995	Bryant et al.	6,069,343 A	5/2000	Kolowich
5,476,368 A	12/1995	Rabenau et al.	6,109,254 A	8/2000	Reinke et al.
5,482,440 A	1/1996	Dennehey et al.	6,110,617 A	8/2000	Feres
5,487,649 A	1/1996	Dorsey, III et al.	6,122,972 A	9/2000	Crider
5,522,769 A	6/1996	DeGuseppi	6,126,403 A	10/2000	Yamada
5,536,412 A	7/1996	Ash	6,129,699 A	10/2000	Haight et al.
5,542,919 A	8/1996	Simon et al.	6,133,547 A	10/2000	Maynard
5,554,013 A	9/1996	Owens et al.	6,139,528 A	10/2000	Kistner et al.

US 7,815,595 B2

Page 4

6,142,974 A	11/2000	Kistner et al.	GB	2124511 A	2/1984
6,146,359 A	11/2000	Carr et al.	JP	6-106033	4/1994
6,165,154 A	12/2000	Gray et al.	SE	331736	1/1971
6,175,688 B1	1/2001	Cassidy et al.	WO	WO 84/02473	7/1984
6,208,107 B1	3/2001	Maske et al.	WO	85/01879	5/1985
6,223,130 B1	4/2001	Gray et al.	WO	WO 85/04813	11/1985
6,228,047 B1	5/2001	Dadson	WO	WO 86/01115	2/1986
6,229,957 B1	5/2001	Baker	WO	87/05223	9/1987
6,231,320 B1	5/2001	Lawless et al.	WO	89/01795	3/1989
6,234,992 B1	5/2001	Haight et al.	WO	90/13795	11/1990
6,236,809 B1	5/2001	Cassidy et al.	WO	WO 91/02484	3/1991
6,246,831 B1	6/2001	Seitz et al.	WO	WO 92/15349	9/1992
6,248,093 B1	6/2001	Moberg	WO	93/00938	1/1993
6,254,567 B1	7/2001	Treu et al.	WO	93/01845	2/1993
6,257,265 B1	7/2001	Brunner et al.	WO	WO 93/01845	2/1993
6,259,074 B1	7/2001	Brunner et al.	WO	94/20158	9/1994
6,261,261 B1	7/2001	Gordon	WO	95/02559	1/1995
6,270,673 B1	8/2001	Belt et al.	WO	WO 95/35124	12/1995
6,278,084 B1	8/2001	Maynard	WO	98/17333	4/1998
6,283,719 B1	9/2001	Frantz et al.	WO	WO 99/06082	2/1999
6,293,921 B1	9/2001	Shimoto et al.	WO	99/37335	7/1999
6,293,926 B1	9/2001	Sorensen et al.	WO	00/20050	4/2000
6,302,653 B1	10/2001	Bryant et al.	WO	00/31553	6/2000
6,343,614 B1	2/2002	Gray et al.	WO	00/44018	7/2000
6,364,857 B1	4/2002	Gray et al.	WO	00/57928	10/2000
6,416,293 B1	7/2002	Bouchard et al.	WO	WO 01/91829	12/2001
6,491,656 B1	12/2002	Morris			
6,595,948 B2	7/2003	Suzuki et al.			
6,743,201 B1	6/2004	Dönig			
6,764,761 B2	7/2004	Eu et al.			
6,814,547 B2	11/2004	Childers et al.			
6,869,538 B2	3/2005	Yu et al.			
6,929,751 B2	8/2005	Bowman, Jr. et al.			
6,939,111 B2	9/2005	Huitt et al.			
6,953,323 B2	10/2005	Childers et al.			
7,033,539 B2	4/2006	Krensky et al.			
7,087,036 B2	8/2006	Busby et al.			
7,115,228 B2	10/2006	Lundtveit et al.			
7,153,286 B2	12/2006	Busby et al.			
7,175,606 B2	2/2007	Bowman, Jr. et al.			
2001/0018937 A1	9/2001	Nemoto			
2002/0045851 A1	4/2002	Suzuki et al.			
2002/0077598 A1	6/2002	Yap et al.			
2002/0081109 A1	6/2002	Mitsunaga et al.			

FOREIGN PATENT DOCUMENTS

EP	0028371	5/1981
EP	0 033 096 A2	8/1981
EP	0 052 004 A1	5/1982
EP	0 097 432 A1	1/1984
EP	0 157 024 B1	10/1985
EP	0 206 195 A2	12/1986
EP	0204260	12/1986
EP	0 248 632 B1	12/1987
EP	0 258 424 B1	3/1988
EP	0319272	6/1989
EP	0 402 505 B1	12/1990
EP	0402505	12/1990
EP	0011935	5/1991
EP	0248632	4/1992
EP	0 660 725 B1	7/1995
EP	0 847 769 A	6/1998
EP	0 856 320 A	8/1998
EP	0 856 321 A1	8/1998
EP	0 856 322 A1	8/1998
EP	0 928 615 A1	7/1999
EP	1110564 A2	6/2001
EP	1110565 A2	6/2001
FR	2371931	6/1978
FR	2440740	6/1980
GB	1326236	8/1973
GB	2122509 A	1/1984

OTHER PUBLICATIONS

U.S. Appl. No. 10/155,752, filed May 24, 2002, Busby et al.

U.S. Appl. No. 10/155,753, filed May 24, 2002, Hopping.

U.S. Appl. No. 10/155,787, filed May 24, 2002, Busby et al.

U.S. Appl. No. 11/187,515, filed Jul. 22, 2005, Childers et al.

U.S. Appl. No. 11/614,850, filed Dec. 21, 2006.

Opening Expert Witness Report of Dr. Juan Santiago Regarding Anticipation and Obviousness of the Claims of U.S. Patents Nos. 6,503,062 and 6,808,369 in view of the Prior Art and based on the Indefiniteness, Lack of Enablement, and Lack of Written Description of Certain Claims of U.S. Patent Nos. 6,503,062 and 6,808,369, Apr. 24, 2009.

Opening Expert Witness Report of William K. Durfee Regarding whether Certain Claims of U.S. Patent No. 5,324,422, U.S. Patent No. 5,421,823, U.S. Patent No. 5,431,626 and U.S. Patent No. 5,438,510 were Ready for Patenting, Apr. 24, 2009.

Expert Witness Report of Fred K. Forster: Analysis of Obviousness of Certain Asserted Claims of U.S. Patent Nos. 5,431,626; 5,324,422; and 5,438,510, Apr. 24, 2009.

Expert Witness Report of Ronald J. Adrian Regarding Lack of Written Description, Lack of Enablement, and Indefiniteness of the Asserted Claim (Claim 12) of U.S. Patent No. 6,814,547, Apr. 24, 2009.

Expert Report on Development of the PD700 and Motivation to Combine the PD700 and U.S. Patent No. 5,088,515, Sven Olofsson, Apr. 24, 2009.

Expert Witness Report of Juan G. Santiago Regarding Lack of Written Description, Non-Enablement, and Indefiniteness of the Asserted Claims of U.S. Patent Nos. 5,421,823; 5,324,422; 5,438,510; and 5,431,626, Apr. 24, 2009.

Opening Expert Witness Report of Dr. Martin Roberts Regarding a History of Peritoneal Dialysis and the Obviousness and Consequent Invalidity of the Asserted Claims of U.S. Patent No. 5,421,823, Apr. 24, 2009.

Opening Expert Witness Report of Dr. Darrell Long Regarding Technical Features of the High Flow Peritoneal Dialysis and Personal Cycler Machines, Apr. 24, 2009.

Defendants' Preliminary Invalidity Contentions for U.S. Patent No. 5,421,823, *Baxter Healthcare Corporation v. Fresenius Medical Care Holdings*, Case No. C 07-01359 PJH (JL), filed Aug. 24, 2007.

Defendants' Supplemental Invalidity Contentions for U.S. Patent No. 5,421,823, *Baxter Healthcare Corporation v. Fresenius Medical Care Holdings*, Case No. C 07-01359 PJH (JL), filed Jan. 31, 2008.

US 7,815,595 B2

Page 5

- Defendants' Preliminary Invalidity Contentions for U.S. Patent No. 6,503,062, *Baxter Healthcare Corporation v. Fresenius Medical Care Holdings*, Case No. C 07-01359 PJH (JL), filed Jan. 31, 2008.
- Defendants' Preliminary Invalidity Contentions for U.S. Patent No. 6,808,369, *Baxter Healthcare Corporation v. Fresenius Medical Care Holdings*, Case No. C 07-01359 PJH (JL), filed Jan. 31, 2008.
- Defendants' Preliminary Invalidity Contentions for U.S. Patent No. 5,324,422, *Baxter Healthcare Corporation v. Fresenius Medical Care Holdings*, Case No. C 07-01359 PJH (JL), filed Jan. 31, 2008.
- Defendants' Preliminary Invalidity Contentions for U.S. Patent No. 5,438,510, *Baxter Healthcare Corporation v. Fresenius Medical Care Holdings*, Case No. C 07-01359 PJH (JL), filed Jan. 31, 2008.
- Defendants' Preliminary Invalidity Contentions for U.S. Patent No. 6,814,547, *Baxter Healthcare Corporation v. Fresenius Medical Care Holdings*, Case No. C 07-01359 PJH (JL), filed Jan. 31, 2008.
- Defendants' Preliminary Invalidity Contentions for U.S. Patent No. 5,431,626, *Baxter Healthcare Corporation v. Fresenius Medical Care Holdings*, Case No. C 07-01359 PJH (JL), filed Jan. 31, 2008.
- Defendants' Preliminary Invalidity Contentions for U.S. Patent No. 6,929,751, *Baxter Healthcare Corporation v. Fresenius Medical Care Holdings*, Case No. C 07-01359 PJH (JL), filed Jan. 31, 2008.
- Defendants' Preliminary Invalidity Contentions for U.S. Patent No. 7,083,719, *Baxter Healthcare Corporation v. Fresenius Medical Care Holdings*, Case No. C 07-01359 PJH (JL), filed Jan. 31, 2008.
- Defendants' Final Invalidity Contentions for U.S. Patent No. 6,814,547, *Baxter Healthcare Corporation v. Fresenius Medical Care Holdings*, Case No. C 07-01359 PJH (JL), filed Apr. 1, 2009.
- Fresenius 90/2 Peritoneal Therapy Cyclers (on information and belief, on sale in United States by 1991).
- Blumenkrantz et al., Applications of the Redy Sorbent System to Hemodialysis and Peritoneal Dialysis, *Artificial Organs*, vol. 3, No. 3 (Aug. 1979).
- Blumenkrantz et al., Development of a Sorbent Peritoneal Dialysate Regeneration System—A Progress Report, *European Dialysis and Transplant Association* 1978.
- Blumenkrantz and Roberts, Progress in Peritoneal Dialysis: a Historical Prospective, *Contributions to Nephrology*, vol. 17, pp. 101-110 (1979).
- Diaz-Buxo, CCPD is even better than CAPD, *Kidney International*, vol. 28, Suppl. 17, pp. S-26-S-28 (1985).
- Diaz-Buxo, CCPD Technique and Current Clinical Experience (1982).
- Diaz-Buxo, et al., Continuous Cyclic Peritoneal Dialysis: A Preliminary Report, *Artificial Organs*, vol. 5, No. 2, pp. 157-161 (May 1981).
- Diaz-Buxo, Current Status of Continuous Cyclic Peritoneal Dialysis (CCPD), *Peritoneal Dialysis International*, vol. 9, pp. 9-14 (1989).
- Diaz-Buxo, Issues in Nephrology: Continuous Cyclic Peritoneal Dialysis, *NAPHT News*, Feb. 1983, pp. 12-14.
- Diaz-Buxo, Peritoneal Dialysis Reverse Osmosis Machines and Cyclers, *Dialysis Therapy*, pp. 41-48 (1986).
- Drukker et al., Replacement of Renal Function by Dialysis, 2nd Ed., Ch. 21, 1983.
- Lewin and Maxwell, Sorbent-Based Regenerating Peritoneal Dialysis, Sorbents and Their Clinical Applications, pp. 353-374 (1980).
- Lewin et al., Sorbent for Application in the Treatment of ESRD Patients, Annual Progress Report re Contract #N01-AM-9-2215, submitted Jun. 22, 1982.
- Ratnu, et al., A New Technique—Semicontinuous Rapid Flow, High Volume Exchange—For Effective Peritoneal Dialysis in Shorter Periods, *Nephron*, vol. 31, pp. 159-163 (1982).
- Twardowski, Peritoneal Dialysis: Current Technology and techniques, *Postgraduate Medicine*, vol. 85, No. 5 (Apr. 1989).
- Product Evaluation Reports: Peritoneal Dialysis Machine "Pac-X," *Hospital Materials Management*, vol. 12, No. 11, p. 16 (Nov. 1987).
- Brochure entitled, AP Hauni: Automatisches Peritonealdialyse-Great (1970).
- Peritoneal Dialyser PD700 Service Manual, Jun. 1977.
- PD700 Peritoneal Dialyser Users Hand-book, Dec. 1977.
- Brochure entitled, Peritoneal Dialyser PD700, May 1979.
- Brochure entitled, For Volume Measurement, Temperature Control and Cycling of Dialysing Fluid, Peritoneal Dialyser PD700, 1970.
- Bergstrom et al., An Automated Apparatus for Peritoneal Dialysis with Volumetric Fluid Balance Measurement, reprinted from *Dialysis & Transplantation*, Jun./Jul. 1976.
- U. Callsen, Peritoneal-Dialysator PD700, *Prakt. Anasth.* 9 (1974).
- Piazolo et al., Erfahrungen mit einem neuen vollautomatischen Peritoneal-dialysegerat, *Munchener Medizinische Wochenschrift*, 1972.
- Technical Note, PD700 Peritoneal Dialyser, Jan. 29, 1979.
- Elsevier Science Ltd., Air-Operated Diaphragm Pumps, *World Pumps*, Jan. 1996, at 38.
- Bran & Luebbe GmbH, Diaphragm Metering Pumps, *Chem. Eng'g Progress*, Apr. 1987, at 18-24.
- W.M. Phillips, J.A. Brighton & W.S. Pierce, Artificial Heart Evaluation Using Flow Visualization Techniques, published in *Transactions: American Society for Artificial Internal Organs*, vol. XVIII (1972).
- J.A. Brighton, W.S. Pierce, D.Landis & G. Rosenberg, Measuring Cardiac Output of Pneumatically Driven Total Artificial Hearts, published in 30th Anniversary Conference on Engineering in Medicine and Biology: Proceedings, vol. 19 (Nov. 5-9, 1977).
- Operator's Instructions, Fresenius 90/2 Peritoneal Therapy Cyclers (Rev. C. copyright 1991-2000).
- Memorandum of Donald X. Vaccarino entitled 90/2 History File (1991-1992).
- Document entitled 90/2 Cyclers Software, Version 3.96 (Jan. 24, 1992).
- Software Change Requests (Jul. 8, 1991-Oct. 3, 1992).
- Brochure entitled Fresenius Delivers 90/2 Peritoneal Therapy Cyclers (Apr. 2001).
- 90/2 Cyclers Parts List (Nov. 6, 1997).
- 90/2 Brochure (Jul. 1993).
- 90/2 Brochure (Apr. 2001).
- Specification entitled Inpersol Cyclers 3000 Operating Manual, List No. 21952-04, dated 1990.
- Training aid entitled Learning to Use the Inpersol Cyclers 3000, dated Jul. 1991.
- Fresenius USA/Delmed 90/2 Peritoneal Dialysis System Operators Manual, dated Feb. 6, 1991.
- Fresenius 90/2 Peritoneal Therapy Cyclers Operator's Instructions, dated 2000.
- Defendants' Preliminary Invalidity Contentions for U.S. Patent No. 5,421,823 Aug. 24, 2007.
- U.S. Appl. No. 11/617,527, filed Dec. 28, 2006, Childers et al.
- U.S. Appl. No. 11/617,543, filed Dec. 28, 2006, Childers et al.
- U.S. Appl. No. 11/773,148, filed Jul. 3, 2001, Childers et al.

* cited by examiner





FIG. 2






FIG. 3A

FIG. 3B





FIG. 4A



U.S. Patent

Oct. 19, 2010

Sheet 6 of 37

US 7,815,595 B2

FIG. 4B




FIG. 5





FIG. 6






FIG. 7




FIG. 8


U.S. Patent

Oct. 19, 2010

Sheet 11 of 37

US 7,815,595 B2

FIG. 9




U.S. Patent

Oct. 19, 2010

Sheet 12 of 37

US 7,815,595 B2



U.S. Patent

Oct. 19, 2010

Sheet 13 of 37

US 7,815,595 B2

FIG. 11




FIG.12




FIG. 13




FIG. 14







FIG. 15








FIG.17A




FIG.17B

FIG. 18




FIG. 19




FIG. 20




FIG. 21




FIG. 22




FIG. 23




FIG. 24

500
↓

	Parameter	Effect
486	Begin Stroke Acceleration	Big impact on overshoot but undesirable from standpoint of stroke time.
488	Proximity Threshold	Good method but may need to be set far below pressure set point.
490	DP/dt	Look at rate of change of pressure and limit to reset to the new velocity setpoint.
492	Max. Travel Velocity	May or may not effect result depending on the initial setting.
494	Conversion to Pressure Deceleration	Potential large impact
496	Kp	Will affect in significant ways the overshoot once the proximity threshold is reached, however, it also affects the entire control to pressure PID control period of the pulse.
498	Kd	Will affect in small ways the overshoot once the proximity threshold is reached, however, it also affects the entire control to pressure and PID control period of the pulse.
502	Ki	Will affect in small ways the overshoot once the proximity threshold is reached, however, it also affects the entire control to pressure and PID control period of the pulse.




FIG. 26




FIG. 27





FIG. 28






FIG. 30A




FIG. 30B




FIG. 30C




FIG. 30D




FIG. 30E




FIG. 30F




FIG. 30G




FIG. 30H




FIG. 30I




FIG. 30J




FIG. 30K




FIG. 30L





FIG. 30M



US 7,815,595 B2

1

AUTOMATED DIALYSIS PUMPING SYSTEM**PRIORITY**

This application claims priority to and the benefit as a continuation application of U.S. patent application "Automated Dialysis System", Ser. No. 10/155,603, filed May 24, 2002.

BACKGROUND

The present disclosure generally relates to dialysis systems. More specifically, the present disclosure relates to automated peritoneal dialysis systems. The present disclosure also relates to methods of performing automated peritoneal dialysis and devices for performing same.

Due to disease, insult or other causes, a person's renal system can fail. In renal failure of any cause, there are several physiological derangements. The balance of water, minerals and the excretion of daily metabolic load is no longer possible in renal failure. During renal failure, toxic end products of nitrogen metabolism (urea, creatinine, uric acid, and others) can accumulate in blood and tissues.

Kidney failure and reduced kidney function have been treated with dialysis. Dialysis removes waste, toxins and excess water from the body that would otherwise have been removed by normal functioning kidneys. Dialysis treatment for replacement of kidney functions is critical to many people because the treatment is life saving. One who has failed kidneys could not continue to live without replacing at least the filtration functions of the kidneys.

Hemodialysis and peritoneal dialysis are two types of dialysis therapies commonly used to treat loss of kidney function. Hemodialysis treatment utilizes the patient's blood to remove waste, toxins and excess water from the patient. The patient is connected to a hemodialysis machine and the patient's blood is pumped through the machine. Catheters are inserted into the patient's veins and arteries to connect the blood flow to and from the hemodialysis machine. As blood passes through a dialyzer in the hemodialysis machine, the dialyzer removes the waste, toxins and excess water from the patient's blood and returns the blood back to the patient. A large amount of dialysate, for example about 120 liters, is used to dialyze the blood during a single hemodialysis therapy. The spent dialysate is then discarded. Hemodialysis treatment lasts several hours and is generally performed in a treatment center about three or four times per week.

Peritoneal dialysis utilizes a dialysis solution or "dialysate", which is infused into a patient's peritoneal cavity through a catheter implanted in the cavity. The dialysate contacts the patient's peritoneal membrane in the peritoneal cavity. Waste, toxins and excess water pass from the patient's bloodstream through the peritoneal membrane and into the dialysate. The transfer of waste, toxins, and water from the bloodstream into the dialysate occurs due to diffusion and osmosis, i.e., an osmotic gradient occurs across the membrane. The spent dialysate drains from the patient's peritoneal cavity and removes the waste, toxins and excess water from the patient. This cycle is repeated.

There are various types of peritoneal dialysis therapies, including continuous ambulatory peritoneal dialysis ("CAPD"), automated peritoneal dialysis and continuous flow peritoneal dialysis. CAPD is a manual dialysis treatment, in which the patient connects an implanted catheter to a drain and allows a spent dialysate fluid to drain from the peritoneal cavity. The patient then connects the catheter to a bag of fresh dialysate and manually infuses fresh dialysate

2

through the catheter and into the patient's peritoneal cavity. The patient disconnects the catheter from the fresh dialysate bag and allows the dialysate to dwell within the cavity to transfer waste, toxins and excess water from the patient's bloodstream to the dialysate solution. After a dwell period, the patient repeats the manual dialysis procedure.

In CAPD the patient performs several drain, fill, and dwell cycles during the day, for example, about four times per day. Each treatment cycle typically takes about an hour. Manual peritoneal dialysis performed by the patient requires a significant amount of time and effort from the patient. This inconvenient procedure leaves ample room for improvement and therapy enhancements to improve patient quality of life.

Automated peritoneal dialysis ("APD") is similar to CAPD in that the dialysis treatment includes a drain, fill, and dwell cycle. APD machines, however, automatically perform three to four cycles of peritoneal dialysis treatment, typically overnight while the patient sleeps. The APD machines fluidly connect to an implanted catheter. The APD machines also fluidly connect to a source or bag of fresh dialysate and to a fluid drain.

The APD machines pump fresh dialysate from the dialysate source, through the catheter, into the patient's peritoneal cavity and allow the dialysate to dwell within the cavity so that the transfer of waste, toxins and excess water from the patient's bloodstream to the dialysate solution can take place. The APD machines then pump spent dialysate from the peritoneal cavity, through the catheter, to the drain. APD machines are typically computer controlled so that the dialysis treatment occurs automatically when the patient is connected to the dialysis machine, for example, when the patient sleeps. That is, the APD systems automatically and sequentially pump fluid into the peritoneal cavity, allow for a dwell, pump fluid out of the peritoneal cavity and repeat the procedure.

As with the manual process, several drain, fill, and dwell cycles will occur during APD. A "last fill" is typically used at the end of APD, which remains in the peritoneal cavity of the patient when the patient disconnects from the dialysis machine for the day. APD frees the patient from having to manually performing the drain, dwell, and fill steps.

However, continuing needs exist to provide improved APD systems. For example, needs exist to provide simplified APD systems that are easier for patients to use and operate. Further, needs exist to provide lower cost APD systems and APD systems which are less costly to operate. Particularly, needs exist to clinically, economically and ergonomically improve known APD systems.

APD systems need to be improved for home use. One common problem with current home systems is that they are susceptible to electrical shock due to "leakage current". Current that flows from or between conductors insulated from one another and from earth is called "leakage current". If any conductor is raised to a potential above earth potential, then some current is bound to flow from that conductor to earth. This is true even of conductors that are well insulated from earth, since there is no such thing as perfect insulation or infinite resistance. The amount of current that flows depends on: (i) the potential, (ii) the capacitive reactance between the conductor and earth and (iii) the resistance between the conductor and earth.

For medical equipment, several different leakage currents are defined according to the paths that the leakage currents take. An "earth leakage current" is the current which normally flows in the earth conductor of a protectively earthed piece of equipment. In medical equipment, impedance to earth from an enclosure is normally much lower through a protective earth conductor than it is through the patient. However, if the

US 7,815,595 B2

3

protective earth conductor becomes open circuited, the patient could be at risk of electrical shock.

"Patient leakage current" is the leakage current that flows through a patient connected to an applied part or parts. It can either flow from the applied parts via the patient to earth or from an external source of high potential via the patient and the applied parts to earth. Other types of leakage currents include "enclosure leakage current", and "patient auxiliary current".

Leakage currents are normally small, however, the amount of current required to produce adverse physiological effects in patients is also small. Accordingly, leakage currents must be limited as much as possible by the design of the equipment and be within safety limits.

SUMMARY

Generally, the present disclosure provides improved dialysis systems and improved methods of performing dialysis. More particularly, the present disclosure provides systems and methods for performing automated peritoneal dialysis ("APD"). The systems and methods of the present disclosure automatically provide dialysis therapy by providing dialysis fluid to the patient and draining spent dialysis fluid from the patient.

Also, the systems and methods of the present disclosure can perform various dialysis therapies. One example of a dialysis therapy which can be performed according to the present disclosure includes an automatic dialysis fluid exchange of a patient fill, dwell and a patient drain. The dialysis system of the present disclosure can automatically perform dialysis therapy on a patient, for example, during nighttime while the patient sleeps.

To this end, in an embodiment a dialysis system is provided. The system includes a fluid supply line. A disposable unit is in fluid communication with the fluid supply line. The disposable unit has at least two flexible membranes that bond together at selected locations and to a rigid plastic piece or manifold. The membranes can be single or double layer. One suitable membrane material is described herein. The membranes seal to one another so as to define a fluid pump receptacle and a fluid heating pathway. The membranes and plastic manifold define a number of flexible valve chambers. The disposable unit also fluidly communicates with a patient line and a drain line.

The manifold and other areas of the disposable unit include reduced or tapered edges that provide an area to seal the membranes. The reduced thickness or tapered area requires less heat than the full thickness, which reduces the heat sinking disparity between the thickness of the manifold of the disposable unit and the thinner flexible membranes. The frame of the manifold is bowed or curved to provide rigidity. The frame is also asymmetrical and designed to be placed into the hardware unit in only one direction.

The hardware unit can be manually transported to a patient's home and opened so that the patient can place a disposable unit therein and closed so that the dialysis unit and the disposable unit cooperatively form a pump chamber that enables dialysis fluid to be pumped to and from the patient. The hardware unit has an enclosure that defines a pump shell, a valve actuator and a heater. The disposable unit is placed in and removed from the enclosure. The fluid pump receptacle of the disposable unit and the shell of the hardware unit form a pump chamber. The pump chamber operates with a pump actuator, which is also located inside the transportable hardware unit.

4

When packaged, a plurality of tubes extend from the disposable unit. The ends of the tubes have connectors that attach to a single body. The body defines or provides a plurality of tip protectors that hold the tubes in an order according to steps of the therapy. The body is configured to slide into the hardware unit of the system from one direction, so that a patient can readily pull the tubes and connectors from the tip protector organizer.

The tip protector used to house the patient fluid connector includes a hydrophobic filter that allows air but not fluid to escape. This vented tip protector enables the system to be primed without having to perform elevation balancing or controlled fluid metering. The system performs a prime by flowing fluid through the system and into the patient fluid line until the dialysate backs up against the filter, causing a fluid pressure increase, which is sensed by the system. The system then stops the pump.

The hardware unit also provides a controller. The controller includes a plurality of processors, a memory device for each processor and input/output capability. One of the processors coordinates operation of the pump actuator, the valve actuator and the heater with the various stages of dialysate flow, such as the fill, dwell and drain stages. The processor also controls or obtains feedback from a plurality of different types of sensors. The sensors include, among others, a capacitance fluid volume sensor, a dialysis fluid temperature sensor, a pressure sensor, a vacuum sensor, an air detection sensor and a mechanical positioning sensor.

In an embodiment, the system uses both preset motion control and adaptive pressure control to control the pressure of fluid within the pump receptacle. The system uses a preset pump motor acceleration to overcome system compliance (i.e., membrane and tubing expansion), which would not otherwise be readily overcome by known proportional, differential or integral control. After the system overcomes compliance, the system converts to an adaptive control using adaptive techniques for controlling pressure by precisely controlling the velocity of a pump motor shaft. The adaptive parameters are modified over time to fine tune the system. This method is especially important for the patient fill and drain cycles, wherein the patient can feel pressure fluctuations. The method also readily compensates for pressure variations due to bag height, bag fullness, etc.

The capacitance fluid volume sensor indicates a volume of fluid in the pump chamber, wherein the sensor generates a voltage signal that is indicative of the volume of fluid in the receptacle. The controller receives the voltage signal and converts the signal into an amount of fluid or an amount of air within the flexible fluid receptacle of the pump chamber.

The pump actuator can be mechanically or pneumatically operated. When mechanically driven, a pump motor drives a vacuum source, such as a piston-cylinder, which pulls a vacuum on the membranes of the fluid receptacle of the disposable unit. Here, a mechanical positioning sensor, such as an encoder, senses the angle of a pump motor shaft relative to a home position and sends a signal to the controller, wherein the controller can adjust the pump motor accordingly. The encoder also provides safety feedback to the controller, whereby the controller, once therapy starts, prevents the camshaft from rotating to a position where the valves can free fill the patient. When the pump actuator is pneumatically operated, the system in an embodiment uses a vacuum pump to pull apart the membranes of the fluid receptacle. Here, the system uses a vacuum sensor to sense the state of the vacuum pump and a mechanical sensing device, such as a linear encoder, to sense the state of a pump piston.

US 7,815,595 B2

5

Thus, in an embodiment, the system maintains a negative pressure on one of the membranes of the fluid receptacle of the disposable unit to pull same away from the other membrane and draw dialysis fluid into the fluid receptacle. The negative pressure on the active membrane is then released, which pushes the membrane towards the other membrane and dispels the dialysis fluid from the pump receptacle. In another embodiment, a mechanical pump piston can be pneumatically attached to one of the membranes, wherein the system mechanically pulls the membrane away from the other membrane. In an embodiment, the membrane is coupled to the pump piston through negative pressure. The pump also includes a diaphragm that is pulled to a bottom side of the piston head, wherein the membrane is pulled to a top side of same. In a further embodiment, the system mechanically pushes one of the membranes while applying the negative pressure to same.

The system also performs other necessary tasks automatically. For example, the system automatically heats the dialysate to a desired temperature while pumping dialysate to the patient. The heater heats the fluid heating pathway defined by the flexible membranes of the disposable unit. In an embodiment, the heater includes an electrical heating plate. Alternatively, or in addition to the heating plate, the heater includes an infrared heating source. In an embodiment, the fluid heating pathway and the heater define an in-line heater that heats dialysate as it travels from the supply bag to the patient.

The system employs a method of heat control that uses a knowledge-based algorithm and a fuzzy logic based algorithm. The former uses laws of physics, empirical data and sensed inputted signals. The latter inputs a difference between desired and actual temperatures and uses fuzzy logic membership functions and fuzzy logic rules. Each algorithm operates at a different update frequency. Each algorithm outputs a duty cycle, wherein the system weights the fuzzy logic based duty cycle relative to the knowledge based duty cycle and produces an overall heater control duty cycle. This method enables accurate dialysate temperature control.

The system automatically purges air from the dialysate, for example, through the pump chamber. The system also senses a total volume of fluid pumped to the patient, records and logs same. Furthermore, the system knows the instantaneous flow rate and fluid pressure of fluid entering or leaving the patient's peritoneal cavity.

The disposable unit includes a valve manifold. The manifold defines a plurality of valve chambers. The hardware unit includes a valve actuator that selectively and sequentially presses against one or more of the valve chambers. In an embodiment, a mechanically operated valve actuator includes a single camshaft and a plurality of cams. The cams press against one of the membranes of the disposable unit to engage the other membrane and block or disallow fluid flow. As stated above, the system uses a sensing device, such as a rotary encoder, to sense the angle of the camshaft relative to a home position, so that the controller can rotate the camshaft to open or close one or more valves as desired. The single camshaft toggles back and forth between: supply and pump chamber fill positions; patient drain and system drain positions; and between pump chamber fill and patient fill positions. These positions are actuated by a unique rotational position on an overall cam profile (i.e., the superposition of each of the individual cams as seen from the end of the camshaft).

The disposable unit of the present invention is provided in a variety of different forms. In an embodiment, the portion of the disposable unit forming the heating path is formed by the same membranes that seal to the rigid member or manifold

6

that forms the valve chambers. The same membranes also form the pump receptacle. In another embodiment, the disposable unit includes a first set of membranes that form the pump receptacle and the valve manifold via the rigid member. Here, the disposable unit includes a second set of membranes, distinct from the first membranes, which form the fluid heating path. In an embodiment, medical grade tubing connects the first set of membranes to the second set. In particular, the tubing enables the fluid heating path to fluidly connect to the valve manifold.

The disposable unit in another embodiment includes a first flexible membrane and a second flexible membrane that house the pump receptacle, the fluid heating path and the rigid valve manifold. The disposable unit also includes a rigid frame that attaches to at least one of the first and second flexible membranes. The rigid frame enables a patient or operator to place the frame and the disposable unit into the enclosure of the hardware unit of the automated dialysis system. The rigid frame is sized to securely fit into a dedicated place in the enclosure. The rigid frame further holds the disposable unit stable while the patient or operator connects tubes to same. For example, the valve manifold provides ports or other types of connectors for connecting to a supply line, a drain line and a patient line. In an embodiment, the rigid frame extends around or circumvents the membranes including the pump receptacle, fluid heating path and valve manifold. In an embodiment, the rigid frame is plastic. In an embodiment, the rigid frame is bowed along at least two sides to increase the rigidity of the disposable unit and to keep the disposable unit from deforming during the heat sealing portion of its manufacture.

In an embodiment, the rigid member or manifold of the disposable unit includes interfaces that allow the membranes to be more easily sealed to the manifold. The manifold edges are tapered to reduce the heat needed to form a cohesive bond between the membranes and the plastic valve manifold. The knife-like tapered edges also reduce or eliminate the gap between the top and bottom membranes, which minimizes the opportunity for leaks to occur in the disposable unit. The chamfered edges also reduce the likelihood that the heat sealing process will burn through the membranes.

The hardware unit described above includes a display device that provides dialysis system information. The display device also enables the patient or operator to enter information and commands into the controller. For example, the display device can include an associated touch screen that enables the patient or operator to initiate automatic flow of the dialysate through the disposable unit. The system begins to pneumatically and/or mechanically pump dialysate through the pump chamber, past the in-line heater and into the patient's peritoneal cavity. Thereafter, the system automatically runs the other cycles of dialysis therapy, for example, while the patient sleeps and/or at night. The automated system not only transfers dialysate from a supply container to the patient, the system allows the dialysate to dwell inside the patient for an amount of time and automatically operates to transfer the dialysate from the patient to a drain.

The system provides a graphical user interface ("GUI"). The GUI in an embodiment employs an embedded web browser and an embedded web server. The web browser and server operate on a main microprocessor for the system. The GUI also employs instrument access and control software, which operates on the main system processor and on one or more delegate processors. The instrument access and control software controls lower level devices, such as the heater and

US 7,815,595 B2

7

the pump. The GUI also provides intermediate software that allows the web browser to communicate with the instrument access and control software.

The GUI displays a number of therapy set-up screens and a number of dialysis treatment screens. The set-up screens generally walk the patient through the set-up portion of the therapy. The system waits for an operator input before proceeding to the next set-up screen. The set-up screens provide information to the patient in the form of real-life images of the equipment and through animations of the actions needed to connect the system to the patient.

The therapy treatment screens display the various cycles of the therapy to the patient in real-time or substantially in-real time. The therapy treatment screens display information such as cycle time in both a graphical and quantitative manner. The therapy treatment screens do not require input from a patient, who may be sleeping while these screens are displayed. When the therapy is complete, the system once again displays a number of disconnection screens which, like the set-up screens, wait for an input from the patient before performing an action.

The treatment screens are colored and lighted for night time viewing, and may be easily seen from a distance of about ten to fifteen feet, however, the screens are lighted so as not to wake a sleeping patient. In an embodiment, the background of the screens is black, while the graphics are ruby red. In contrast, the set-up screens are lighted and colored for daytime viewing.

With the above embodiments, one advantage of the present disclosure is to provide improved systems and methods for performing dialysis.

Another advantage of the present disclosure is to provide improved systems and methods for performing peritoneal dialysis.

A further advantage of the present disclosure is to provide an automated peritoneal dialysis system and method of operating same.

Still another advantage of the present disclosure is to provide an automated peritoneal dialysis system that provides dialysis therapy advantages.

Still a further advantage of the present disclosure is to provide an automated peritoneal dialysis system that has economic advantages.

Yet another advantage of the present disclosure is to provide an automated peritoneal dialysis system that has quality of life advantages.

A still further advantage of the present disclosure is to provide a disposable unit having bowed sides, which increase rigidity and decrease flexing of disposable unit.

Moreover, an advantage of the present disclosure is to provide a disposable unit having tapered interfaces that decrease the heat sinking of the semi-rigid manifold and provide a more robust seal.

Various features and advantages of the present invention can become apparent upon reading this disclosure including the appended claims with reference to the accompanying drawings. The advantages may be desired, but not necessarily required to practice the present disclosure.

BRIEF DESCRIPTION OF THE FIGURES

FIG. 1 schematically illustrates an embodiment of an automated dialysis system of the present disclosure having a mechanically actuated fluid pump.

FIG. 2 schematically illustrates another embodiment of an automated dialysis system of the present disclosure having a fluidly actuated fluid pump.

8

FIGS. 3A and 3B illustrate perspective views of the hardware unit and disposable unit of the present disclosure.

FIG. 4A is a plan view of one embodiment of the hardware and disposable units of the present disclosure.

FIG. 4B is a cross-sectional view taken along line 4B-4B in FIG. 4A, which shows one possible configuration of the system components within the hardware unit.

FIGS. 5 and 6 illustrate additional embodiments of the disposable unit of the present disclosure.

FIG. 7 is a perspective view of one embodiment of a valve manifold that includes a reduced thickness interface for sealing to membranes of a disposable dialysis unit.

FIG. 8 is a perspective view of one embodiment of a multiple tip protector organizer of the present disclosure.

FIG. 9 is an elevation sectional view of the multiple tip protector organizer illustrated in FIG. 8.

FIG. 10 is an elevation sectional view of one embodiment of a vented tip protector of the present disclosure showing the tip protector housing a patient fluid line connector.

FIG. 11 is an elevation sectional view of one embodiment of the patient fluid line connector that couples to the vented tip protector of the present disclosure.

FIG. 12 is an elevation sectional view of one embodiment of the vented tip protector of the present disclosure.

FIG. 13 is a sectional view of one embodiment of a single layer film structure for the disposable unit membranes of the present disclosure.

FIG. 14 is a sectional view of one embodiment of a multiple layer film structure for the disposable unit membranes of the present disclosure.

FIG. 15 is a perspective view of one embodiment of a valve actuator in combination with the fluid manifold of the present disclosure.

FIGS. 16A and 16B illustrate features of the camshaft and cam arrangement of the present disclosure.

FIGS. 17A and 17B illustrate an embodiment of a mechanically operated fluid pump and capacitance type fluid volume sensor of the present disclosure.

FIG. 18 illustrates an alternate embodiment of a fluidly operated fluid pump and capacitance sensor of the present disclosure.

FIG. 19 is a graphical illustration of one embodiment of the present disclosure for the control of the pressure inside a fluid pump through precise velocity control of a pump piston.

FIG. 20 is a schematic illustration of one embodiment of an algorithm of the present disclosure for performing proportional, integral and derivative type adaptive pressure control.

FIG. 21 is a graphical illustration of one embodiment of the present disclosure for the control of the pressure inside a fluid pump during repeated patient fill and pull from supply bag strokes.

FIG. 22 is a graphical illustration of one embodiment of the present disclosure for the control of the pressure inside a fluid pump during repeated patient drain and pump to drain strokes.

FIG. 23 is a schematic illustration of one embodiment of an algorithm of the present disclosure for adapting pressure error correction parameters over time to optimize pressure control efficiency.

FIG. 24 is a table illustrating one set of the correction parameters illustrated in connection with FIG. 23.

FIG. 25 is a schematic representation of one embodiment of a heater control method of the present disclosure.

FIG. 26 is a flow diagram of a knowledge based algorithm of the method discussed in connection with FIG. 25.

FIG. 27 is a flow diagram of a fuzzy logic based algorithm of the method discussed in connection with FIG. 25.

US 7,815,595 B2

9

FIG. 28 is an electrical insulation diagram illustrating one embodiment for providing double electrical insulation in the medical fluid unit of the present disclosure.

FIG. 29 is a schematic representation of one embodiment of the web based graphical user interface of the present disclosure.

FIGS. 30A to 30M are screen shots from a display device employing the graphical user interface of the present invention.

DETAILED DESCRIPTION

The present disclosure relates to dialysis systems and methods of performing dialysis. In particular, the present disclosure relates to a system and method for automatically providing peritoneal dialysis therapy to patients. The present disclosure provides automatic multiple exchanges of dialysis fluid to and from the patient's peritoneal cavity. The automatic exchanges of dialysate include drain, fill, and dwell periods, which usually occur while the patient sleeps. A typical therapy can include three to five exchanges of dialysis fluid. The present disclosure, in an embodiment, provides a single pass system, wherein the dialysate passes through the peritoneal cavity only once before being disposed. While the present disclosure performs peritoneal dialysis, it is also suitable for other types of dialysis and other medical fluid transfer operations.

I. The System Generally

Referring now to the drawings and in particular to FIG. 1, a typical therapy performed by the system 10 of the present disclosure begins by draining dialysis solution that is already in the patient's peritoneal cavity 12. The system 10 pumps fresh dialysate from one of a plurality of supply bags 14, through an in-line heater 16 to the patient or peritoneal cavity 12. After a dwell period in the peritoneal cavity 12, the spent dialysate in the cavity is pumped out of the patient or cavity 12 to a drain 18 or other disposal means. The system 10 then pumps fresh dialysate from the supply bags 14 to the patient or peritoneal cavity 12 and the procedure is repeated as defined in the therapy protocol. The system 10 in an embodiment pumps a last bag of dialysate (usually, a dialysate having a different formulation than the dialysate in the other supply bags) to the peritoneal cavity 12 for an extended dwell, such as a daytime dwell.

In an embodiment, the system 10 includes a mechanically operated diaphragm pump 20. The mechanically operated diaphragm pump 20 employs a pump motor 22 and a linear pump actuator 24. A vacuum may also be used with the mechanical actuator for the diaphragm pump 20, as described in further detail below. In another embodiment illustrated in FIG. 2, the pump is completely fluidly activated.

In FIG. 1 the system 10 also includes a valve actuator 26, which mechanically actuates valves V1 to V5. A controller 30 controls the valve actuator 26 to open valves V1 to V5 as necessary to achieve the desired direction of dialysate fluid flow. In an embodiment, the valve actuator 26 includes a valve motor 28 and a camshaft (illustrated below), which opens one or more of the valves V1 to V5 to achieve the desired dialysate flow.

The controller 30 includes a plurality of processors and a memory device for each processor. The processors include a main microprocessor and a number of delegate processors. The main microprocessor runs certain higher level tasks such as the graphical user interface ("GUI") described below. The delegate processors perform lower level tasks, such as mov-

10

ing valves, reading sensors, controlling heater duty cycle, etc. An additional processor is provided solely for the purpose of tracking safety parameters, such as heater plate and medical fluid temperature. For purposes of the present disclosure, except where otherwise specified, the term "processor 34" refers collectively to all of the processors and the term "memory device 32" refers collectively to all of the corresponding memory devices.

The controller 30 also includes an input/output ("I/O") module 36. The memory device 32 stores a computer program that contains a step by step sequence for the system 10 and configures certain outputs to occur upon specified inputs. The processor 34 runs the program in the memory device 32. The I/O module 36 accepts signal lines from various sensors. The I/O module 36 also connects to power lines including input power lines (including if battery powered) and power lines outputted to the various electrical components.

The controller 30, in an embodiment, includes a video controller 38, which may be a video card. The controller 30 also includes a display device or video monitor 40 that displays medical treatment or dialysis information to a patient or operator. In an embodiment, the controller 30 further includes a touch screen 42 that interfaces with the video monitor 40 and electrically communicates with the I/O module 36. The touch screen 42 enables the patient or operator to input medical treatment or dialysis information into the controller 30.

The controller 30 controls the heater 16, the pump 20 and the valve actuator 26 in a number of different phases that make up a single medical or dialysis treatment. In a first pump fill phase, controller 30 activates the pump 20 to pump medical fluid or dialysate from one of the supply bags 14. In FIG. 1, the controller 30 commands a vacuum source 44, including an air pump motor 46, to pull a vacuum on both sides of the pump 20 through a first vacuum line 48 and a second vacuum line 50. The vacuum lines 48 and 50 pull respective vacuums through first and second pump chamber walls to suction one of a pair of opposing membranes inside the pump chamber against the interior of the pump chamber. The other membrane is held against a piston head in the pump 20. The other membrane alternatively temporarily or permanently mechanically attaches to the piston head, rendering the vacuum on the piston side of the pump 20 unnecessary.

With the membranes maintained against the interior of the pump chamber and the piston head, the controller 30 commands the linear actuator 24 to withdraw within the pump 20. The withdrawal causes the membranes inside the pump chamber to pull further apart. At this time, the controller 30 controls the valve actuator 26 so that only valve V1 is open. The pulling apart of the membranes causes a negative pressure to occur in fill line 52, wherein the negative pressure pulls medical fluid or dialysate from the supply bag 14, through the fill line 52, into a receptacle created by the opened membranes inside the pump chamber of pump 20.

In a patient fill phase, with the negative pressure still maintained by the vacuum source 44, through the pump chamber walls, on the interior membranes, the controller 30 causes the linear pump actuator 24 to move upwards within the pump 20. The upward movement of the actuator 24 and an attached piston head provides a positive mechanical pressure that closes the membrane receptacle and thereby pumps the medical fluid out of the pump 20. At this time, the controller 30 controls the valve actuator 26 so that only valves V2 and V3 are open. Consequently, all of the fluid exiting pump 20 is pumped through a heater line 54, past the in-line heater 16, through a catheter line 56, and into the patient, for example, the patient's peritoneal cavity 12. The catheter line 56 in an embodiment connects to a single lumen catheter, which is

US 7,815,595 B2

11

implanted into the patient 12. Although, in other embodiments, the system 10 can employ a multi-lumen catheter.

The heater 16 in an embodiment includes one or more electrical heating plates, which heat the medical fluid to roughly body temperature. The controller 30 energizes and de-energizes the heater 16 as necessary to obtain the proper fluid temperature. The controller 30 can close valves V2 and V3, located on opposing sides of the heater 16 in the heater line 54, if the medical fluid is too hot or too cold. The improperly heated dialysate does not enter the peritoneal cavity 12.

The controller 20 repeats the pump fill phase and the heater fill phase until the patient's the peritoneal cavity 12 becomes full of fluid according to the therapy protocol. In an embodiment, the volume inside the pump is about thirty to fifty milliliters, and an adult patient typically uses about two liters of dialysis fluid. Accordingly, the pump fill phase and the heater fill phase can be repeated on the order of fifty times. In an embodiment, the pump actuator 24 maintains a fluid pressure at the pump 20 of about three pounds per square inch ("psi").

The system 10 provides a fluid volume sensor 60, which measures the actual volume of medical fluid that has been forced through the pump 20. By summing multiple individual pump volumes, the controller accurately knows how much medical fluid or dialysate has been delivered to the patient 12. The system 10 in an embodiment repeats the pump fill phase and the heater fill phase until the pump 20 has delivered a predetermined volume of medical fluid. The predetermined volume can be inputted into the controller 30 by a patient or operator via the touch screen 42.

In a dwell phase, the controller 30 lets the medical fluid or dialysate remain within the patient 12 for an amount of time, which can be controlled by the controller 30, the patient 12 or an operator. In an embodiment, the controller 30 determines the dwell time, but the patient 30 12 or operator can override the system 10 and command that the system 10 remove the medical fluid from the patient 12.

In a second pump fill phase, the medical fluid is removed from the patient 12. The controller 30 and the actuator 26 open valve V4, while shutting the remaining valves. With the vacuum source still maintaining a negative pressure on the membranes inside the pump 20, the linear actuator 24 withdraws the pump piston within the chamber of pump 20 and reopens the receptacle between the membranes. The negative pressure created by the opening receptacle pulls the medical fluid from the patient 12, through the catheter line 56 and into the membrane receptacle formed inside the pump 20.

In a drain phase, with the negative pressure still maintained by the vacuum source 44, through the pump chamber walls, on the interior membranes, the controller 30 causes the linear pump actuator 24 to move upwardly within the pump 20. The upward movement of the actuator 24 causes a positive mechanical pressure to close the membrane receptacle and thereby pump the medical fluid out of the pump 20. At this time, the controller 30 controls the valve actuator 26 so that only valve V5 is open. Consequently, all of the fluid exiting pump 20 is pumped through a drain line 58 and into the drain 18. Drain 18 can be a drain bag or a drain pipe inside a home, a hospital or elsewhere.

One embodiment of the fluid volume sensor 60 is described in more detail below in connection with the description of the diaphragm pump 20. Besides the fluid volume sensor 60, the system 10 includes various other desired types of sensors.

The system 10 includes temperature sensors 62, such as the sensors T1 to T4, which measure the temperature at relevant places within the system 10. In an embodiment, the sensors 62 are non-invasive, however, any other types of temperature

12

sensors may be employed. As illustrated in FIG. 1, sensors T1 and T2 provide redundant post heater feedback of the fluid temperature to the controller 30. Sensor T3 provides a temperature of the medical fluid prior to heating. Sensor T4 provides the ambient temperature.

The system 10 also provides temperature sensors 62 that monitor the temperature of the heater 16. In an embodiment, the heater 16 is an in-line plate heater. The in-line plate heater 16 can have one or more heater plates, for example, two heater plates having a disposable unit placed between same. Separate temperature sensors PT1 and PT2 are provided to monitor the temperature of each of the plates of the plate heater. The system 10 can thereby control each plate heater individually.

The system 10 includes one or more air sensors 64, such as the sensor AS1, placed directly at the throat of the inlet and outlet of the pump 20. Another air sensor AS2 monitors air in the medical fluid after it leaves the heater 16 and just before the final shut-off valve V3 leading to the catheter line 56. The controller 30 monitors the air content sensed by the air sensors 64 and thereby controls the system 10 to perform any necessary air purge. The system 10 can separate and discharge the air from the fluid or simply convey the air to the drain 18. The system 10 also includes an air vent solenoid 66, which is operated by the controller 30. The air vent solenoid 66 enables the system 10 to relieve the vacuum applied to one or both of the membranes in the pump 20.

The system 10 can accumulate air for various reasons. For example, the valves V1 to V5 and fluid lines, such as lines 52, 54, 56 and 58 may contain air prior to priming the system 10. The supply bags 14 may also introduce air into the pump 20. The patient 12 can also produce certain gasses, which become entrained in the dialysate and enter the pump 20. Further, if minor leaks exist in the fluid disposable or the connections to the supply bag 14, the catheter at the patient 12, or the drain bag, the pump 20 can draw air in through the leaks.

The system 10 provides various fluid pressure sensors 68. Fluid pressure sensors FP1 and FP2 provide a redundant pressure reading of the fluid in the fill line 52 leading to the pump 60. The fluid pressure sensors 68 provide a signal to the controller 30 that indicates the respective fluid pressure at that location. Based on the signals from the pressure sensors FP1 and FP2, the controller 30 operates the fluid pumps and valves to obtain and maintain a desired fluid pressure. As stated above, the system 10 maintains the pump pressure, for example, at about three psi.

The system 10 also provides various valve pressure sensors 70. Valve pressure sensors VP1 to VP5 detect the fluid pressure at the valves V1 to V5. The system 10 further provides one or more vacuum pressure sensors 72, for example, at the vacuum source 44, to ensure that a proper vacuum is maintained on the membrane receptacle within the pump 20.

In an embodiment, the fluid pressure, valve pressure and vacuum sensors 68, 70 and 72, respectively, are non-invasive sensors. That is, the sensors do not physically contact (and possibly contaminate) the medical fluid or dialysate. Of course, the system 10 can include other flow and pressure devices, such as flow rate sensors, pressure gauges, flowmeters, or pressure regulators in any suitable quantity and at any desired location.

The system 10 also includes various positioning sensors. In an embodiment, the positioning sensors include a linear encoder 74 that monitors the position of the linear pump actuator 24 and a rotary encoder 76 that monitors the angular position of the valve actuator 26 or camshaft. An encoder is one type of positioning feedback device that can be employed. Other types of positioning feedback systems

US 7,815,595 B2

13

include proximity sensors and magnetic pick-ups that sense a pulse, e.g., a gear tooth of a gear attached to the camshaft, and output the pulse to a counter or microprocessor.

The encoders 74 and 76 also typically provide a pulsed output, which is sent to the controller 30. The pulsed output tells the controller 30 how many steps or how far the linear pump actuator 24 or the valve actuator 26 is from a home position or home index 78. For example, the home position 78 can be the pump fully open or pump fully closed position for the linear encoder 74 and the zero degree position for the rotary encoder 76.

In an embodiment, the encoders 74 and 76 are absolute type encoders that know the location of the home position 78 even after a power loss. In another embodiment, the encoders 74 and 76 are incremental encoders and a battery back-up is provided to the controller so that the system 10 can maintain the location of the home position 78 even when no external power is applied. Further alternatively, system 10 can be programmed to automatically move the pump actuator 24 and the valve actuator 26 upon power-up until a home position is sensed, wherein the system 10 can begin to run the main sequence.

Referring now to FIG. 2, an alternative system 100 is illustrated. The system 100 includes many of the same components having the same functionality (and the same reference numbers) as previously described. These components therefore do not need to be described again except to the extent that their functioning with the new components of system 100 differs. The primary difference between the system 100 and the system 10 is that the pump 120 of the system 100 is completely fluidly actuated and does not use the linear pump actuator 24 of the system 10.

In the pump fill phases, described above, the controller 30 activates the pump 120 to pump medical fluid or dialysate from one of the supply bags 14. To do so, the controller 30 commands vacuum source 44 (shown separately from motor 46 in FIG. 2), including a vacuum pump motor 46, to pull a vacuum on both sides of the pump 120, i.e., on both pump membranes, through vacuum lines 148 and 149. The vacuum pump motor 46 in this embodiment includes a rotary encoder 76 and a home position or home index 78. The rotary encoder 76 provides positional feedback of a member 150 within the vacuum source 44. The system 100 therefore knows if the vacuum source 44 can provide any additional suction or if the member 150 has bottomed out within the vacuum source 44.

To draw in medical fluid, the vacuum line 148 pulls a vacuum through first and second pump chamber walls to the pair of opposing membranes inside the pump chamber. The vacuum pulls the membranes against the interior of the pump chamber. At this time, the controller 30 controls the valve actuator 26 so that only valve V1 is open. The pulling apart of the membranes causes a negative pressure to occur in fill line 52, wherein the negative pressure pulls medical fluid or dialysate from the supply bag 14, through the fill line 52, into a receptacle created by the volume between the membranes inside the pump chamber of pump 120.

In an alternative embodiment, the pump 120 maintains a constant vacuum on one of the membranes, wherein the opposing membrane does the pumping work. To pump fluid out, the vacuum on one or both membranes is released. The membranes, which have been stretched apart, spring back to a closed position. This operation is described in detail below.

The system 100 also includes a slightly different valve manifold than the system 10. The system 100 includes one less valve than the system 10, wherein the system 100 does not provide an extra valve (V3 in system 10) directly after the fluid heater 16. Obviously, those of skill in the art can find

14

many ways to configure the valves and fluid flow lines of the systems 10 and 100. Consequently, the configuration of the valves and fluid flow lines of the systems 10 and 100 as illustrated merely represent practical examples, and the present disclosure is not limited to same.

II. Hardware Unit and Disposable Unit

Referring now to FIGS. 3A, 3B, 4A and 4B, both systems 10 and 100 include a hardware unit 110 and a disposable unit 160. The hardware unit 110 in an embodiment is portable and can be transported to and from a person's home. The hardware unit 110 includes a housing 112 that includes a base 114 and a lid 116. In an embodiment, the lid 116 is hinged to the base 114. Alternatively, the lid 116 is completely removable from the base. The lid 116 in either case opens to provide access to the interior of the housing 112, so as to allow the patient or operator to place and remove the disposable unit 160 into and from the hardware unit 110. The hardware unit 110 can be made of any protective, hard, resilient and/or flexible material, for example, plastic or metal sheet, and can have a decorative and/or finished surface.

Once the disposable unit 160 is placed inside the hardware unit 110, the operator closes the lid 116 and uses one or more locking or latching mechanism 118 (FIG. 3B) to safely house the disposable unit 160 within the hardware unit 110. FIG. 4A illustrates members 119 of the housing 112 to which the latching mechanism 118 of the lid 116 attaches. The hardware unit 110 displays the video monitor 40, which can have an associated touch screen 42 to input commands as described above. Alternatively, or in addition to the touch screen 42, the hardware unit 110 can provide one or more electromechanical switches or pushbuttons 43, 124, 125 and 127, analog controls 122 and/or lighted displays. The pushbuttons or switches 43, 124, 125 and 127 and knob 122 enable the patient or operator to input commands and information into the systems 10 and 100. The video monitor 40 provides medical treatment information 126 to the patient or operator.

FIG. 3B illustrates one set of dimensions for the hardware unit 110 of the present disclosure. The size and weight of the present disclosure are less than previous automated dialysis system. This feature belies the portability and ease of use of the system 10, 100 of the present disclosure. The size and weight enable the hardware unit 110 to be shipped economically by standard overnight courier services. In the event that the system 10, 100 of the present disclosure breaks down, a replacement unit can be economically shipped to the patient in time for the next therapy.

The hardware unit 110 in an embodiment is approximately 23 to 30 cm high and deep and in one preferred embodiment, as illustrated, about 25 cm high and deep. The hardware unit 110 in an embodiment is approximately 32 to 40 cm wide and in one preferred embodiment, as illustrated, about 34 cm wide. The internal volume of the unit 110 is therefore about 17,000 cm³ to about 36,000 cm³, and in one preferred embodiment, approximately 21,250 cm³ (1310 in³). Section view 4B aptly illustrates the many components maintained within this compact space and the efficient use of same. All these components and the hardware unit 110 have a total mass of about six to nine kilograms (kg) and in one preferred embodiment about seven kilograms.

FIGS. 3A to 4B also illustrate that the architecture, configuration and layout of the hardware unit 110 provides an automated system that is also convenient to use. The components of the system 10, 100 with which the patient must interact are placed on the top, front and sides of the unit 110. The flow control components are placed below the heater 116,

US 7,815,595 B2

15

which is placed below the disposable unit loading station. The monitor **40** and controls **43**, **122**, **124**, **125** and **127** are placed in the front of the unit **110**.

The hardware unit **110** contains the pump **20** or **120** and the linear pump actuator **24** if system **10** is employed. The hardware unit **110** also contains the valve actuator **26** including the valve motor **28**, the in-line heater **16**, the various sensors, the vacuum source **44** including the air pump motor **46** and the controller **30** as well as the other hardware described above. FIG. **4B** illustrates that one of the pump chamber walls of the pump **20** or **120** is disposed in the lid **116** of the housing. In FIG. **4B**, the heater **16** is disposed in the base **114** of the housing **112**. Alternatively or additionally, the heater may be placed in the lid **116**. The base **114** also contains the opposing pump chamber wall.

Referring now to FIGS. **3A**, **4A**, **4B**, **5** and **6**, various embodiments of the disposable unit **160** are illustrated. In each of the embodiments, the disposable unit **160** includes a pair of flexible membranes, including an upper flexible membrane **162** and a lower flexible membrane **164**. The disposable unit **160** of FIG. **6** includes two pairs of flexible membranes, namely, membrane pair **166** and membrane pair **168**. Each of the membrane pairs **166** and **168** also includes the upper flexible membrane **162** and the lower flexible membrane **164**.

The flexible membranes **162** and **164** can be made of any suitable sterile and inert material, such as a sterile and inert plastic or rubber. For example, the membranes **162** and **164** can be buna-N, butyl, hypalon, kel-F, kynar, neoprene, nylon, polyethylene, polystyrene, polypropylene, polyvinyl chloride, silicone, vinyl, viton or any combination of these. One suitable material for the flexible membrane is described below in connection with FIGS. **13** and **14**.

The membranes **162** and **164** are sealed together in various places to create fluid flow paths and receptacles between the membranes **162** and **164**. The seals are heat seals, adhesive seals or a combination of both. FIGS. **3A**, **4A**, **5** and **6** illustrate that a generally circular seal **170** creates a substantially circular fluid pump receptacle **172** between the membranes **162** and **164**. The pump receptacle **172** operates with the fluid pumps. Instead of the seal **170**, one alternative embodiment is for the base **114** and lid **116** to press the membranes together to form the seal. FIGS. **4A** and **5** illustrate that in an embodiment, the disposable unit **160** provides a secondary seal **174** to protect the systems **10** and **100** in case the primary seal **170** leaks or degrades during use.

FIGS. **3A**, **4A** and **4B** illustrate that the fluid pump receptacle **172** fits between the clamshell shapes of the pumps **20** and **120** in the lid **116**. The clamshell shapes defined by the base **114** and lid **116** of the hardware unit **110** together with the fluid pump receptacle **172** form the pump chamber of the pumps **20** and **120** of the present disclosure. The clamshell shapes in the base **114** and lid **116** include one or more ports with which to draw a vacuum on the membranes **162** and **164**. In this manner, the membranes **162** and **164** are pulled towards and conform to the clamshell shapes in the base **114** and lid **116** and thereby create a negative pressure inside the receptacle **172** that pulls medical fluid from a supply bag **14** located outside the hardware unit **110**, into the receptacle **172**.

FIGS. **3A**, **4A**, **5** and **6** illustrate that a generally rectangular, spiral seal **178** creates a spiral heating path **180** between the membranes **162** and **164**. The fluid heating path **180** runs from a valve manifold **190**, through the spiral section, and back to the valve manifold **190**. FIG. **4A** illustrates that the fluid heating path **180** fits between the heating plates of the heater **16**, which reside in the base **114** and lid **116** of the hardware unit **110**. Providing a heat source on either side of the fluid heating path **180** enables the medical fluid to be

16

quickly and efficiently heated. In alternative embodiments, however, the heater **16** can include only a single heater on one side of the fluid heating path **180** defined by the disposable unit **160** or multiple heaters on each side of the disposable unit **160**.

The upper and lower membranes **162**, **164** are attached to the disposable unit **160** utilizing heat sealing techniques as described herein. The membranes **162** and **164** are expandable so that when the disposable unit **160** is placed between a predefined gap between the upper and lower plates of the heater **16**, the membranes **162** and **164** expand and contact the heater plates. This causes conductive heating to take place between the plates of the heater **16** and the membranes **162**, **164** and between the membranes and the medical fluid. The predefined gap is slightly larger than the thickness of the disposable unit **160**. Specifically, when dialysate moves through the fluid heating path **180** of the disposable unit **160**, the membranes **162**, **164** of the spiral wound fluid heating pathway **180** expand between the spiral seal **178** and touch the plates of the heater **16**.

A. Separate Sets of Membranes

The disposable unit **160** of FIG. **6** is similar to the disposable units **160** of FIGS. **3A** through **5**. The in-line fluid heating path **180**, however, is placed in a separate membrane pair **166** from the fluid pump receptacle **172** and the valve manifold **190**, which are placed in a separate membrane pair **168**. A pair of flexible tubes **182** and **184**, which can be any suitable medical grade tubing, fluidly connect the valve manifold **190** to the fluid heating path **180**. The tubes **182** and **184** can be connected to the membrane pairs **166**, **168** by any desired means, such as, heat sealing, bonding, press-fitting or by any other permanent or removable fluid connection. When placed in the hardware unit **110**, the heater **16** heats each side of the heater membrane pair **166**, as in the other embodiments.

Separating the fluid heating path **180** from the fluid pump receptacle **172** and the valve manifold **190** enables the membranes of the respective pairs to be made of different materials. It is desirable that the membranes **162** and **164** of the heating pair **166** conduct or radiate heat efficiently. On the other hand, it is desirable that the membranes **162** and **164** of the fluid flow pair **168** withstand the forces of suction and mechanical actuation. It may therefore be desirable to use dissimilar materials for the membrane pair **166** and the membrane pair **168**.

The membrane pair **166**, defining the heater fluid flow path **180**, additionally defines alignment holes **176** that align with pegs protruding from the base **114** or the lid **116** of the hardware unit **110**. Each of the embodiments of the disposable unit **160** disclosed herein may be adapted to include alignment holes **176**, which aid the patient or operator in properly placing the disposable unit **160** within the housing **112** of the hardware unit **110**.

B. Rigid Frame and Bowed Sides

As shown in FIGS. **3A**, **4A** and **5**, each of the embodiments of the disposable unit **160** disclosed herein may also be adapted to provide a rigid or semi-rigid member or frame **186**, which in an embodiment, surrounds or substantially circumscribes the membranes **162** and **164** of the disposable unit **160**. In an embodiment, the rigid member or frame **186** is made of a sterile, inert, rigid or semi-rigid plastic, for example, from one of or a combination of the plastics listed above for the membranes **162** and **164**. The frame **186** aids the

US 7,815,595 B2

17

patient or operator in properly placing the disposable unit 160 within the housing 112 of the hardware unit 110.

In an embodiment, the housing 112 defines a pin or guide into which the frame 186 of the disposable unit 160 snugly fits. FIG. 5 illustrates that the frame 186 defines an aperture 161 that fits onto the pin or guide of the housing 112. The frame 186 can provide a plurality of apertures, such as the aperture 161, which fit onto a like number of pins or guides provided by the housing 112. FIG. 5 also illustrates that the frame 186 includes an asymmetrical member or chamfer 163. The chamfer 163 forms an angle, such as forty-five degrees, with respect to the other sides of the frame 186. The housing 112 defines or provides an area into which to place the disposable unit 160. The area has the asymmetrical shape of the frame 186 or otherwise provides guides that only allow the unit 160 to be placed in the housing 112 from a single direction. The chamfer 163 and the cooperating housing 112 ensure that when the patient places the disposable unit 160 in the housing 112, the bottom of the disposable unit 160 is placed in the housing 112 and the fluid inlets/outlets 196 face in the proper direction.

As discussed above, the disposable unit 160 includes a valve manifold 190. In an embodiment, the valve manifold 190 is made of a rigid or semi-rigid plastic, such as, from one of or a combination of the plastics listed above for the membranes 162 and 164. The valve manifold 190 is covered on either side by the upper and lower membranes 162 and 164 to thereby create a sealed and inert logic flow path for the systems 10 and 100.

In FIG. 5, the manifold 190 defines holes 192 and slots 194. The holes 192 define the location of the valves, for example, valves V1 to V5 of the system 10. The slots 194 define the fluid flow paths from the valves to the fluid pump receptacle 172, the fluid heating path 180 or to fluid inlets/outlets 196. The fluid inlets/outlets 196 individually lead to the supply bag 14, the catheter line 56, the patient 12 and the drain 18. The fluid inlets/outlets 196 may have various configurations and orientations, as contrasted by FIG. 3A. The drain 196 may also be adapted to connect to an external flexible tube via a method known to those of skill in the art.

In an embodiment, the rigid or semi-rigid frame 186 includes bowed sides 187 and 189, as illustrated in FIG. 5. The bowed sides 187 and 189 are formed with the frame 186 before the membranes 162 and 164 heat seal or adhesively seal to the frame 186 and manifold 190. The frame 186 and bowed sides 187 and 189 can be extruded plastic or plastic injection molded. The frame 186 can include as little as one bowed side, any number less than all, or have all sides be bowed.

In the illustrated embodiment, the sides 187 and 189 bow outward although they can alternatively bow inward. In an embodiment, the sides are bowed in a direction of the plane of the frame 186 of the disposable unit 160. The bowed sides 187 and 189 increase the rigidity of the frame 186 and the disposable unit 160. The disposable unit is accordingly more easily placed in the housing 112 of the hardware unit 110. The bowed sides 187 and 189 reduce the amount of flexing or distortion of the frame 186 due to heat sealing or mechanically pressing membranes 162 and 164 onto the frame 186 and manifold 190.

C. Heat Seal Interface

Referring now to FIG. 7, an embodiment for heat sealing the membranes 162 and 164 to the manifold 190 is illustrated. In an embodiment, the manifold 190 is made of a rigid or semi-rigid plastic material as described above. Heat sealing

18

the membranes 162 and 164 to the semi-rigid manifold 190, which in an embodiment is an injection molded component, requires different processing parameters than heat sealing the individual membranes 162 and 164 together, for example, at seal 170 of the fluid pump receptacle 172. In particular, heat sealing the membranes 162 and 164 to the manifold 190 can require more heat, more pressure and more heating time. The semi-rigid or rigid manifold 190 is appreciably thicker than the individual membranes 162 and 164. Consequently, relative to the thin membranes, the thicker manifold 190 acts as a heat sink. The bond between the thin membrane and thicker manifold 190 therefore requires more heat or energy than the heat seal bond between the thin membranes 162 and 164.

As illustrated in FIGS. 3A, 4A, 5 and 6, the disposable unit 160 requires both membrane to manifold and membrane to membrane seals. It is desirable to heat seal the entire disposable unit 160 in one step or process for obvious reasons. It should also be obvious that the heat sealing process should be performed so as avoid burning or melting one of the thin membranes 162 or 164.

FIG. 7 illustrates one embodiment for solving the heat sinking disparity between varying materials. FIG. 7 illustrates a portion of the manifold 190, which is shown in its entirety in FIG. 5. In FIG. 5, the manifold 190 illustrates a port that connects to the fluid pump receptacle 172. This port is illustrated as port 205 in FIG. 7. FIG. 5 also illustrates two ports extending from the manifold 190 that fluidly connect to the fluid heating path 180. These ports are illustrated as ports 201 and 203 in FIG. 7. Both FIG. 5 and FIG. 7 illustrate that the injection molded manifold 190 defines a plurality of holes 192 and slots 194. The holes 192 operate with the valve actuator and the slots 194 to form fluid pathways when enclosed by the membranes 162 and 164.

To reduce the amount of heat necessary to seal the membranes 162 and 164 to the manifold 190, the manifold 190 includes a side 193 having a lesser thickness than the remaining portion of the manifold 190. The thinner side 193 has less mass and therefore absorbs less localized heat than would a manifold of constant thickness. The side 193 also defines or includes a tapered portion 195. The tapered portion 195 provides flat surfaces on which to seal the membranes 162 and 164 and also positions the membranes 162 and 164 together so that in an embodiment a membrane to membrane seal may also be made in addition to the membrane to manifold 190 seal.

The tapered edges 195 form an interface for the membranes 162 and 164 to seal to the manifold 190, which occurs along continuous stretches of the sides 193 of the manifold 190 that require sealing or that would otherwise come into contact with the medical fluid. Therefore, as illustrated in FIG. 5, the side of the manifold 190 defining the input/output ports 196 does not need to be tapered as illustrated in FIG. 7. Also, as illustrated in FIG. 7, the tapered edges 195 of the thin sides 193 discontinue where the ports 201, 203, and 205 extend from the manifold 190.

The ports 201, 203 and 205 also form tapered edges 207. Tapered edges 207 form an interface for heat sealing the parts to the membranes 162 and 164. As described above, the tapered edges 207 of the ports 201, 203 and 205 also enable a membrane to membrane seal to take place directly next to the membrane to tapered edge 207 seal. The tapered edges 195 and 207 in an embodiment taper gradually towards the knife-like edge. In other embodiments, the tapered edges 195 and 207 may take on different forms or shapes, such as a rounded edge, a blunter edge or may simply be further reduced in thickness from side 193 of the manifold 190. As illustrated, the ports 201, 203 and 205 in an embodiment form ovular

openings. The tapered oval openings provide a smoother transition angle than would a circular outer diameter. The oval openings perform as well as round openings from a fluid flow standpoint as long as open area of the inner oval is not less than the open area of a suitable circular port.

The ports **201**, **203** and **205** also form raised portions **209**. The raised portions **209** form a bead of polymeric material along the tops of the ports **201**, **203** and **205** and the tapered edges **207**. The beads can be additionally or alternatively placed along the tapered edges **195** and or the sides **193**. The raised portions or beads **209** provide an extra thin area of plastic that melts or deforms to provide a flux-like sealant that enables the membranes **162** and **164** to seal to the manifold **190**. The beads create a concentrated strip of higher temperature plastic than the surrounding plastic of the manifold **190**. The membranes **162** and **164** seal to the manifold **190** without having to heat a larger area of the manifold **190**. The raised portions or beads **209** help to seal curved portions and corners created by the manifold **190**.

D. One-Piece Tip Protector Organizer and Vented Tip Protector

Referring now to FIG. 8, one embodiment of a one-piece tip protector organizer **270** is illustrated. In the HOMECHOICE® peritoneal dialysis system provided by the assignee of the present invention, a disposable set is prepackaged and provided to the patient. The patient opens up the package, wherein each of the components is sterilized and maintained within the disposable set. The disposable set includes a disposable unit and a number of tubes emanating from the disposable unit. Like the present invention, the HOMECHOICE® disposable unit includes a drain line tube that connects to one or more fill bag tubes, and a tube that connects to a patient transfer set. Each of these tubes requires a separate tip protector. That is, after sterilizing the inside of the disposable unit and the tubes, for example, using ethylene oxide, the ends of the tubes have to be capped off so that the sterilization of the inside of the system is maintained. The HOMECHOICE® system provides a separate tip protector for each tube.

The one-piece tip protector organizer **270** of the present disclosure provides a single body **272** (which may actually be made of a plurality of pieces) that defines or provides a plurality of tip protectors **274**, **276**, **278** and **280**. The vented tip protector **270** not only houses and protects the connectors at the ends of the tubes emanating from the disposable unit **160**, the one-piece tip protector **270** also organizes and orders the tubes according to the steps of the dialysis therapy. In the illustrated embodiment, the tip protector **274** is a tip protector for a drain line connector **284** connected to a drain line **285** that leads to the appropriate port of the disposable unit **160**. The tip protectors **276** and **278** are supply bag protectors that protect the connectors **286** and **288** that connect to the ends of the tubes **287** and **289** that run to a “Y” connection **287/289**, wherein the leg of the “Y” connection **287/289** runs to the appropriate port of the disposable unit **160**. The tip protector **280** is a patient fluid line protector. The tip protector **280** houses and protects a connector **290** that connects to patient tube **292**, which runs to the appropriate port of the disposable unit **160**.

Each of the tubes **285**, the “Y” connection **287** and **289** and the patient fluid tube **292** in an embodiment are made of polyvinylchloride (“PVC”) having an inner diameter of 4 mm and an outer diameter of 5 mm. As illustrated, the one-piece tip protector organizer **270** is adaptable to receive and protect various types of fluid connectors. The fluid connector **284** that

runs via tube **285** to the drain line port of the disposable unit **160** is in an embodiment largely the same as the port that emanates from the supply bags **14**. The ports that emanate from the supply bags **14** also include a membrane that is pierced by the sharp stem of the supply bag connectors **286** and **288**. The drain line connector **284** does not include the membrane of the supply bag **14** as it is not needed. The tip protector **290** that connects to the end of the patient fluid tube **292** is discussed in detail below.

In one embodiment, the system **10**, **100** of the present disclosure provides two, six liter supply bags **14**. The two, six liter bags provide an economic amount of peritoneal dialysis fluid, which is enough fluid to provide a number of fill, dwell and drain cycles during the evening while the patient sleeps. The one-piece organizer **270** therefore provides two tip protectors **276** and **278**, which house and protect the supply connectors **286** and **288**. In alternative embodiments, the one-piece organizer **270** can define or provide any number of supply bag tip protectors. Any number of supply bags can be additionally linked via “Y” or “T” type tubing links.

The one-piece organizer **270** can provide additional tip protectors such as a last bag protector, which protects a line that runs to a bag that holds enough peritoneal fluid, e.g., two liters, for a final fill for the patient during the daytime. In this case, an additional last bag tube, not illustrated, would connect to a connector, which would be a bag piercing connector, the same as or similar to the fill bag connectors **286** and **288**.

The body **272** of the tip protector organizer **270** is in an embodiment also made of PVC. The tip protectors **274**, **276**, **278** and **280** are injection molded or blow molded. Alternatively, the tip protectors can be separately applied to the body **272**. As seen in FIG. 8, one or more of the tip protectors can include flutes, threads or other protrusions that aid in grasping and holding the respective tube connector. Further, while the organizer **270** is generally referred to herein as a “one-piece” organizer, the organizer **270** may itself be comprised of any number or pieces. “One-piece” refers to the feature that a single unit houses a multitude or tip protectors.

The one-piece organizer **270** also includes a rim **294** that extends outwardly from the main portion of the body **272**, and which circumvents the main portion of the body **272**. Referring now to FIG. 9, a cross section of the one-piece organizer **270** illustrates that the rim **294** tapers downwardly from the drain line tip protector **274** towards the patient fluid tip protector **280**. That is, the rim **294** is higher or thicker at the drain line end than it is at the patient fluid line end. This enables the one-piece tip protector organizer **270** to be mounted to the hardware unit **110** in only one orientation.

FIG. 3A illustrates that the one-piece tip protector organizer **270** in an embodiment slides into the hardware unit **110** vertically. The hardware unit **110** includes or provides a pair of members **296** that extend outwardly from a side wall of the hardware unit **110**. FIGS. 3B and 4A illustrate another embodiment, wherein the rim **294** of the organizer **270** slides vertically into a notch **297** defined or provided by the base **114** of the housing **112** of the hardware unit **110**. The rim **294** of the organizer **270** slides between the members **296** and the side wall of the hardware unit **110**. The members **296** extend further outwardly as they extend running towards the top of the hardware unit **110**. The taper of the members **296** corresponds to the taper of the rim **294** of the organizer **270** so that the organizer **270** can only slide into the hardware unit **110** vertically from one direction.

FIG. 9 also illustrates that the tip protectors **274**, **276**, **278** and **280** can have various cross-sectional shapes. Each of the tip protectors includes a solid bottom and sides that seal around the respective connectors **284**, **286**, **288** and **290**, so

US 7,815,595 B2

21

that the one-piece organizer 270 maintains the sterility of the system even after the patient removes the disposable set from a sealed sterilized container. The one-piece organizer 270 illustrated in FIGS. 8 and 9 mounts in a sturdy fashion to the side of the hardware unit 110. Via this solid connection, the patient is able to remove the tubes 285, 287, 289 and 292 using only one hand in many cases. The interface between the hardware unit 110 and the organizer 270 simplifies the procedure for the patient and provides a solid, sterile environment for the tubes and associated connectors until used.

FIG. 3A also illustrates another possible embodiment wherein an alternative one-piece organizer 298 is integral to or provided by the frame 186 of the disposable unit 160. Here, the tubes 196, indicated generally, are horizontally organized as opposed to the vertical arrangement of the tip protector 270 in the housing 112. The horizontal one-piece organizer 298 illustrates that the concept of protecting and organizing the tubes before use can be provided in a variety of places and orientations in the system 10.

In one embodiment, the tip protector and organizer 270 structures the tubes 285, 287, 289 and 292 in a downwardly vertical order, such that the first tube that the patient is supposed to pull when starting the dialysis therapy is provided on top, the next tubes that the patient is supposed to pull are provided in the middle and the final tube is provided lowest on the vertically oriented one-piece organizer 270. According to one protocol, the patient first removes the drain connector 284 from the tip protector 274 and runs the drain line 285 to a toilet, drain bag or other drain. The patient then removes the supply connectors 286 and 288 and punctures the supply bags 14 (FIGS. 1 and 2). At this point, dialysate can be pumped to the disposable unit 160 and throughout the system 10. The controller 30 of the system 10, 100 begins a priming cycle, which is discussed in more detail below.

Once priming is complete, system 10, 100 prompts the patient to remove the primed patient line 292 and connect same to the transfer set implanted into the patient. The transfer set (not illustrated) includes a catheter positioned into the patient's peritoneal cavity and a tube running to the catheter. The tube also includes a connector that couples to the connector 290. At this point, system 10, 100 can begin to either drain spent peritoneal fluid from the patient 12 to the drain 18 or pull new fluid from one or both of the supply bags 14 and fill the patient's peritoneal cavity 12.

Referring now to FIGS. 10 to 12, one embodiment for the patient line tip protector 280 of the present disclosure is illustrated. The HOMECHOICE® system produced by the assignee of the present disclosure primes the patient fluid line by allowing the patient connector to be held vertically approximately at the same level as the supply bag. In this manner, when the HOMECHOICE® system primes the disposable unit, gravity feeds peritoneal fluid into the patient fluid line up to the end of the patient fluid connector. The patient fluid connector is open so that air can freely escape when the peritoneal fluid is fed by gravity through the patient line. HOMECHOICE® system enables the patient fluid line to be primed without counting pump strokes or having to meter out a known volume of dialysate, techniques which are complicated and prone to failure.

The system 10, 100 of the present disclosure provides a different apparatus and method of priming without having to calculate the amount of fluid that is needed to just reach but not surpass the patient connector of the patient fluid line. FIG. 10 shows a cross-section of the patient fluid connector 290 that has been inserted into the vented tip protector 280. FIG. 11 illustrates a cross section of the patient fluid connector 290 only. FIG. 12 illustrates a cross section of the tip protector 280

22

only. A hydrophobic membrane 300 is placed on the outer edge of the tip protector 280. The tip protector 280 defines a fluid lumen 302 that runs through the entire length of the tip protector 280. The hydrophobic membrane 300 covers the fluid lumen 302. The hydrophobic membrane 300 allows air to purge from inside the patient's fluid line but does not allow water or peritoneal fluid to flow through same.

It should be appreciated that the vented tip protector 280 including the hydrophobic membrane 300 is not limited to being placed in the one-piece tip protector organizer 270. FIG. 9 illustrates that the one-piece organizer 270 does include the patient tip protector 280 having the hydrophobic membrane 300 and the fluid lumen 302. The vented tip protector 280 in an alternative embodiment, however, can be provided as a separate or stand alone tip protector, similar to the one used on the HOMECHOICE® system provided by the assignee of the present disclosure.

Hydrophobic membranes, such as the hydrophobic membrane 300 employed herein, are commercially available. One suitable hydrophobic membrane is produced by Millipore, 80 Ashby Road, Bedford, Mass. 01730. FIG. 12 best illustrates that the hydrophobic membrane heat seals or sonically seals to the tip protector 280. The fluid lumen 302 in an embodiment is relatively small in diameter, such as approximately fifty to seventy thousandths of an inch (1.25 to 1.75 mm).

The vented tip protector 280 and the patient fluid connector 290 also cooperate so that when the system 10, 100 is completely primed, the tip protector 280 and connector 290 minimize the amount of fluid that spills when the patient removes the patient fluid connector 290 from the tip protector 280. The connector 290 includes or provides a male lure 304 that mates with a female lure 306 best seen in FIG. 10. The mating lures 304 and 306 prevent peritoneal fluid from filling the cavity of the tip protector 280, which must be wide enough to house the flange 308 of the patient fluid connector 290. FIG. 12 illustrates that the seal interface between the male lure 304 of the connector 290 and the female lure 306 of the vented tip protector 280 reduces the volume significantly from an interior volume 310 existing around the male lure 304 to the fifty to seventy thousandths diameter of the lumen 302.

To prime the system 10, 100 the patient removes the drain line 285 from the tip protector 274 and places it into a tub, toilet or drain bag 18. The patient removes the two or more supply bag connectors 286 and 288 and punctures seal membranes (not illustrated) of the supply bags 14. System 10, 100 may then automatically begin pump priming or may begin pump priming upon a patient input. In either case, system 10, 100 pumps fluid from one or both of the supply bags 14 through the connectors 286 and 288 and tubes 287 and 289, into the disposal disposable unit 160, out the patient fluid line 292 and into the patient fluid connector 290, which is still housed in the vented tip protector 280 of the one-piece organizer 270. The organizer 270 is vertically housed in the hardware unit 110 as seen in FIGS. 3A and 3B.

When the peritoneal fluid reaches the patient fluid connector 290, most all the air within the system 10 has been pushed through the hydrophobic membrane 300 attached at the end of the tip protector 280 housed in the one-piece tip protector 270. The nature of the hydrophobic membrane 300 is that it allows air to pass through but filters or does not allow water or peritoneal fluid to pass through same. Thus, when the fluid finally reaches the hydrophobic membrane 300, the lack of any additional space in which to flow fluid causes the pressure to increase within the system 10, 100. The system 10, 100 provides one or more pressure sensors, for example pressure sensors 68 (marked as FP1, FP2 and FPT in FIGS. 1 and 2).

US 7,815,595 B2

23

One or more of the pressure sensors **68** sense the increase in pressure due to the peritoneal fluid backing up against the hydrophobic filter **300**. The pressure sensor(s) sends a signal to the I/O module **36** of the controller **30**. The controller **30** receives the signal and is programmed in memory device **32** to shut down the diaphragm pump **20**, **120**. In this manner, the system **10** self-primers each of the fill lines **287** and **289**, the disposable unit **160** and the patient fluid line **292** automatically and without need for controlled volume calculations or gravity feeding.

System **10**, **100** also includes one or more safety features that may be based upon a volume calculation. That is, under normal operations, the system **10**, **100** does not control the priming using a volume calculation. However, in the case where for example the patient removes the patient fluid connector **290** from the vented tip protector **280** of the one-piece tip organizer **270** before the system **10**, **100** senses a pressure increase and stops the pumps **10**, **100**, the system **10**, **100** can employ an alarm calculation, wherein the system **10**, **100** knows that it has pumped too much peritoneal fluid (e.g., a predetermined amount more than the internal volume of the system) and shuts down pump **20**, **120** accordingly.

III. Membrane Material for the Disposable unit

Referring now to FIGS. **13** and **14**, upper and lower membranes **162**, **164** can be fabricated from a monolayer film structure **312** (FIG. **13**) or a multiple layer film structure **312** (FIG. **14**). The film **312** is constructed from a non-PVC containing polymeric material and must satisfy numerous physical property requirements. The film **312** must have a low modulus of elasticity so that it can be deformed under low pressure to function as a pumping element. What is meant by low modulus is the film **312** has a modulus of elasticity when measured in accordance with ASTM D882, of less than about 10,000 psi, more preferably less than about 8,000 psi and even more preferably less than about 5,000 psi and finally, less than about 3,000 psi, or any range or combination of ranges defined by these numbers. The film **312** must have adequate thermal conductivity to allow for in-line heating. The film has a thermal conductivity of greater than 0.13 W/meters-° K when measured using a Hot Disk™ sold by Mathis Instruments Ltd. The film **312** must be capable of being heat sealed to cassette **160**. The film **312** must be capable of being sterilized by exposure to gamma rays, by exposure to steam for a period of time (typically 1 hour), and exposure to ethylene oxide without significant degradation of the film or having an adverse effect on the dialysis solution. Finally, the film **312** must be capable of being extruded at high rates of speed of greater than 50 ft/min.

The monolayer structure **312** is formed from a blend of from about 90% to about 99% by weight of a first component containing a styrene and hydrocarbon copolymer and from about 10% to about 1% of a melt strength enhancing polymer and more preferably a high melt strength polypropylene.

The term "styrene" includes styrene and the various substituted styrenes including alkyl substituted styrene and halogen substituted styrene. The alkyl group can contain from 1 to about 6 carbon atoms. Specific examples of substituted styrenes include alpha-methylstyrene, beta-methylstyrene, vinyltoluene, 3-methylstyrene, 4-methylstyrene, 4-isopropylstyrene, 2,4-dimethylstyrene, o-chlorostyrene, p-chlorostyrene, o-bromostyrene, 2-chloro-4-methylstyrene, etc. Styrene is the most preferred.

The hydrocarbon portion of the styrene and hydrocarbon copolymer includes conjugated dienes. Conjugated dienes which may be utilized are those containing from 4 to about 10

24

carbon atoms and more specifically, from 4 to 6 carbon atoms. Examples include 1,3-butadiene, 2-methyl-1,3-butadiene (isoprene), 2,3-dimethyl-1,3-butadiene, chloroprene, 1,3-pentadiene, 1,3-hexadiene, etc. Mixtures of these conjugated dienes also may be used such as mixtures of butadiene and isoprene. The preferred conjugated dienes are isoprene and 1,3-butadiene.

The styrene and hydrocarbon copolymers can be block copolymers including di-block, tri-block, multiblock, and star block. Specific examples of diblock copolymers include styrene-butadiene, styrene-isoprene, and selectively hydrogenated derivatives thereof. Examples of tri-block polymers include styrene-butadiene-styrene, styrene-isoprene-styrene, alpha-methylstyrene-butadiene-alpha-methylstyrene, and alpha-methylstyrene-isoprene-alpha-methylstyrene and selectively hydrogenated derivatives thereof.

The selective hydrogenation of the above block copolymers may be carried out by a variety of well known processes including hydrogenation in the presence of such catalysts as Raney nickel, noble metals such as platinum, palladium, etc., and soluble transition metal catalysts. Suitable hydrogenation processes which can be used are those wherein the diene-containing polymer or copolymer is dissolved in an inert hydrocarbon diluent such as cyclohexane and hydrogenated by reaction with hydrogen in the presence of a soluble hydrogenation catalyst. Such procedures are described in U.S. Pat. Nos. 3,113,986 and 4,226,952, the disclosures of which are incorporated herein by reference and made a part hereof.

Particularly useful hydrogenated block copolymers are the hydrogenated block copolymers of styrene-isoprene-styrene, such as a polystyrene-(ethylene/propylene)-polystyrene block polymer. When a polystyrene-polybutadiene-polystyrene block copolymer is hydrogenated, the resulting product resembles a regular copolymer block of ethylene and 1-butene (EB). This hydrogenated block copolymer is often referred to as SEBS. When the conjugated diene employed is isoprene, the resulting hydrogenated product resembles a regular copolymer block of ethylene and propylene (EP). This hydrogenated block copolymer is often referred to as SEPS. When the conjugated diene is a mixture of isoprene and butadiene the selectively hydrogenated product is referred to as SEEPS. Suitable SEBS, SEPS and SEEPS copolymers are sold by Shell Oil under the tradename KRATON, by Kuraray under the tradename SEPTON® and HYBRAR®.

The block copolymers of the conjugated diene and the vinyl aromatic compound can be grafted with an alpha,beta-unsaturated monocarboxylic or dicarboxylic acid reagent. The carboxylic acid reagents include carboxylic acids per se and their functional derivatives such as anhydrides, imides, metal salts, esters, etc., which are capable of being grafted onto the selectively hydrogenated block copolymer. The grafted polymer will usually contain from about 0.1 to about 20%, and preferably from about 0.1 to about 10% by weight based on the total weight of the block copolymer and the carboxylic acid reagent of the grafted carboxylic acid. Specific examples of useful monobasic carboxylic acids include acrylic acid, methacrylic acid, cinnamic acid, crotonic acid, acrylic anhydride, sodium acrylate, calcium acrylate and magnesium acrylate, etc. Examples of dicarboxylic acids and useful derivatives thereof include maleic acid, maleic anhydride, fumaric acid, mesaconic acid, itaconic acid, citraconic acid, itaconic anhydride, citraconic anhydride, monomethyl maleate, monosodium maleate, etc.

The first component containing a styrene and hydrocarbon block copolymer can be modified by adding an oil, such as a mineral oil, paraffinic oil, polybutene oil or the like. The

amount of oil added to the styrene and hydrocarbon block copolymer is from about 5% to about 40%. The first component can also contain a polypropylene up to about 20% by weight of the first component. One particularly suitable first component is an oil modified SEBS sold by the Shell Chemical Company under the product designation KRATON G2705.

The melt strength enhancing polymer preferably is a high melt strength polypropylene. Suitable high melt strength polypropylenes can be a homopolymer or a copolymer of polypropylene and can have free end long chain branching or not. In one preferred form of the invention, the high melt strength polypropylene will have a melt flow index within the range of 10 grams/10 min. to 800 grams/10 min., more preferably 10 grams/10 min. to 200 grams/10 min. or any range or combination of ranges therein. High melt strength polypropylenes are known to have free-end long chain branches of propylene units. Methods of preparing polypropylenes which exhibit a high melt strength characteristic have been described in U.S. Pat. Nos. 4,916,198; 5,047,485; and 5,605,936, which are incorporated herein by reference and made a part hereof. One such method includes irradiating a linear propylene polymer in an environment in which the active oxygen concentration is about 15% by volume with high energy ionization radiation at a dose of 1×10^4 megarads per minute for a period of time sufficient for a substantial amount of chain scission of the linear propylene polymer to occur but insufficient to cause the material to become gelatinous. The irradiation results in chain scission. The subsequent recombination of chain fragments results in the formation of new chains, as well as joining chain fragments to chains to form branches. This further results in the desired free-end long chain branched, high molecular weight, non-linear, propylene polymer material. Radiation is maintained until a significant amount of long chain branches form. The material is then treated to deactivate substantially all the free radicals present in the irradiated material.

High melt strength polypropylenes can also be obtained as described in U.S. Pat. No. 5,416,169, which is incorporated in its entirety herein by reference and made a part hereof, when a specified organic peroxide (di-2-ethylhexyl peroxydicarbonate) is reacted with a polypropylene under specified conditions, followed by melt-kneading. Such polypropylenes are linear, crystalline polypropylenes having a branching coefficient of substantially 1, and, therefore, has no free end long-chain branching and will have an intrinsic viscosity of from about 2.5 dl/g to 10 dl/g.

Suitable copolymers of propylene are obtained by polymerizing a propylene monomer with an α -olefin having from 2 to 20 carbons. In a more preferred form of the disclosure the propylene is copolymerized with ethylene in an amount by weight from about 1% to about 20%, more preferably from about 1% to about 10% and most preferably from 2% to about 5% by weight of the copolymer. The propylene and ethylene copolymers may be random or block copolymers. In a preferred form of the disclosure, the propylene copolymer is obtained using a single-site catalyst.

The components of the blend can be blended and extruded using standard techniques well known in the art. The film 312 will have a thickness of from about 3 mils to about 12 mils, more preferably from 5 mils to about 9 mils.

FIG. 14 shows a multiple layer film having a first layer 314 and a second layer 316. FIG. 14 shows the use of two layers but the present disclosure contemplates using more than two layers provided the above-mentioned material property requirements are met. The first layer 314 can be of the same polymer blend used to fabricate the monolayer structure and

in a more preferred form of the disclosure will define a seal layer for joining the film the cassette 160. The second layer 316 can be made from non-PVC containing materials and preferably is selected from polyolefins, polybutadienes, polyesters, polyester ethers, polyester elastomers, polyamides and the like and blends of the same. A tie layer or tie layers (not shown) may be required to adhere additional layers to the first layer 314.

Suitable polyolefins include homopolymers and copolymers obtained by polymerizing α -olefins containing from 2 to 20 carbon atoms, and more preferably from 2 to 10 carbons. Therefore, suitable polyolefins include polymers and copolymers of propylene, ethylene, butene-1, pentene-1, 4-methyl-1-pentene, hexene-1, heptene-1, octene-1, nonene-1 and decene-1. Most preferably the polyolefin is a homopolymer or copolymer of propylene or a homopolymer or copolymer of polyethylene.

Suitable homopolymers of polypropylene can have a stereochemistry of amorphous, isotactic, syndiotactic, atactic, hemiisotactic or stereoblock. In one preferred form of the disclosure the homopolymer of polypropylene is obtained using a single site catalyst.

It is also possible to use a blend of polypropylene and α -olefin copolymers wherein the propylene copolymers can vary by the number of carbons in the α -olefin. For example, the present disclosure contemplates blends of propylene and α -olefin copolymers wherein one copolymer has a 2 carbon α -olefin and another copolymer has a 4 carbon α -olefin. It is also possible to use any combination of α -olefins from 2 to 20 carbons and more preferably from 2 to 8 carbons. Accordingly, the present disclosure contemplates blends of propylene and α -olefin copolymers wherein a first and second α -olefins have the following combination of carbon numbers: 2 and 6, 2 and 8, 4 and 6, 4 and 8. It is also contemplated using more than 2 polypropylene and α -olefin copolymers in the blend. Suitable polymers can be obtained using a catalytic procedure.

It may also be desirable to use a high melt strength polypropylene as defined above.

Suitable homopolymers of ethylene include those having a density of greater than 0.915 g/cc and includes low density polyethylene (LDPE), medium density polyethylene (MDPE) and high density polyethylene (HDPE).

Suitable copolymers of ethylene are obtained by polymerizing ethylene monomers with an α -olefin having from 3 to 20 carbons, more preferably 3-10 carbons and most preferably from 4 to 8 carbons. It is also desirable for the copolymers of ethylene to have a density as measured by ASTM D-792 of less than about 0.915 g/cc and more preferably less than about 0.910 g/cc and even more preferably less than about 0.900 g/cc. Such polymers are oftentimes referred to as VLDPE (very low density polyethylene) or ULDPPE (ultra low density polyethylene). Preferably the ethylene α -olefin copolymers are produced using a single site catalyst and even more preferably a metallocene catalyst systems. Single site catalysts are believed to have a single, sterically and electronically equivalent catalyst position as opposed to the Ziegler-Natta type catalysts which are known to have a mixture of catalyst sites. Such single-site catalyzed ethylene α -olefins are sold by Dow under the trade name AFFINITY, DuPont Dow under the trademark ENGAGE® and by Exxon under the trade name EXACT. These copolymers shall sometimes be referred to herein as m-ULDPE.

Suitable copolymers of ethylene also include ethylene and lower alkyl acrylate copolymers, ethylene and lower alkyl substituted alkyl acrylate copolymers and ethylene vinyl acetate copolymers having a vinyl acetate content of from

US 7,815,595 B2

27

about 5% to about 40% by weight of the copolymer. The term "lower alkyl acrylates" refers to comonomers having the formula set forth in Diagram 1:




Diagram 1

The R group refers to alkyls having from 1 to 17 carbons. Thus, the term "lower alkyl acrylates" includes but is not limited to methyl acrylate, ethyl acrylate, butyl acrylate and the like.

The term "alkyl substituted alkyl acrylates" refers to comonomers having the formula set forth in Diagram 2:




Diagram 2

R₁ and R₂ are alkyls having 1-17 carbons and can have the same number of carbons or have a different number of carbons. Thus, the term "alkyl substituted alkyl acrylates" includes but is not limited to methyl methacrylate, ethyl methacrylate, methyl ethacrylate, ethyl ethacrylate, butyl methacrylate, butyl ethacrylate and the like.

Suitable polybutadienes include the 1,2- and 1,4-addition products of 1,3-butadiene (these shall collectively be referred to as polybutadienes). In a more preferred form of the disclosure the polymer is a 1,2-addition product of 1,3 butadiene (these shall be referred to as 1,2 polybutadienes). In an even more preferred form of the disclosure the polymer of interest is a syndiotactic 1,2-polybutadiene and even more preferably a low crystallinity, syndiotactic 1,2 polybutadiene. In a preferred form of the disclosure the low crystallinity, syndiotactic 1,2 polybutadiene will have a crystallinity less than 50%, more preferably less than about 45%, even more preferably less than about 40%, even more preferably the crystallinity will be from about 13% to about 40%, and most preferably from about 15% to about 30%. In a preferred form of the disclosure the low crystallinity, syndiotactic 1,2 polybutadiene will have a melting point temperature measured in accordance with ASTM D 3418 from about 70° C. to about 120° C. Suitable resins include those sold by JSR (Japan Synthetic Rubber) under the grade designations: JSR RB 810, JSR RB 820, and JSR RB 830.

Suitable polyesters include polycondensation products of di- or polycarboxylic acids and di or poly hydroxy alcohols or alkylene oxides. In a preferred form of the disclosure the polyester is a polyester ether. Suitable polyester ethers are obtained from reacting 1,4 cyclohexane dimethanol, 1,4 cyclohexane dicarboxylic acid and polytetramethylene glycol ether and shall be referred to generally as PCCE. Suitable PCCE's are sold by Eastman under the trade name ECDEL. Suitable polyesters further include polyester elastomers which are block copolymers of a hard crystalline segment of polybutylene terephthalate and a second segment of a soft (amorphous) polyether glycols. Such polyester elastomers are sold by Du Pont Chemical Company under the trade name HYTREL®.

28

Suitable polyamides include those that result from a ring-opening reaction of lactams having from 4-12 carbons. This group of polyamides therefore includes nylon 6, nylon 10 and nylon 12. Acceptable polyamides also include aliphatic polyamides resulting from the condensation reaction of diamines having a carbon number within a range of 2-13, aliphatic polyamides resulting from a condensation reaction of di-acids having a carbon number within a range of 2-13, polyamides resulting from the condensation reaction of dimer fatty acids, and amide containing copolymers. Thus, suitable aliphatic polyamides include, for example, nylon 66, nylon 6,10 and dimer fatty acid polyamides.

In a preferred form of the disclosure, the cassette **160** is fabricated from a material that is adhesively compatible with the upper and lower membrane **162**, **164**. What is meant by adhesive compatibility is the membrane can be attached to the cassette using standard heat sealing techniques. One particularly suitable material is a polymer blend of a polyolefin and a styrene and hydrocarbon copolymer. More particularly, the polyolefin of the polymer blend is a polypropylene and even more preferably a polypropylene copolymer with ethylene with an ethylene content of from about 1% to about 6% by weight of the copolymer. The styrene and hydrocarbon copolymer is more preferably an SEBS tri-block copolymer as defined above. The polypropylene copolymer should constitute from about 70% to about 95% and more preferably from about 80% to about 90% of the blend, and the SEBS will constitute from about 5% to about 30% and more preferably from about 10% to about 20% SEBS. In a preferred form of the disclosure, the polypropylene used to fabricate the cassette will have a lower melting point temperature than the high melt strength polypropylene used to fabricate the membrane. In a preferred form of the disclosure the polypropylene of the cassette **160** will have a melting point temperature of from about 120° C.-140° C. and for the film from about 145° C.-160° C. The cassette **160** can be injection molded from these polymer blends.

The upper and lower membranes **162**, **164** are attached to the cassette **160** utilizing heat sealing techniques. The film has a peel strength of greater than 5.0 lbf/inch when tested with a tensile instrument until film failure or bond failure. Also, when the film is attached to the cassette it can be deformed under a pressure of 5 psi. The film maintains its low modulus and deformability properties even after sterilization to continue to meet the pumping requirement. The film has an extended shelf life. The film retains its pumping abilities even after two years shelf storage.

IV. Valve Actuator

Referring now to FIG. **15**, one embodiment of an interface between the valve actuator **26** and the valve manifold **190** is illustrated. The valve motor **28** (not illustrated) of the valve actuator **26** drives a camshaft **200** through a mechanical linkage determinable to those of skill in the art. In an embodiment, a single camshaft **200** attaches to a series of cams **202**, for example, one of each of the valves in the system **10** or **100**. The cams **202** are fixed to the camshaft **200** and rotate in a one to one relationship with same.

The cams **202** drive pistons **204**, which engage in a friction reduced way with the cams, for example, via rollers **206**. The cams **202** drive pistons **204** up and down (only two of five cams shown having associated pistons to show other features of the actuator **26**). When a cam **202** drives its associated piston **204** upward, the piston **204** engages one of the membranes **162** or **164** (typically the lower membrane **164**, which is not shown in FIG. **15** for clarity) and pushes the membrane

US 7,815,595 B2

29

up into the respective hole 192 defined by the rigid manifold 190. This action stops the flow of medical fluid or dialysate through the respective valve.

The pistons 204 are also spring-loaded inside a respective housing 208. When the camshaft 200 turns so that a lower cam profile appears below one of the pistons 204, the spring inside the housing 208 pushes the piston 204 so that the roller 206 maintains contact with the respective cam 202. The piston 204 consequently moves away from the respective hole 192 defined by the rigid manifold 190, wherein the membrane 162 or 164, which has been stretched upward by the piston 204, springs back to its normal shape. This action starts the flow of medical fluid or dialysate through the respective valve.

The motor 28 is of a type, for example a stepper or servo motor, that can rotate a fraction of a rotation and stop and dwell for any predetermined period of time. Thus, the motor 28 can hold a valve open or closed for as long as necessary. The cams 202 are shaped to provide a unique combination of bumps and valleys for every flow situation. In certain situations, such as with valves V2 and V3 of the system 10, the valves always open and close together, so that both valves use the same cam 202 oriented in the same way on camshaft 200.

Referring now to FIGS. 16A and 16B, the camshaft 200 and cams 202 are illustrated figuratively. FIG. 16A illustrates a composite cam profile 370, i.e., a combination of each of the cams 202a to 202f illustrated in FIG. 16B. FIG. 16B illustrates that the cams 202a to 202f mount to the camshaft 200 via hubs 384. The hubs 384 may employ set screws as is well known camshaft 200 can also have indentations, etc., for aligning the hubs 384. In an alternative embodiment, one or more of the cams 202a to 202f may be integrally formed with the otherwise camshaft 200. In an embodiment, the camshaft 200 is a single molded piece, which prevents the cams 202a to 202f from rotating with respect to one another. The single molded camshaft 200 supports or attaches to a plurality of or to all of the cams 202a to 202f.

As illustrated above in FIG. 15, each of the cams 202a to 202f of FIG. 16B drives a single piston 204 and roller 206 to operate a single valve head 192 of the rigid manifold 190. The cams 202a to 202f open or occlude the valve heads 192 according to the shape of the respective cam. FIG. 16B illustrates that the camshaft 200 supports six cams 202a to 202f. FIG. 15 illustrates five cams 200. The cam provided in the embodiment of FIG. 16B may be to open a last bag, illustrated by the "last bag valve open" position 382. Either of the systems 10 or 100 may include a last bag. The last bag is a final dialysate fill of about two liters into the patient before the patient disconnects from the system and resumes normal daily activities.

The valve motor 28 and the valve actuator 26 (FIGS. 1 and 2) rotate the camshaft 200 to open or close the valve heads 192 to create a desired solution flow path. The arrangement of the cams 202a to 202f on the camshaft 200 is made such that, at any time during the therapy, there is no more than one fluid path open at any given time. Further, when the valve actuator 26 rotates the camshaft 200 from one flow path open position to the next, the series of cams 202a to 202f close all the valves for a moment of time. The closing of each of the valves prevents dialysate from back-flowing or moving in the wrong direction. Still further, the cams 202a to 202f are arranged such that only one valve head 192 of the valve manifold 190 of the disposable unit 160 may be open at any given time. Therefore, there is no open fluid path in the event of a system failure or inadvertent power down. This safety feature prevents dialysate from free-flowing into the patient 12 or overfilling the patient 12.

30

The lid 116 for the housing 112 of the hardware unit 110 may be freely opened by an operator or patient to load the disposable unit 160 into the hardware unit. When this occurs, the controller 30 automatically commands the camshaft to rotate so that an "all valves open" position 372, illustrated by the composite profile 370, resides beneath the rollers 206 and pistons 204. In the "all valves open" position 372, the camshaft 200 is rotated such that a depression exists under each of the pistons 204 and associated rollers 206. Accordingly, the pistons 204 sit in a relatively low position, i.e., out of the way, when the operator or patient loads the disposable unit 160 and valve manifold 190 into the hardware unit 110. This enables the patient or operator to place a disposable unit 160 into the unit 110 without encountering an obstruction or opposing force by one or more of the pistons 206.

After the patient or operator loads the disposable unit into the hardware unit 110 and closes the lid 116, the controller 30 automatically rotates the camshaft 200 so that an "all valves closed" position 386a resides beneath the pistons 204 and rollers 206. As illustrated, the "all valves closed" position 386a resides adjacent to the "all valves open" position 372. When the camshaft 200 is rotated to the "all valves closed position" 386a, no fluid can flow through the system 10, 100. As the camshaft 200 rotates from the "all valves open" position 372 to the first "all valves closed" position 386a, a mechanical interlock (not illustrated) is moved into the camshaft 200, which prevents the rotation of the camshaft 200 back to the "all valves open" position 372. This prevents uncontrolled flow of the dialysate, which could occur when each of the valve heads 192 is open, in the event that the operator tries to open the lid 116 during therapy.

In an alternative embodiment, an interlock can be provided through software. An encoder provides positional and velocity feedback to the controller 30. The controller 30 therefore knows the position of the cam shaft 200. Thus, the controller 30 is able to prevent the rotation of the camshaft 200 back to the "all valves open" position 372.

When the patient closes lid 116, a second mechanical interlock (not illustrated) locks the lid in place, so that the patient cannot open the lid 116 during therapy. The system 10, 100 senses when the patient has removed the patient fluid line 292 and connector 290 from the transfer set, which is implanted in the patient 12. Only then will the system 10, 100 allow the patient to open the lid 116. The mechanical interlocks prevent free-filling, overfilling and the patient from tampering with the system while it is running. The valve configuration provides a fail safe system that prevents fluid flow in the event of a failure or power down.

In many instances, when the patient begins dialysis therapy, the patient is already full of dialysate. In the illustrated embodiment of FIG. 16A, therefore, the composite profile 370 provides the "all valve open" position 372 next to the "from patient valve open" position 374. The "from patient valve open" position resides next to the "drain valve open" position 376. In this manner, upon therapy startup, camshaft 200 is readily positioned to be able to cooperate with the pump 20, 120 to drain spent dialysate from the patient. It should be appreciated that any of the cams 202a to 202f may be the cam that provides the "from patient valve open" position 374, the "drain valve open" position 376, etc.

Between the "from patient valve open" position 374 and the "drain valve open" position 376 resides a second "all valves closed" position 386b. Between each opening of a new valve and closing of a previously opened valve, each valve is momentarily closed. The controller 30 causes the motor (e.g., a stepper, servo or DC motor) and actuator 26 to toggle the camshaft 200 back and forth between the "from patient valve

US 7,815,595 B2

31

open" position **374**, past the "all valves closed" position **386b**, to the "drain valve open" position **376**. In this manner, the pump **20, 120** is able to sequentially pull apart fluid from the patient **12** and dump it to drain **18**.

When the system **10, 100** completes the initial patient drain cycle, the controller **30** causes the actuator **26** of motor **28** to rotate camshaft **200** past the "all valves closed" position **386c** to the "supply valve open position" **378**. To fill the patient full of fresh dialysate, the controller **30** causes the camshaft **200** to toggle back and forth between the "supply valve open" position **378** and the "to patient valve open" position **380**, each time passing over the "all valves closed" position **386d**. Again, for the drain and fill cycles, only one valve head **192** is open at any given period of time. The toggling always includes an "all valves closed" position between the dosing of one valve head **192** and the opening of another. The single pump sequentially pulls fluid into the disposable unit **160** and pushes fluid from same.

After the initial fill, camshaft **200** is positioned so that the camshaft **200** can once again toggle back and forth between the "from patient valve open" position **374**, past the intermediate "all valves closed" position **386b**, to the "drain valve open" position **376**. When the patient is once again empty, the camshaft **200** is positioned so that the camshaft may be toggled back and forth between the "supply valve open" position **378** and the "to patient valve open" position **380**. The system **10, 100** repeats this series of cycles as many times as necessary. Typically, the patient receives approximately 2 to 2.5 liters of dialysate in a single fill cycle. The two supply bags **14** each hold six liters of dialysate in an embodiment. This provides the system **10, 100** with four to six complete fill, dwell and drain cycles, which are provided, for example, through the night while the patient sleeps.

In many instances, the patient will receive a last bag fill at the end of the therapy, which the patient will carry for the day. To perform this procedure, the camshaft **200** toggles back and forth between the "from patient valve open" position **374** to the "drain valve open" position **376** to dump the preceding fill of peritoneal fluid to drain **18**. Thereafter, the camshaft **200** is positioned to toggle back and forth between the "last bag valve open" position **382** and the "to patient valve open" position **380**. In doing so, the camshaft **200** rotates past one of the all valves closed positions, namely, the "all valves closed position" **386e**.

To prime the system, the camshaft **200** may be positioned and toggled in a number of different ways. In one embodiment, the camshaft **200** toggles back and forth between the "supply valve open" position **378** and the "drain valve open" position **376**, passing over the "all valves closed" position **386c**. This toggling in cooperation with the pumping of pump **20** or **120** causes the dialysate to flow from the supply bags **14**, through the disposable unit **160**, to drain **18**. In another embodiment, using the vented tip protector **280** illustrated in connection with the FIGS. **8** to **12**, the camshaft **200** toggles back and forth between the "supply valve open" position **378** and the "to patient valve open" position **380**. This causes dialysate to flow from the bags **14**, through the disposable unit **160**, and into the patient fluid line **292** to the end of the vented tip protector **280**. When dialysate reaches the hydrophobic membrane **300** of the vented tip protection **28**, the pressure in the system **10, 100** rises, wherein a signal is received by the

32

controller **30**, which causes the pump **20, 120** to stop pumping and the camshaft **200** to stop toggling.

V. Medical Fluid Pump

A. Pump Hardware and Operation

Referring now to FIGS. **17A** and **17B**, one embodiment of the pump **20** is illustrated. The lid **116** of the hardware unit **110** defines an upper chamber wall **216**. Disposed within the housing **112** of the hardware unit **110** (FIGS. **3A** to **4B**) is a lower chamber wall **218**. The chamber walls **216** and **218** define an internal chamber **210**. The chamber **210** can have any desired shape, for instance the clamshell shape as illustrated in FIGS. **17A** and **17B**.

The lower chamber wall **218** defines or provides a sealed aperture **219** that allows a pump piston **212** to translate back and forth within the chamber **210**. The piston **212** is attached to or integrally formed with a piston head **214**. The piston head **214** in an embodiment has an outer shape that is similar to or the same as an internal shape of the upper chamber wall **216**.

The pump piston **212** connects to or is integrally formed with the linear actuator **24**. The linear actuator **24** in an embodiment is a device, such as a ball screw, that converts the rotary motion of a motor **22** into the translational motion of the piston **212**. In one preferred embodiment, the motor **22** is a linear stepper motor that outputs a translationally moving shaft. Here, the actuator **24** may simply couple the motor shaft to the piston **212**. The linear or rotary stepper motor provides quiet linear motion and a very high positional resolution, accuracy and repeatability. Stepper motors are commercially available, for example, from Hayden Switch and Instrument Inc., Waterbury, Conn.

As described above, the flexible fluid receptacle **172** (seen in FIG. **17A** but not in FIG. **17B**) is defined by the expandable upper and lower membranes **162** and **164**, respectively, of the disposable unit **160**. In FIG. **17A**, when the pump **20** is full of medical fluid, the pump chamber **210** and the membrane receptacle **172** have substantially the same shape. In FIG. **17B**, when the pump **20** has displaced all or most all of the medical fluid, the pump chamber **210** maintains the same volume but the membranes **162** and **164** of the fluid receptacle **172** have collapsed to virtually a zero volume along the interior surface of the upper chamber wall **216**.

Vacuum source **44** for the pump **20** is described above in connection with FIG. **1**. The vacuum source **44** exerts a vacuum on the upper membrane **162**, through the aperture or port **222**. The aperture or port **222** extends through the upper chamber wall **216**. The vacuum source **44** exerts a vacuum on the lower membrane **164**, through an aperture **221** defined or provided by housing **223**, and through the port or aperture **220**. The port or aperture **220** extends through the piston **212**, including the piston head **214**. When a vacuum is applied, the lower membrane **164** seals against the piston head **214**. The upper membrane **162** seals against the upper chamber wall **216**.

The port **222** fluidly connects to channels (not illustrated) defined by the interior wall of the upper chamber wall **216**. The channels extend radially outwardly from port **222** in various directions. The channels help to distribute the negative pressure applied through the port **222** to further enable the upper membrane **162** to substantially conform to the interior shape of the upper chamber wall **216**. In a similar manner, the outer surface of the piston head **214** can include radially extending channels to further enable the lower mem-

US 7,815,595 B2

33

brane 164 to substantially conform, upon application of the vacuum, to the outer surface of the piston head 214.

The pump 20 also includes a diaphragm 232 tensioned between the upper and lower chamber walls 216 and 218, respectively. The diaphragm 232 defines, together with the upper chamber wall 218, a known, predictable and repeatable maximum volume of dialysate, which can be drawn from one or more of the supply bags 14 and transported to the patient 12. The diaphragm 232 also enables the volume of a partial stroke to be characterized, which also enables accurate and repeatable volume measurements.

The diaphragm 232 is disposed beneath the piston head 214 and around the piston 212. When the vacuum is applied to the port or aperture 220, the diaphragm 232, as well as the lower membrane 164, are pulled against the piston head 214. When the piston head 214 is actuated upwardly away from the lower chamber wall 218, with the vacuum applied through aperture 220, the membrane 164 and the diaphragm 232 remain drawn to the piston head 214. An inner portion of the membrane 164 conforms to the shape of the outer surface of the piston head 214. The remaining outer portion of the membrane 164 conforms to the shape of the exposed surface of the diaphragm 232.

The diaphragm 232 in an embodiment includes a flexible, molded cup-shaped elastomer and a fabric reinforcement, such as fabric reinforced ethylene propylene diene methylene ("EPDM"). The fabric can be integrally molded with the elastomer. The fabric prevents unwanted deformation of the diaphragm while under pressure. The diaphragm 232 can stretch when the piston 212 and head 214 move downwardly towards the lower chamber wall 218, pulling the diaphragm 232 along the crimped edges of the upper and lower chamber walls 216 and 218. The diaphragm 232 also moves and remains sealed to the piston head 214 when the piston 212 and head 214 move upwardly towards the upper chamber wall 216.

In operating the pump 20, negative pressure is constantly applied through the port 222 to hold the upper membrane 162 against the upper chamber wall 216. The manifold 190 of the disposable unit 160 (see FIGS. 3A and 5) define a fluid port opening 230 to the membrane receptacle 172. The fluid port opening 230 allows medical fluid or dialysate to enter and exit the membrane receptacle 172. The membrane receptacle 172 seats in place with the crimped edges of the upper and lower chamber walls 216 and 218. The seal 170 of the receptacle 172 may actually reside slightly inside the crimped edges of the upper and lower chamber walls 216 and 218 (see FIG. 4A).

During a pump fill stroke, with the upper membrane 162 vacuum-pressed against the upper chamber wall 216, and the lower membrane 164 and the diaphragm 232 vacuum-pressed against the piston head 214, the motor 22/actuator 24 cause the piston head 214 to move downwardly towards the lower chamber wall 218, increasing the volume within the flexible receptacle 172, and producing a negative pressure within same. The negative pressure pulls dialysate from the supply bags 14 or the patient 12 as dictated by the current valve arrangement. The opened receptacle 172 fills with fluid. This process occurs when the pump moves from the position of FIG. 17B to the position of FIG. 17A. FIG. 17A shows the pump 20 at the end of the stroke, with the receptacle 172 fully opened (i.e., full of fluid).

During a patient fill or drain stroke, again with the upper membrane 162 vacuum-pressed against the upper chamber wall 216, and the lower membrane 164 and the diaphragm 232 vacuum-pressed against the piston head 214, the motor 22/actuator 24 cause the piston head 214 to move upwardly

34

towards the upper chamber wall 216, decreasing the volume within the flexible receptacle 172 and producing a positive pressure within same. The positive pressure pushes dialysate from the receptacle 172 to the patient 12 or the drain 18 as dictated by the current valve arrangement. The receptacle 172 closes as the lower membrane 164 moves upward towards the upper membrane 162. This process occurs when the pump moves from the position of FIG. 17A to the position of FIG. 17B. FIG. 17B shows the pump 20 at the end of the stroke, with the receptacle 172 empty or virtually empty.

In the event that air ("air" for purposes of this disclosure includes air as well as other gases which may be present, particularly those that have escaped from the patient's peritoneal cavity) enters the fluid receptacle 172, it must be purged to maintain accuracy. It should be appreciated that if air enters between the membranes 162 and 164, system 10, 100 does not have the ability to pull a vacuum between the membranes 162 and 164. The elasticity of the membranes 162 and 164, however, naturally tend to purge air therefrom. In an alternative embodiment the system 10, 100 can be adapted to provide a vacuum source that pulls a vacuum between the membranes 162 and 164 to purge air therefrom.

To purge air from between the membranes, the system 10, 100 also provides a positive pressure source. In systems 10, 100, for example, the pump motor 46 can be used in reverse of normal operation and, instead of producing vacuum source 44 (FIGS. 1 and 2), produce a positive pressure. The system 10 applies a positive pressure through the aperture or port 222 in the upper chamber wall 216 when air is detected between the membranes 162 and 164 or elsewhere in the disposable unit 160 or tubing. In one purge procedure, the controller 30 causes the motor 22/actuator 24 to move the piston head 214 to approximately a halfway point in either the positive or negative strokes. With the upper membrane 162 vacuum-pressed against the upper chamber wall 216, and the lower membrane 164 and the diaphragm 232 vacuum-pressed against the piston head 214 maintained at the halfway point, the controller causes the negative pressure source in through the aperture 222 to change to a positive pressure source, which pushes the upper membrane 162 conformingly against the lower membrane 164, which is supported by the piston head 214 and the diaphragm 232. Any air or fluid residing in the receptacle 172 is purged to drain as is any air between the receptacle 172 and drain 18.

B. Capacitance Volume Sensor

FIGS. 17A and 17B also illustrate that the pump 20 cooperates with an embodiment of the capacitance fluid volume sensor 60 of the system 10. One embodiment of a capacitance sensor 60 is disclosed in greater detail in the patent application entitled, "Capacitance Fluid Volume Measurement," Ser. No. 10/054,487, filed on Jan. 22, 2002, incorporated herein by reference. The capacitance sensor 60 uses capacitance measurement techniques to determine the volume of a fluid inside of a chamber. As the volume of the fluid changes, a sensed voltage that is proportional to the change in capacitance changes. Therefore, the sensor 60 can determine whether the chamber is, for example, empty, an eighth full, quarter full, half full, full, or any other percent full. Each of these measurements can be made accurately, for example, at least on the order of the accuracy achieved by known gravimetric scales or pressure/volume measurements. The present disclosure, however, is simpler, non-invasive, inexpensive and does not require the medical operation to be a batch operation.

US 7,815,595 B2

35

Generally, the capacitance C between two capacitor plates changes according to the function $C=k \times (S/d)$, wherein k is the dielectric constant, S is the surface area of the individual plates and d is the distance between the plates. The capacitance between the plates changes proportionally according to the function $1/(R \times V)$, wherein R is a known resistance and V is the voltage measured across the capacitor plates.

The dielectric constant k of medical fluid or dialysate is much higher than that of air, which typically fills the pump chamber **210** when the piston head **214** is bottomed out against the upper chamber wall **216**, as illustrated in FIG. **17B**. Therefore, the varying distance, Δd , of the low dielectric displacement fluid between the expanding and contracting receptacle **172** and the lower chamber wall **218** may have some effect on the capacitance between ground capacitance plate **224** and the active capacitance plate **226**. Likewise the surface area, S , of the capacitance plates and the moving membrane **164** may have some effect on the capacitance. Certainly, the changing overall dielectric from the high dielectric dialysate replacing the low dielectric air (or vice versa) affects the overall capacitance between the plates **224** and **226**.

As the membranes **162** and **164** expand and fill with medical fluid, the overall capacitance changes, i.e., increases. The sensor **60** generates a high impedance potential across the grounded and active capacitor plates **224** and **226**. The high impedance potential is indicative of an amount of fluid in the receptacle **172**. If the potential does not change over time when it is expected to change, the sensor **60** can also indicate an amount or portion of air within the receptacle **172**.

A capacitance sensing circuit amplifies the high impedance signal to produce a low impedance potential. The low impedance potential is also fed back to the guard plate **228**, which protects the sensitive signal from being effected by outside electrical influences. The amplified potential is converted to a digital signal and fed to the processor **34**, where it is filtered and/or summed. The video monitor **40** can then be used to visually provide a volume and/or a flowrate indication to a patient or operator. Additionally, the processor **34** can use the summed outputs to control the pump **20** of the system **10**, for example, to terminate dialysate flow upon reaching predetermined overall volume.

Referring now to FIG. **18**, the pump **120** of the system **100** is illustrated in operation with the capacitance sensor **60** of the present disclosure. The pump **120** forms a clamshell with first and second portions **246** and **248**, which together form the pump chamber **250**. The portions **246** and **248** are rigid, fixed volume, disk shaped indentations in the base **114** and lid **116** of the hardware unit **110**. The clamshell first and second portions **246** and **248** are closed and sealed on the pump receptacle portion **172** of the disposable unit **110**, which includes the expandable membranes **162** and **164**.

An opening or aperture **252** is defined between the first and second clamshell portions **246** and **248** and the flexible membranes **162** and **164**. The opening **252** enables medical fluid, for example, dialysate, to enter and exit the chamber **250** between the membranes **162** and **164** in the receptacle portion **172**. The receptacle portion **172** fluidly communicates with the valve manifold **190**.

FIG. **18** shows the pump chamber **250** in an empty state with both membranes **162** and **164** in relaxed positions, so that the flexible receptacle portion **172** is closed. The empty volume state is achieved when the membranes **162** and **164** have collapsed so that substantially all the fluid is removed from the sterile receptacle **172** and likewise the pump chamber **250**.

36

The empty volume state can be achieved, for example, by allowing the elastic membranes **162**, **164** to return to their relaxed, unstressed state as shown in FIG. **18**. Also, both membranes **162** and **164** can be forced together against each other or against either one of the inside portions **246** and **248** of the pump chamber **250**. When the pump chamber **250** is in the full state, the medical fluid resides between the membranes **162** and **164**, wherein the membranes have been suctioned against the inner walls of portions **246** and **248**.

It should be appreciated that either one or both of the membranes **162** and **164** can be moved towards and away from the clamshell portions **246** and **248** by any suitable fluid activation device. In various embodiments, the diaphragm pump is pneumatically or hydraulically actuated.

The diaphragm pump **120** of the system **100** does not require a separate piston or mechanical actuator as does the pump **20** of the system **10**. The clamshell portions **246** and **248** define ports **254** and **256**, respectively, to allow for movement of a displacement fluid (for example, pneumatic or hydraulic fluid) into and out of the chamber areas outside of the receptacle **172** to operate the diaphragm pump.

In an embodiment, the medical fluid, for example, dialysate, is suctioned into the receptacle **172** in the chamber **250**. The receptacle **172**, defined by membranes **162** and **164**, may be filled with medical fluid by applying negative pressures to one or both of the chamber ports **254** and **256**. The medical fluid can be emptied from the receptacle **172** by applying a positive pressure to at least one of the ports **254** and **256**, or by allowing the membranes **162** and **164** to spring back into shape. In an alternative embodiment, the medical fluid, for example, dialysate, is pressurized from an external source to move in and out of the pump chamber **250** between the membranes **162** and **164**.

The clamshell portions **246** and **248** form and hold the capacitor plates of the capacitance sensor **60**. In an embodiment, upper clamshell portion **246** includes an active metal or otherwise conductive capacitance plate **258** between electrically insulative or plastic layers. A metal guard plate **260** is provided on the outer plastic layer of the upper clamshell portion **246**. The guard plate **260** provides noise protection for the high impedance signal that transmits from the active capacitor plate **258**.

As with the pump **20** of system **10**, the active capacitor plate **258** of upper clamshell portion **246** of the pump **120** of the system **100** electrically couples to a capacitance sensing circuit. The guard plate **260** likewise electrically couples to the feedback loop of the capacitance sensing circuit as described above.

In an embodiment, lower clamshell portion **248** is also made of an inert plastic, wherein a metal capacitor plate **262** attaches to the outer surface of the lower clamshell portion **248**. The metal capacitor plate **262** disposed on the outside of the clamshell portion **248** electrically couples to ground.

In one implementation, a negative pressure is constantly maintained at the lower port **256**, so that the lower membrane **164** is pulled to conform to the inner surface of the grounded clamshell portion **248** during a multitude of fill and empty cycles. In this implementation, the upper membrane **162** does the pumping work. That is, when a negative pressure is applied to upper port **254** of upper clamshell **246**, upper membrane **162** is suctioned up against and conforms with the inner surface of upper clamshell **246**. This action draws fluid from the supply bag **14**, through the manifold **190**, and into the receptacle **172**. To expel fluid, the negative pressure is released from upper port **254**, wherein upper membrane **162** collapses to push the fluid from the receptacle **172**. Alternatively, a positive pressure is applied through one or both ports.

In operation, the capacitance sensor **60** operates substantially as described in FIGS. **17A** and **17B**. The receptacle **172** expands between the portions **246** and **248**. A varying distance, Δd , of the low dielectric displacement fluid between the expanding and contracting receptacle **172** and the portions **246** and **248** may have some effect on the capacitance between the ground plate **262** and the active plate **258**. Likewise the surface area, S , defined by the ground and active capacitance plates and the expanding membranes may have some effect on the overall capacitance. Certainly, the changing overall dielectric from the high dielectric dialysate replacing the low dielectric air (or vice versa) affects the overall capacitance between the plates **258** and **262**.

As the membranes **162** and **164** expand and fill with medical fluid, the capacitance changes, i.e., increases. Each different amount of medical fluid within the chamber **250** has a unique overall capacitance. A unique capacitance value can therefore be associated with each specific fluid volume in the chamber, for example, substantially empty, partially full, or substantially full.

As an alternative to the capacitance volume sensor **60** described above, the volume of dialysate fluid flowing through the automated systems **10** and **100** can be determined using other methods, such as through an electronic balance. In such a case, the electronic balance keeps track of the amount of dialysate that is supplied to the system during a priming of the system. The electronic balance also monitors any additional dialysate added to the system during dialysis treatment.

In other alternative embodiments, any of the systems described herein can be sensed using other types of flowmeters or devices employing Boyle's Law, which are known to those of skill in the art. Further, various other types of fluid volume measurement or flowrate devices can be used with the automated systems **10** and **100**, such as orifice plates, mass flow meters or other flow measuring devices known to those of skill in the art.

VI. Precision Pressure Control

As discussed above, the system **10** employs a valve actuator **24** and a pump motor **22**. In one embodiment the pump motor **22** is a stepper motor. In another embodiment, the motor **22** may be a DC motor or other type of repeatable and accurately positionable motor. Each of these types of motors enable system **10** to position the piston **212** and piston head **214** very accurately within the pump chamber **210**. In the case of a high precision rotary motor **22**, the actuator **24** converts the rotary motion into a translation motion precisely and moves the piston **212** back and forth within the chamber **210** within the accuracy and repeatability requirement of the system. The resolution of the linear stepper motor in an embodiment is about 0.00012 inches per step to about 0.00192 inches per step.

The pump motor **22** is also programmable. The programmable nature of the pump motor **22** enables acceleration, velocity and positional data to be entered into the controller **30**, wherein the controller **30** uses the information to position the piston **212** and piston head **214** within the pump chamber **210**, within an appropriate amount of time, to produce a desired amount of force or fluid pressure. The ability to preset the acceleration, velocity and position of the piston head **214** provides an advantage over purely pneumatic systems that respond relatively sluggishly to pneumatic signals.

The flexible nature of the PVC medical tubing described, e.g., in connection with FIG. **8** and the membrane material, described above in connection with FIGS. **13** and **14**, causes

the system **10** to have what is known as "compliance". Compliance is caused when the system **10** attempts to create fluid pressure, e.g., by moving the pump piston **212** and head **214**, but instead causes the flexible tubing and membranes to expand. With the flexible tubing and membranes, compliance is inevitable. Eventually, when the tubing and membranes have expanded to their elastic limit, the pressure in the pump chamber **210** (i.e., in the receptacle **172**) and throughout the tubing rises sharply. It is desirable to overcome the compliance of the tubing and membranes **162** and **164** as quickly as possible so that pressure may be built to drive the fluid.

The present invention uses a hybrid pressure control system that combines the ability to preset the pump piston acceleration and velocity with an adaptive pressure control scheme, which causes the pressure to achieve a desired pressure set point for any given stroke and causes the pressure to be fine tuned over time, i.e., over repeated strokes. That is, the present invention employs a method of controlling pressure within the system that seeks first to overcome system compliance and then seeks to achieve a desired pressure set point. The output of the present method of controlling pressure within the pump chamber **210** is illustrated by the velocity and pressure curves of FIG. **19**.

In general, the system **10** controls the pressure within the receptacle **172** in the pump chamber **210** by controlling the velocity of the piston **212** and piston head **214**. The velocity profile **390** of FIG. **19** illustrates a single pump stroke that occurs over a time "t" beginning at the start of stroke position **392**. In the beginning of the stroke, the velocity ramps up at a preset acceleration **394**. The preset acceleration **394** is programmed into the controller **30**. When the velocity due to the preset acceleration **394** reaches a max velocity **396**, the acceleration **394** changes to a zero acceleration, and the piston **212** moves at the constant max velocity **396**.

During the time period of the acceleration **394** and the max velocity **396**, which is designated by the dashed vertical line **398**, the corresponding pressure as illustrated by a pressure curve **401** of pressure profile **400**, ramps up beginning very slowly and exponentially increasing as the time reaches that of the dashed line **398**. In the initial portion of the pressure curve, i.e., just after the start of stroke position, the pressure builds slowly as the compliance in the system is taken up. As the compliance is taken up, the pressure builds at faster and faster rates.

When the pressure reaches a pressure proximity threshold **402**, set in software, the software within the controller **30** converts from the previous motion (acceleration, velocity, position) control to an adaptive control. It should therefore be appreciated that the method of controlling pressure within the fluid pump of the present disclosure is a hybrid type of control method, employing a combination of techniques.

The motion control portion, accented by the acceleration **394** and max velocity **396**, represents a period in time when the method of control is forcing the system to overcome the pressure compliance. Upon reaching the pressure proximity threshold **402**, the controller **30** causes the velocity to sharply decelerate at deceleration **404**. Deceleration **404** reduces the velocity of the piston **212** and piston head **214** to a velocity **406**, which is a velocity that aids in the ability of the adaptive control portion of the pressure control system to achieve a pressure set point **408**. That is, without the programmed deceleration **404**, the adaptive control portion would have a more difficult (i.e., longer) time controlling the velocity to make the pressure reach or substantially reach the pressure set point **408**.

As explained in more detail below, the acceleration **394** is adaptively controlled in an embodiment, so as to reduce the

US 7,815,595 B2

39

amount of initial overshoot. The adaptive control over the acceleration **394** is fine tuned over time to further reduce the amount of initial overshoot. Each of these measures affects the amount of controlled deceleration **404** needed.

After the controlled deceleration **404** reaches the velocity **406** and until the time of the second dashed line **410**, the system **10** operates in an adaptive mode. The second vertical line **410** occurs near the end of the stroke. As illustrated, the adaptive portion of the stroke is broken down into a number of areas, namely area **412** and area **414**. Area **412** is characterized by the overshoot or undershoot caused by the programmed acceleration **394**. In applying adaptive techniques, the adjustments or parameters that overcome area **414** error are tailored in software to combat overshoot or undershoot. The area **414** focuses on attempting to minimize the error between the actual pressure curve **401** and the pressure set point **408**. During the area **414**, the parameters and adaptive measures are tailored in software reduce the oscillation of the pressure curve **401** to achieve a pressure set point **408** as much as possible and as quickly as possible.

Upon reaching the time denoted by the dashed line **410**, the pressure control method once again resumes motion control and decelerates the velocity at a controlled and predetermined deceleration **416** down to a final travel velocity **418**, which is also the initial velocity at the start of the stroke **392**. In an alternative embodiment, the method can simply let the adaptive control continue past the time line **410** and attempt to achieve the final travel velocity **418**. After the time line **410**, the pressure along pressure curve **401** falls off towards zero pressure as illustrated by the pressure profile **400**. Comparing the pressure profile **400** to the velocity profile **390**, it should be appreciated that pressure remains in the receptacle **172** of the pump chamber **210** even after the stroke ends at time "t". In some cases, the pressure overshoots as the piston **212** suddenly stops, wherein the momentum of the liquid produces a pressure spike after time "t".

Referring now to FIG. 20, an algorithm **420** for employing the adaptive pressure control during the areas **412** and **414** of the pressure profile **400** is illustrated. In an embodiment, the adaptive control portion of the pressure control method employs a proportional, integral and derivative ("PIED") adaptive parameters. In the method, a pressure reading is taken from a pressure sensor which senses the pressure inside the receptacle **172** of the pump chamber **210**, and which provides a pressure sensor input **422** to the controller **30**, as illustrated by the algorithm **420**. Pressure sensor input **422** is sent through a digital filter **424**, producing a measured variable **426**. The measured variable **426** is compared with a desired variable, i.e., the pressure set point **408** illustrated in FIG. 19, wherein an error **428** is produced between the measured variable **426** and the desired pressure set point **408**.

Next, the error **428** is entered into a PID calculation **430**, which uses a proportional coefficient **432**, an integral coefficient **434** and a differential coefficient **436**. The output of the PID calculation **430** is an adaptive pressure change **438**. The controller **30** then changes the velocity up or down to produce the pressure change **438**.

In the pressure profile **400** of FIG. 19, the algorithm **420** of FIG. 20 is constantly being performed during the adaptive areas **412** and **414**. As discussed below, the corrective parameters, e.g., the coefficients **432**, **434** and **436**, are used differently during the areas **412** and **414** because correction in the area **412** is focused on minimizing overshoot and undershoot, while correction in the area **414** however is focused on reducing error to zero about the pressure set point **408**.

As described above, a single pump **20** is used in the system **10**. The single pump **20** provides positive pressure during the

40

patient fill stroke and the pump to drain stroke. The pump **20** also provides negative pressure during the pull from supply bag **14** stroke and the pull from patient **12** stroke. Of the four strokes, it is most important to accurately control the pressure during the patient fill and patient drain stroke. It is not as critical to control the pressure when pumping fluid from the supply bags **14** or when pumping fluid from the receptacle **172** of the pump chamber **210** to drain **18**. In the two positive pressure strokes, one stroke, namely the patient fill stroke, it is critical to properly control pressure. In the two negative pressure strokes, one of the strokes, namely the pull from patient stroke, it is critical to properly control pressure. In the other two strokes, pressure is controlled without taxing the controller, motor **22** and disposable unit **160** needlessly.

Referring now to FIG. 21, pressure and velocity curves are shown for a number of strokes during the patient fill cycle. The upper profile **440** shows the actual pressure **444** versus the desired pressure **442** in milli-pounds per square inch ("mPSI"). The lower profile **450** shows corresponding velocity curves. In the pressure profile **440**, the darkened line **442** corresponds to the desired pressure in mPSI. The curve **444** illustrates the actual pressure in mPSI. The curves **452a**, **452b** and **452c** in the velocity profile **450** illustrate the piston velocities that produce the pressure fluctuations along the pressure curve **444** of the pressure profile **440**. The velocity is measured in some increment of steps per second, such as milli-steps per second or micro steps per second when the motor **22** employed is a stepper motor. Different stepper motors for use in the present disclosure may be programmed in different increments of a step. The actual velocity is therefore a function of the resolution of the stepper motor.

At time zero, the desired pressure **442** changes virtually instantaneously to 2000 mPSI. The desired pressure curve **442** maintains this constant 2000 mPSI until reaching approximately 1.6 seconds, at which point the desired pressure **442** returns virtually instantaneously to zero. This step by the desired pressure curve **442** represents one complete patient fill stroke, wherein one full positive up-stroke of the piston **212** and piston head **214** within the pump chamber **220** occurs. In this step it is critical to control pressure because dialysate is being pumped into the patient's peritoneal cavity **12**. The actual pressure curve **444** ramps up exponentially and oscillates about the 2000 mPSI set point in the manner described in connection with FIG. 19. It should also be noted that the velocity curve **452a** follows a similar pattern to that shown in FIG. 19.

At about 1.6 seconds, i.e., when the piston head has reached the upper chamber **216** of the valve chamber **210**, controller **30** stops the piston **212** from moving. The velocity of the piston head remains at zero until approximately 3.4 seconds. In this period, the valves have all been closed via one of the "all valves closed" positions illustrated in connection with FIG. 16A. As illustrated by pressure curve **444**, residual fluid pressure resides within the pump chamber **210** even though the piston head **214** is not moving.

At about time 3.4 seconds, the desired pressure curve **442** switches virtuously instantaneously to -2000 MPsi. The pump **20** is now being asked to expand and form a negative pressure that pulls fluid from the supply bags **14**. During this stroke, it is not as critical to control pressure as accurately in the patient fill stroke. Accordingly, the method may be programmed to bypass the motion control portion of the pressure control method and simply adaptively seek to find the pressure set point along line **442**. Dialysate moves through the fluid heating path **180** of the disposable unit **160** (see FIGS. 3A and 5, etc.) during the patient fill stroke. Much of the compliance, i.e., stretching of the system occurs when the

US 7,815,595 B2

41

fluid passes through the path **180**. Pumping fluid from the supply bag **14**, however, does not require the fluid to pass through the heating path **180**. The system **10** does not therefore experience the same level of compliance during this stroke. It is possible to pump from the bags **14** without using the motion control portion illustrated in connection with FIG. **19**, since the lessened compliance may not require the “brute force” supplied by the controlled acceleration.

In FIG. **21**, the pump completes the stroke that pulls dialysate from the supply bag at about five seconds. The demand pressure along curve **442** returns to zero accordingly. Next, the valve switches to an all closed position, the controller **30** sets the piston speed to zero, and the piston head resides substantially along the lower chamber wall **218**, with the receptacle **172** full of fluid until approximately 6.8 seconds has passed, wherein the system **10** repeats the patient fill stroke as described previously.

Referring now to FIG. **22**, a pressure profile **452** and a velocity profile **460** are illustrated for the patient drain stroke and the pump to drain stroke of the patient drain cycle. In the pressure profile **452**, the demand pressure curve **454** illustrates that the controller calls for a negative 2500 MPsi to pull dialysate from the patient. The controller **30** calls for a positive pressure of 2500 MPsi to push fluid from the receptacle **172** of the pump chamber **210** to the drain bag **18**. In the velocity profile **460** shown below the pressure profile **452**, the actual velocity **462** in some increment of steps per second is illustrated. It should be appreciated that both velocity profiles **450** and **460** of FIGS. **21** and **22** are absolute velocities and do not illustrate that the pump piston **212** moves in positive and negative directions.

The actual pressure curve **456** of the profile **452** illustrates that the pressure is controlled to conform to the demand pressure line **454** more closely during the pull from patient portion than during the pump to drain portion of the profile **452**. In an embodiment, the controller **30** is programmed to provide a motion controlled velocity **464** for a portion of the pull from patient stroke and use an adaptive control during the time “tadapt”. The method also uses, in an embodiment, a controlled deceleration **466** at the end of the pull from patient stroke. Alternatively, the method allows the PID control to seek to find zero pressure. Similarly, during the pump to drain stroke, the controller **30** can switch to PID control only.

Referring now to FIG. **23**, one embodiment of an algorithm **470** illustrating the “fine tuning” adaptive control of the PID portion of the pressure control method of the present disclosure is illustrated. FIG. **23**, like FIG. **20**, includes a measured pressure variable **426** and a desirable pressure set point **408**. The pressure error **472** represents an error in either the overshoot area **412** or the oscillation area **414** illustrated in the pressure velocity profile **400** of FIG. **19**. For each area, the algorithm **470** looks at two error components, namely, the error **474** determined in the current stroke and the error **476** stored for previous strokes. The controller **30** compares the two errors **476** and **478** and makes a decision as illustrated in decision block **478**.

In the block **476**, if the current stroke error **474** is less than the previous stroke error **476**, the method uses the previous coefficient because the previous coefficient is currently having a desirable result. If the current stroke error **474** is greater than the previous strokes error **476**, two possibilities exist. First, the coefficient or corrective measure taken is not large enough to overcome the error increase. Here, the coefficient or corrective setting can be increased or another tactic may be employed. Second, the previous corrective procedure may be having an adverse impact, in which case the parameter connection can be reversed or another tactic can be employed.

42

Obviously, to employ algorithm **470**, the method provides that the controller **30** store the manner of the previous corrective attempts and outcomes of same. Based on what has happened previously, the controller decides to increment or decrease one or more of the parameters. The amount of increase or decrease is then applied to one or more coefficients stored in an increment table **480**. The adjusted or non-adjusted increment is then summed together with the currently used one or more coefficients **482** to form an adjusted one or more coefficients **484**.

Referring now to FIG. **24**, table **500** illustrates various different coefficients and adaptive perimeters for the pressure control method of the present disclosure. Certain of the coefficients and parameters apply more to the motion control portion of the profiles illustrated above, i.e., the set acceleration, deceleration and velocity portions of the profiles. The motion control parameters, however, effect the error, which influences the adaptive parameters in the PID portion of the pressure control. Other parameters apply to the adaptive control portions of the profiles. Adjusting the beginning stroke acceleration parameter **486** (illustrated by the acceleration **394** of the velocity profile **390** of FIG. **19**) affects the motion control portion of the present method. Acceleration as illustrated, affects overshoot and the efficient use of stroke time. That is, it is desirable to have a high acceleration to overcome compliance quickly, however, the cost may be that overshoot increases. On the other hand, a lower acceleration may reduce overshoot but require more time to overcome the compliance in the system.

The proximately threshold parameter **488** (illustrated by pressure line **402** in the pressure profile **400** of FIG. **19**) also affects overshoot and undershoot. Here, setting the pressure threshold **488** too low may cause undershoot, whereas setting the parameter **488** too high may cause overshoot. The DP/dt parameter **490** is the change in pressure for a given period of time. This parameter seeks to achieve, for example in FIG. **19**, a certain slope of the pressure curve **401**.

The maximum travel velocity parameter **492**, illustrated as line **396** in the velocity profile **390** of FIG. **19**, also affects overshoot and subsequent resonance. Another corrective factor is the conversion to pressure deceleration **494** corresponding to line **410** of FIG. **19**. The method includes running the system without changing back to motion control and instead leaving the system in the adaptive PID control. The conversion to deceleration can have a large impact on the residual pressure remaining in the pump chamber **210** after the valves close.

The PID factors Kp, Kd and Ki, labeled **496**, **498** and **502**, respectively, affect the adaptive control portion of the present method but also affect, to a lesser extent, the controlled deceleration at the end of the stroke. Each of the PID factors or parameters can be changed and adapted in mid-stroke. Also as illustrated in FIG. **23**, the factors can be changed so as to optimize the system over time.

Each of the above-described factors can be used to insulate the fluid pressure from changes in the environment outside of the system **10**. For example, the factors can overcome changes due to physiological and chemical changes in the patient’s abdomen. Also, the height of the patient supply bags **14** affects the initial loading of the fluid pump **20**. The parameters illustrated in FIG. **24** automatically overcome the changes due to bag height. Further, as the patient sleeps through the night, the supply bags **14** become less and less full, while the drain bag **18** becomes more full, both of which affect the pump pressure. The parameters illustrated in FIG. **24** are automatically adjustable to compensate for these changes and keep the system running smoothly.

US 7,815,595 B2

43

Certain of the above-described factors is changed more and used more during the overshoot area **412** illustrated in the pressure profile **400** of FIG. **19**. Other factors and parameters are used and changed more during the oscillation portion **414** of the profile **400**.

VII. In-Line Heater

In an embodiment, the inline heater **16** includes two electrical plate heaters, which are well known to those of skill in the art. The plate heaters of the heater **16** have a smooth and flat surface, which faces the disposable unit **160**. In an alternative embodiment, the automated systems **10** and **100** provide an in-line heater **16** having a plate heater in combination with an infrared heater or other convective heater.

In the alternative dual mode type heater, both the plate heater and, for example, the infrared heater are in-line heaters that heat the medical fluid that flows through the fluid heating path **180** of the disposable unit **160**. The radiant energy of the infrared heater is directed to and absorbed by the fluid in the fluid heating path **180**. The radiant energy or infrared heater in an embodiment is a primary or high capacity heater, which can heat a relatively large volume of cold fluid to a desired temperature in a short period of time.

The plate heater and the infrared heater of the dual mode heater embodiment of the heater **16** can be arranged in various configurations relative to each other. The dual mode heaters in an embodiment are arranged so that the fluid passes by the heaters sequentially (e.g., first the plate heater and then the radiant or infrared heater). In another embodiment, the fluid passes by the heaters simultaneously (both heaters at the same time). The fluid flow path past the heaters can be a common flow path for both heaters, such as in the fluid heating path **180**, or include independent flow paths for each heater.

The dual mode heater is particularly useful for quickly heating cool dialysate (high heat energy demand) supplied from one of the supply bags **14** to the automated system **10** or **100**. Initial system fills can be cooler than later fills, and the system can lose heat during the dwell phase. The temperature of the dialysate at initial system fill can therefore be quite low, such as 5° C. to 10° C. if the supply bags **14** are stored in cold ambient temperature.

The plate heater and the infrared heater of the dual mode heater embodiment of the heater **16** can be arranged in various configurations relative to each other. The dual mode heaters in an embodiment are arranged so that the fluid passes by the heaters sequentially (e.g., first the plate heater and then the radiant or infrared heater). In another embodiment, the fluid passes by the heaters simultaneously (both heaters at the same time). The fluid flow path past the heaters can be a common flow path for both heaters, such as in the fluid heating path **180** or include independent flow paths for each heater.

VIII. Fuzzy Logic for Heater Control

Similar to the controlling of the fluid pressure, the control of the plate heater **16** is also subject to a number of environmental variables. For example, the ambient temperature inside the patient's home affects the amount of heat that is needed to raise the temperature of the medical fluid to a desired temperature. Obviously, the temperature of the dialysate in the supply bags **14** affects the amount of heat that is needed to raise the fluid temperature to a desired temperature. Plate heater efficiency also affects the amount of heating needed. Further, the voltage provided by the patient's home is another factor. Typically, a doctor or caregiver prescribes the temperature of the dialysate for the patient to be controlled to

44

around a temperature of 37° C. It is, therefore, desirable to have a method of controlling the heater **16** to correct for outside temperature gradients so as to maintain the proper patient fluid temperature.

Referring now to FIG. **25**, one embodiment of a heating control method **510** is illustrated. The method **510** includes two separately performed algorithms **520** and **530** that operate in parallel to form an overall output **544**. The algorithm **520** is termed a "knowledge-based" control algorithm. The knowledge-based control algorithm is based on knowledge, such as empirical data, flow mechanics, laws of physics and lab data, etc.

The knowledge-based algorithm **520** requires a number of inputs as well as a number of constant settings. For example, the control algorithm **520** requires an input pulsatile flowrate. As illustrated below, the pulsatile flowrate is actually calculated from a number of input variables. The system **10**, **100** of the present disclosure provides fluid to the patient **12** in pulses, rather than on a continuous basis. It should be readily apparent from the discussion based on FIGS. **16A** and **16B**, that when all valve heads in the disposable are closed, no fluid can flow through the fluid heating pathway to the patient. The flowrate of fluid to the patient is therefore a pulsatile flowrate, wherein the patient receives the dialysate in spurts or pulses. It is difficult to control fluid temperature with this type of flowrate. To this end, the method **510** provides the dual algorithms **520** and **530**.

Besides the pulsatile flowrate, the knowledge-based control algorithm **520** also receives a measured, i.e., actual, fluid inlet temperature signal. Further, the algorithm **520** stores the plate heater efficiency, which is based on empirical data. In one embodiment, the upper and lower plates of the plate heater **16** are around 95% efficient. Algorithm **520** also inputs the total heater power, which is derived from the voltage input into the system **10**, **100**. Residential voltage may vary in a given day or over a period of days or from place to place.

The algorithm **520** also inputs the desired outlet fluid temperature, which is a constant setting but which may be modified by the patient's doctor or caregiver. As illustrated in FIG. **25**, the desired outlet fluid temperature is inputted into both the knowledge-based control algorithm **520** and the fuzzy logic based control algorithm **530**. As discussed in more detail below, the knowledge-based control algorithm **520** outputs a knowledge-based duty cycle into a summation point **544**.

With respect to the fuzzy logic-based control algorithm **530**, the desired fluid temperature is inputted into a comparison point **514**. The comparison point **514** outputs the difference between the desired fluid temperature and the actual measured fluid temperature exiting the heating system **548**. The fuzzy logic-based control algorithm **530** therefore receives a change in temperature ΔT as an input. As described below, the fuzzy logic-based control algorithm **530** employs the concepts and strategies of fuzzy logic control to output a fuzzy logic duty cycle.

The method **510** uses multiple temperature sensors, such as the sensors **62** illustrated in FIGS. **1** and **2**, which sense the temperature at different times within the method **510** and places within system **10**, **100**. One sensor senses the fluid outlet temperature, which feeds back from the heating system **548** to the comparison point **514**. Another two temperature sensors sense the temperature of the top plate and the bottom plate and feed back to the temperature limit controller **546**, located in software.

It should be appreciated that the weighting block **542** could alternatively be placed in the knowledge-based duty cycle output. As discussed below, however, the update rate of the

US 7,815,595 B2

45

fuzzy logic control loop is substantially higher than the update rate of the input signals entered into the knowledge-based control algorithm 520. It is therefore advantageous to weight the fuzzy logic-based duty cycle, as opposed to the knowledge-based duty cycle.

The weighted fuzzy logic-based duty cycle and the knowledge-based duty cycle are summed together at summing point 544 to produce an overall heater duty cycle. Duty cycle is one way to control the power input and, thus, the plate temperature of the heater. Controlling the duty cycle means controlling the percentage of a time period that full power is applied to the heater, for example, plate heater 16. In an alternative embodiment, the output of the parallel control algorithms 520 and 530 could be a percentage of full power applied at all times. Still further, the output of the parallel control algorithms 520 and 530 could be a percentage of full power applied for a percentage of a time period. For purposes of illustration, the method 510 is described using a duty cycle output which, as explained, is the percent of a time period that full power is applied to the heater.

As described herein, the heating system 548 (i.e., heater 16) in one embodiment is a plate heater, wherein upper and lower plates are disposed about a fluid heating path of the disposable unit 160. It should be appreciated, however, that the method 510 is equally applicable to the infrared heater previously described. Further, the method 510 is equally applicable to the combination of different types of heaters, such as, the combination of a plate heater and an infrared heater.

The method 510 uses multiple temperature sensors, such as the sensors 62 illustrated in FIGS. 1 and 2, which sense the temperature from different areas of the method 510. One sensor senses the fluid outlet temperature, which feeds back from the heating system 548 to the comparison point 514. Another two temperature sensors sense the temperature of the top plate and the bottom plate and feed back to the temperature limit controller 546, located in software.

As illustrated, before the summed heater duty cycle is inputted into the heating system 548, the system determines whether the top and bottom heating plates are already at a maximum allowable temperature. There exists a temperature above which it is not safe to maintain the plates of the plate heater. In a situation where one or both of the plates is currently at the temperature limit, the method 510 outputs a zero duty cycle, regardless of the calculations of the knowledge-based control system 520 and the fuzzy logic-based algorithm 530. To this end, the temperature of the top and bottom plates is fed back into the block 546, wherein the software only allows a heater duty cycle to be applied to the heating system 548 if the current temperature of the top and bottom plates is less than the temperature limit.

In an embodiment, if one of the plates is at the limit temperature, the method 510 provides a zero duty cycle to both plate heaters, even though one of the plate heaters may be below the temperature limit. Further, the software may be adapted so that if the actual temperature of the plate heater is very close to the limit temperature, the method 510 only allows the duty cycle be at or below a predetermined set point. In this manner, when the actual temperature is very near the limit temperature, the method 510 goes into a fault-type condition and uses a safe duty cycle.

Assuming the actual plate temperatures are below the safe temperature limit, the method 510 applies the combined heater duty cycle from the parallel control algorithms at summation point 544. The heater duty cycle applies full power for a certain percentage of a given amount of time. The given amount of time is the update speed of the fuzzy logic control

46

loop. The fuzzy logic control loop, including the fuzzy logic control algorithm 530, updates about nine times per second in one embodiment. It should be appreciated that the update rate of the fuzzy logic control loop is an important parameter and that simply increasing the update rate to a certain value may deteriorate the accuracy of the system. One range of update rates that provide good results is from about 8.5 times per second to about 9.5 times per second.

The update rate should not be evenly divisible into the frequency of the input power. For example, an update rate of nine times per second works when the AC frequency is held steady at 50 or 60 hertz. However, as is the case in some countries, the frequency may be 63 hertz. In such a case, an update rate of nine hertz will cause inaccuracy. Therefore, in one embodiment, an update rate of a fraction of 1 hertz is used, such as 9.1 hertz. Assuming the update rate to be nine times per second, the time per update is approximately 110 milliseconds. Therefore, if the duty cycle is 0.5, i.e., half on, half off, the time at which full power is applied is 55 milliseconds. During the other 55 milliseconds, no power is applied. If the duty cycle is 90%, then full power is applied for 90% of 110 milliseconds.

The update speed of the knowledge-based control algorithm 520 is not as critical as the update speed of the fuzzy logic control loop. For one reason, the signal inputs to the algorithm 520 change gradually over time so that they do not need to be checked as often as the comparison between the desired fluid temperature and the actual fluid temperature. An update rate of about two seconds is sufficient for the signal inputs. The inputs of the control algorithm 520 can be updated from about once every half second to about once every four seconds. The knowledge-based control algorithm 520 can run on the main processor of the system 10, 100, for example, an Intel StrongARM™ Processor. To facilitate the update rate of the fuzzy logic control loop, a high speed processor, such as a Motorola Digital Signal Processor is used. The fuzzy logic-based control algorithm 530 runs, in one embodiment, on a delegate processor, e.g., a Motorola Digital Processor.

Referring now to FIG. 26, the knowledge-based control algorithm 520 is illustrated in more detail. As discussed above, in a first step, the knowledge-based control algorithm receives a number of signal inputs, as indicated by block 522. Some of these inputs are updated at the main processor level of about once every two seconds. Other inputs are set in software as constants. One of the input signals that varies over time, is the number of stroke intervals ("N") per millisecond. The pump piston moves over a certain period of time, stops and dwells, and then moves again for a certain period of time. The pump makes N number of strokes per millisecond, which is inputted into the knowledge-based control algorithm.

Another input signal that varies over time is the input voltage ("Vac"). The input voltage Vac changes over time in a single house or in different locations. Another input signal that changes over time is the measured fluid inlet temperature ("Tin"). Fluid temperature Tin is measured by one of the numerous sensors of the method 510 described above. An input which will like not change over time is the plate heater efficiency ("E"). The heater efficiency E is determined empirically. The heater efficiency E could change depending upon the pressure inside the disposable unit during heating, the material of the disposable unit and the gap tolerance between the top and bottom plate. The heater efficiency E for a particular dialysis device therefore remains substantially constant. As described above, the desired fluid temperature ("Tdesired") may vary, depending on doctor's orders. However, for any given therapy session, Tdesired is a constant.

The knowledge-based control algorithm **520** also calculates the total heater power in Watts, as indicated by block **526**. In the illustrated embodiment, the method **510** calculates the heater power by dividing Vac2 by a plate heater resistance. The knowledge-based control algorithm **520** then uses the above calculations to calculate the knowledge-based duty cycle, as indicated by block **528**. The knowledge-based duty cycle equals, in one embodiment, a factor, e.g., of 0.07, multiplied by the temperature difference, ΔT , which equals T_{desired} minus the T_{in} . This product is then multiplied by the pulsatile flowrate Q. The latter product is then divided by the product of the total heater power W times the heater efficiency E. The knowledge-based duty cycle is then fed into summation point **544** in combination with the fuzzy logic-based duty cycle output as illustrated by FIG. 26.

The knowledge-based control algorithm **520** also calculates the total heater power in Watts, as indicated by block **526**. In the illustrated embodiment, the method **510** calculates the heater power by dividing Vac2 by a plate heater resistance. The knowledge-based control algorithm **520** then uses the above calculations to calculate the knowledge-based duty cycle, as indicated by block **528**. The knowledge-based duty cycle equals, in one embodiment, a factor, e.g., of 0.07, multiplied ΔT , which equals T_{desired} minus the T_{in} . This product is then multiplied by the pulsatile flowrate Q. The latter product is then divided by the product of the total heater power W times the heater efficiency E. The knowledge-based duty cycle is then fed into summation point **544** in combination with the fuzzy logic-based duty cycle output as illustrated by FIG. 26.

Referring now to FIG. 27, one embodiment for the fuzzy logic control algorithm **530** is illustrated. It should be appreciated that fuzzy logic is known generally to systems engineers and in the field of system and process control. The fuzzy logic algorithm described herein is merely one method of implementing fuzzy logic to perform the task of accepting an error input, which is the difference between the desired fluid temperature and the actual fluid temperature, and attempting to minimize this number to zero. Regardless of the method in which fuzzy logic is employed, the method inputs a temperature, sums a ΔT and outputs a power limiter, such as the duty cycle. The first step in the fuzzy logic control logic algorithm **530** is to therefore calculate the difference between T_{desired} and T_{in} , as indicated by block **532**.

Next, a number of membership functions are implemented, as indicated by block **534**. In this embodiment, the algorithm **530** implements five measurement functions. Two of the measurement functions, namely, nlarge and plarge, are trapezoidal membership functions. As is known in the art of fuzzy logic, the trapezoidal membership function consists of four nodes. Three other membership functions, namely nsmall, neutral and psmall, are set up as triangle membership functions, which consists of three nodes. After setting up the membership functions as indicated by block **534**, the fuzzy logic control algorithm **530** performs a fuzzification interface as indicated by block **536**. In the fuzzification interface, the control algorithm **530** converts the temperature difference ΔT , between T_{desired} and T_{in} to a number of fuzzy sets based on the membership functions set up as indicated in block **534**.

Next, the control algorithm **530** applies a number of fuzzy logic heating rules as indicated by block **538**. In an embodiment, the control algorithm **530** employs five fuzzy logic rules. One rules says that, if ΔT is nlarge, the output should decrease at a large pace. Another rules says that, if ΔT is nsmall, the output should decrease at a small pace. The third rule states that if ΔT is neutral, the output should be zero. A further rules states that if ΔT is psmall, the output should

increase at a small pace. The final rule states that if ΔT is plarge, the output should increase at a large pace.

The next step in the fuzzy logic control algorithm **530** is to perform a defuzzification interface, as indicated by block **540**. In the defuzzification interface, the output of the rules is converted to an actual or "crisp" output, which can then be translated into a duty cycle. In the defuzzification step indicated by block **590**, the output of the fuzzy logic rules is converted to a "crisp" or exact number. This number is then converted to the proper output for the heater which, in this embodiment, is the fuzzy heater duty cycle.

As indicated by block **542**, the next step is to determine how much weight to place on the fuzzy logic duty cycle with respect to the knowledge-based duty cycle. The weighting factor is decided by the fuzzy logic rules and the update rates of both the knowledge based and fuzzy logic based control algorithms. The weighted fuzzy logic duty cycle is then summed in summation point **544** with the knowledge-based duty cycle yielded by the knowledge-based control algorithm **520**.

IX. Electrical Insulation for the System

Medical equipment and in particular equipment in intimate contact with a patient needs to be properly electrically insulated against leakage currents. Class I type of equipment provides basic insulation and a means of connecting to a protective earthing conductor in the building in which the equipment resides, which dissipates hazardous voltages if the equipment insulation fails. One primary use for the system **10, 100** of the present disclosure however is in a patient's home. This presents two problems for Class I devices and in particular for dialysis machines. First, in many countries and older homes, the earthing ground is faulty, unreliable or completely absent. Second, many people bypass grounding systems that do exist. The present disclosure overcomes this problem by providing an automated dialysis system **10, 100** that requires no earth ground. The system **10, 100** does not simply rely on the basic insulation provided by Class I devices but provides either double insulation or reinforced insulation.

Double insulation includes two layers of insulation. One layer of insulation can be the basic insulation. At 240 VAC, basic insulation typically requires four millimeters of "creepage" or 2.5 millimeters of "air clearance". Creepage is the shortest distance between two conductive parts when both are disposed along a surface of insulation. Creepage is also the shortest distance between a conductive part and a bounding surface of a piece of equipment, wherein the conductive part and the equipment contact a piece of insulation. Air clearance is the shortest distance between two conductive parts or between a conductive part and a piece of equipment, measured through air.

The additional layer of insulation is called supplemental insulation. Supplemental insulation is independent insulation applied in addition to the basic insulation to ensure protection against electric shock if the basic insulation fails. The supplemental insulation can also be in the form of creepage and clearance.

Reinforced insulation, on the other hand, is a single layer of insulation offering the same degree of protection as double insulation. Reinforced insulation provides the electrical protection equivalent to double insulation for the rated voltage of the double insulation. For 240 VAC, used as the mains voltage of the system **10, 100**, the basic insulation can withstand 1500

US 7,815,595 B2

49

VAC and the supplemental insulation can withstand 2500 VAC. The single layer of reinforced insulation must therefore withstand at least 4000 VAC.

Referring now to FIG. 28, one embodiment of an electrically insulated system 550 of the present disclosure is illustrated. The system 550 is illustrated schematically, however, certain components of the system 550 are identifiable as components illustrated in the hardware drawings discussed above. For example, the system 550 includes the housing or enclosure 112, illustrated above in FIGS. 3A to 4B, which includes the base 114 and the lid 116 of the hardware unit 110. The system 550 also includes the heater 16, which in an embodiment includes upper and lower heating plates illustrated in FIG. 3A and discussed in connection with FIGS. 25 to 27. Further, the system 550 includes the display device 40 and temperature sensors 62 illustrated and discussed in connection with FIGS. 1 and 2.

In FIG. 28, the numbers in parenthesis indicate the working or operating voltage of the respective component. As illustrated, the line 552 and neutral 554 supply a mains voltage of 240 VAC, single phase, in an embodiment, which is the standard voltage used residentially in many countries throughout the world. The line 552 and neutral 554 could otherwise supply the United States residential standard of 120 VAC, single phase, and indeed could provide a voltage anywhere in the range of 90 to 260 VAC. The line 552 and neutral 554 feed the 240 VAC into a mains part 556. It is worth noting that the system 550 does not include or provide a protective earth conductor.

The mains part 556 feeds 240 VAC to a power supply printed circuit board ("PCB") 558. Power supply PCB 558 includes a mains part 562 and a live part 564. For purposes of the present disclosure, a "mains part" is the entirety of all parts of a piece of equipment intended to have a conductive connection with the supply mains voltage. A "live part" is any part that if a connection is made to the part, the part can cause a current exceeding the allowable leakage current for the part concerned to flow from that part to earth or from that part to an accessible part of the same equipment.

As illustrated, the live parts 560 and 564 step down in voltage from the mains parts 556 and 562, respectively, to 24 VDC. Obviously, the voltage may be stepped down to other desired levels. Live part 560 feeds live part 566. Live part 566 is an inverter having a step-up transformer that outputs a voltage of 1200 Vpeak. The inverter 566 powers a number of cathode fluorescent lights, which provide backlighting for the display device 40.

Live part 560 is also electrically isolated from applied part 568, which is maintained at a zero potential. An "applied part" for purposes of the present disclosure is any part of the system 550 that: (i) comes into physical contact with the patient or operator performing the dialysis treatment; (ii) can be brought into contact with the patient or operator; or (iii) needs to be touched by the patient. For instance, it is possible for the patient to touch the upper or lower plates of the plate heater 16, the temperature sensors 62 and the enclosure or housing 112. The applied part 568 represents schematically the casing or insulation around the temperature sensors 62.

In an embodiment, which only includes a display device 40 and not a touch screen 42 (discussed in FIGS. 1 and 2), the housing 112 includes a window 570, such as a glass or clear plastic window. The glass or plastic window provides the same level of insulation as the rest of the, e.g., plastic housing or enclosure 112. In an embodiment which does include a touch screen 42, the touch screen is properly electrically insulated, e.g., by the manufacturer of same. Alternatively,

50

one or more layers of insulation discussed below could be added to system 550 to properly insulate the touch screen 42.

The system 550 makes available an input/output port 572, which can be a serial port or an Ethernet port to connect the system 550 to an external computer, a local area network, a wide area network, an internet and the like. To electrically insulate input/output port 572, the system provides a protective covering or casing 574.

The mains part 556 powers the heater element 576, which is positioned and arranged to heat both the upper and lower plates of the plate heater 16. In an alternative embodiment (not illustrated), the mains part 556 powers the infrared heater discussed above. As illustrated, double insulation is maintained between the heater element 576 and the heater plate 16. The double insulation includes basic insulation B(240), rated for 240 VAC, and supplemental insulation S(240), rated for 240 VAC.

For the heater plate 16 and element 576, at least, the basic and supplemental insulation needs to be electrically insulative but thermally conductive. Polyimides, such as a Kapton®, work very well. In an embodiment, therefore, the B(240) and S(240) layers each include Kapton® tape or sheet of about 0.3 millimeters thickness. As further illustrated, another layer of basic insulation B(240), rated for 240 VAC, and another layer of supplemental insulation S(240), rated for 240 VAC, are disposed between the temperature sensor 62 and the heater plate 16. Thus the heater plate 16 is completely and doubly insulated from the remainder of the system 550. Alternatively, either of the double layers of insulation can be replaced by a single layer of reinforced insulation.

The line 552 and the neutral 554 are insulated by basic operation insulation BOP (240), rated for 240 VAC, which is the electrical insulation wrapped or extruded around the respective wires. Basic insulation B(240), rated for 240 VAC, is provided between the mains part 556 and the enclosure 112 and between the power supply PCB 558 and the enclosure. The basic insulation B(240) can be in the form of a properly separated air gap. The enclosure 112 itself provides supplemental insulation S(240) for 240 VAC. The mains part 556 is therefore doubly insulated from the outside of the enclosure 112.

Since applied part 568 is maintained at a zero operating voltage, there needs to be no additional insulation placed between the applied part 568 and the housing 112. Accordingly, there is simply an operational separation displayed figuratively as OP between the applied part 568 and the housing 112. Double insulation or reinforced insulation D/R (24) for 24 VDC is however provided between live part 560 and the applied part 568, so that applied part 568 maintains its zero potential. Basic insulation B(24), rated for 24 VDC, is provided between live part 560 and the enclosure 112. The basic insulation B(24) can be in the form of a properly separated air gap. As stated above, the enclosure 112 itself provides supplemental insulation S(240) for 240 VAC. Live part 560 is therefore doubly insulated from the outside of the enclosure 112.

No additional insulation is needed and only an operational separation OP is provided between live part 560 and the live part 566. Since live part 566 is stepped up to 1200 Vpeak, the supplemental insulation S(240) rated for only 240 VAC of the enclosure 112 should not be relied upon. Accordingly, double insulation or reinforced insulation D/R (1200) for 1200 Vpeak is provided between the live part 566 and the housing 112.

Double insulation or reinforced insulation D/R (240) for 240 VAC is provided between the mains part 556 and the live part 560. Double insulation or reinforced insulation D/R (240) for 240 VAC is also provided between the line and

neutral line **554** and the upper and lower plates of plate heater **16**. Still further, double insulation or reinforced insulation D/R (**240**) for 240 VAC is provided between the mains part **562** and the live part **564** of the power supply PCB **558**. Here, in the case of double insulation, either the basic or supplementary insulation can be a properly separated creepage distance on the PCB **558**.

Double insulation or reinforced insulation D/R (**24**) for 24 VDC is provided between the housing **112** and the display device **40**. The separation between the display device **40**, maintained at 24 VDC and the inverter, maintained at 1200 Vpeak is only required to be operational. Live part **566** must be separated from the outside of the housing **112** by D/R (**1200**) but not from the LP(**24**). The reason is that the LP(**1200**) is on the secondary side of the live part **566** and if it is shorted to the LP(**24**) due to a failure of the operational insulation, LP(**1200**) will become at most 24 VDC, providing no safety hazard.

X. Graphical User Interface

Referring now to FIG. **29**, one embodiment of a graphical user interface ("GUI") system **600** is illustrated. The GUI system **600** in an embodiment employs web-based software as well as other types of software. As discussed previously in connection with FIG. **28**, the system **10, 100** of the present disclosure is provided with an input/output (e.g., serial or Ethernet) port **572**, which is normally insulated from the patient by a cover **574**. The port **572** allows the controller **30** of the system **10, 100** to access an internet and a variety of other networks. The GUI system **600** of the present disclosure takes advantage of this capability by enabling the controller **30** to interact with software on an internet or other network.

It should be appreciated that the GUI system **600** does not require the patient to have internet or network access in their home. Rather, the port **572** is for a maintenance person or installer to gain access to the controller **30** within the hardware unit **110**. In this manner, the patient may bring their unit to a place having internet or network access, wherein the patient's software may be upgraded. The patient may then bring the unit home and operate it without having to gain internet or network access.

Using web-based software is advantageous because it is based on well established standards, so that the interface screens may be constructed using existing software components as opposed to being hand crafted. Web-based software allows for external communication and multiple access points. The software is portable. For each of these reasons, software constructed using existing software components reduces development time and cost.

The present disclosure includes the construction of a GUI using an embedded web browser **602**. In an embodiment, the embedded web browser **602** is third party software. The embedded web browser **602** can include any third party browser that runs on a target platform and includes support for advanced features such as HTML 4.0, ECMAScript, and animated GIFs. The web browser **602** renders and supplies the various GUI screens to the video monitor **40**. The web browser **602** also handles inputs made by the patient. When the operator interacts with the system (e.g., presses buttons **43, 124, 125** and **127** or turns knob **122**, illustrated in FIG. **3B**), the web browser **602** forwards information about the interaction to the embedded web server **604**.

The web server **604** in turn uses a web server extension software **606** to process the interaction. The embedded web server **604** can also be any third party web server that runs on a target platform and includes support for the web server

extension software **606** and that allows a dynamic definition of the information to be sent to the embedded web browser **602**.

The web server extensions are developed internally using the web server extension software **606** and conform to the specification of a mechanism, such as a Servlet, which works in conjunction with the chosen embedded web server **604**. The web server extension software **606** enables the web server **604** to retrieve back end and real time information from the instrument access and control software **608**. There are a number of different existing web server extension technologies that may be used for the embedded web browser **602**, the embedded web server **604** and the web server extension software **606**, such as CGI, ASP, Servlets or Java Server Pages ("JSP").

The web server extension software **606** interacts with the instrument access and control software **608**. The instrument access and control software **608** is an internally developed operating environment for controlling the various lower level components of the system **10, 100**, such as the valve motor/actuator, pump motor/actuator and heater.

Depending on the operator input and the state of the automated dialysis system **10, 100**, the web server extension software **606** can interact with the instrument access and control software **608** to obtain information from same and to cause one of the devices of the system **10, 100** to take action. The web server extension software **606** then sends information to the embedded web browser **602**, which may then be displayed on the display device **40**. The web server extension software **606** communicates with the instrument access and control software **608** using, in an embodiment, the CORBA standard. This communication, however, may take place using various different protocols known to those of skill in the art.

During the operation of the system **10, 100**, an event may occur that requires high priority information to be displayed to the operator, for example, an alarm and corresponding message either on the display device **40** or on a separate dedicated alarm display. When a high priority event occurs, the instrument access and control software **608** generates an event that is handled by an event-handling software **610**, which can be developed internally. The event-handling software **610** in turn notifies the embedded web browser **602**, through the use of a plug-in or a refresh request simulation from the web server **604**, to refresh whatever display the web browser is currently causing to be displayed on display device **40**.

The event-handling software **610** enables information to flow from the instrument access and control software **608** to the embedded web browser **602** without a request by the embedded web browser **602**, wherein the web browser thereafter requests a refresh. The web server **604** then forwards the request to the web server extension software **606**. The web server extension software **606** determines what information should be displayed on the display device **40** based on the state of the system **10, 110**. The web server extension software **606** then relays that information back to the embedded web browser **602**, which updates the display device, e.g., to show an alarm condition.

In one embodiment of the GUI system **600**, the web client is internal to the hardware unit **110** of the system **10, 100**. As described above in connection with FIG. **1**, the controller **10** includes a plurality of processors (referred to collectively herein as processor **34**). A main microprocessor is provided that resides over a number of delegate processors. Each of the embedded web browser **602**, web server **604**, web server extension software **606** and event handling software **610** run

US 7,815,595 B2

53

on the main microprocessor. The instrument access and control software **608** runs on the main microprocessor and one or more of the delegate processors.

It is alternatively possible that a number of different external web clients may need to access information contained within the system **10, 100**. It is therefore preferred that the HTTP commands to the embedded web server **604** not require predetermined passwords, but instead use a stronger and more flexible security system.

Referring now to FIGS. **30A-30M**, a number of screen shots of the GUI **600** are illustrated that show the overall look and feel of the system **10, 100** as seen by the operator or patient. Further, these drawings illustrate various features provided by the GUI system **600**. The goal of the automated dialysis system of the present disclosure is to make a simple and well operating system. The device only requires two supply bags **14**, weighs less than 10 kg and can be powered virtually anywhere in the world without the risk of electrical shock to the patient. Similarly, the GUI system **600** is designed to be simple, intuitive, effective, repeatable and reliable.

As illustrated in FIG. **3B**, the system **10, 100** includes a display device **40**, a knob **122** that enables the user to interact with the GUI system **600** and a number of dedicated push-buttons **43** that enable the patient to navigate between three different screens namely a parameter change screen, a log screen and a therapy screen. In an embodiment, a display device **40** is provided, wherein the input devices **43, 122, 124, 125** and **127** are each electromechanical. In an alternative embodiment, one or more of the input devices are provided by a touch screen **42** that operates with the display device **40** and a video controller **38**.

A simulated or electromechanical "stop" input **124**, an "OK" button **125** and a "back" button **127** are also provided. The OK button **125** enables the operator to indicate that a particular part of the set-up procedure has been completed and to prompt the GUI **600** to move on to a next step of the set-up stage or to the therapy stage. The stop button **124** enables the operator or patient to stop the set-up or therapy procedures. The system **600** may include a handshake type of response, such as "are you sure you want to stop the set-up". Other parts of the entire procedure, such as the patient fill or drain cycles immediately stop without further input from the operator. At certain points in the procedure, the system enables the operator to move back one or more screens using the back button **127**.

Referring now to FIG. **30A**, the display device **40** and the video controller **38** are adaptable to display animations, which provide the patient with information and instructions **612** in a comfortable format. As illustrated throughout the screen shots, the GUI system **600** waits for the patient to read and understand whatever is being displayed on the display device **40** before moving on to the next step or stage. FIG. **30A** illustrates that the GUI system **600** is waiting until the patient is ready before beginning the therapy. The system **600** prompts the user to press an "OK" input to begin the therapy. FIG. **30A** also illustrates that the therapy screen is being presently displayed by highlighting the word "therapy" at **614**.

In FIG. **30B**, the display device **40** of the GUI system **600** prompts the patient to gather the necessary supplies for the therapy, such as the supply bags **14**. FIGS. **30B** and **30C** illustrate that the system **600** uses static images, such as static image **616** and animations, such as animation **618**, which resemble the actual corresponding supplies or parts to aid the patient in easily, effectively and safely connecting to the system **10, 100**. For example, the animation **618** of FIG. **30C**

54

looks like the actual hose clamp of the system **10, 100**, which aids the patient in finding the proper piece of equipment to proceed with the therapy. The arrow of the animation **618** also illustrates the action that the patient is supposed to perform, reducing the risk that the patient will improperly maneuver the clamp or perhaps break the clamp.

FIGS. **30D** and **30E** illustrate that the GUI system **600** promotes hygienic operation of the system **10, 100** by prompting the patient to: (i) take the steps of covering the patient's mouth and nose at the proper time; and (ii) wash the patient's hands before coming into contact with critical fluid connectors, such as the patient fluid connector and the supply bag connectors. The GUI system **600** waits for the patient to finish and press an OK input at each step before proceeding to the next step. As illustrated in FIGS. **30D** and **30E**, software LEDs **620** located at the top of the display device **40** indicate where the user is in the setup procedure.

Screen shots of FIGS. **30A** to **30E** and **30H** to **30M** each present procedural set-up steps of the therapy. Accordingly, the colors of the screen shots of FIGS. **30A** to **30E** and **30H** to **30M** are chosen so that they are more visible when viewed during the day or with lights on. In one embodiment, the screens are different shades of blue, wherein the static images and animations and inner lettering are white and the outer lettering and borders are black. As illustrated by FIGS. **30F** and **30G** however, the screen shots that illustrate the active stages of the therapy are chosen so that they are more visible when viewed at night or with lights off. In one embodiment, the screen shots of FIGS. **30A** to **30F** are black with ruby red lettering, diagrams and illustrations, etc. The red lettering is configured so as not to be intrusive to a sleeping patient but still visible at distances of about 10 to 25 feet (3 to 7.6 meters).

FIGS. **30F** and **30G** illustrate that during active stages of the therapy, the therapy status information is displayed on the screen shots in the form of both graphics **622** and numerical data **624**. Therapy status information is displayed in real time or in substantially real time with a slight time delay. FIG. **30F** illustrates a screen shot during a fill portion of the therapy. In particular, FIG. **30F** illustrates the first fill of three total fills. The graphical clock **622** illustrates that the fill cycle time is approximately $\frac{1}{8}$ th elapsed. The arrow graphic **622** indicates that the therapy is in a fill cycle. Also the graphical representation of the body **622** has a very low percentage of dialysate. The numerical data **624** illustrates that the system **10, 100** has pumped 150 ml of dialysate into the patient.

FIG. **30G** illustrates that the patient is currently undergoing the first drain cycle of three drain cycles that will take place overnight. The graphical representation of the clock illustrates that the drain cycle time is approximately $\frac{1}{8}$ th elapsed. The graphical arrow is pointing downward indicating a drain cycle. The body is shown as being substantially full of dialysate. The numerical data **624** illustrates that 50 ml of dialysate has been removed from the patient.

FIGS. **30H** and **30I** illustrate that in the morning when the therapy is complete, the screen reverts back to the daytime colors, or colors which are more easily seen in a lighted room. FIG. **30H** includes information and instructions **612** that prompt the patient to disconnect from the system **10, 100**. The system waits for the patient to select the OK button **125** (FIG. **3B**) before proceeding. FIG. **30I** includes an animation **618**, which illustrates an action and equipment that the patient while disconnecting from the system. For each action in the disconnection sequence, system **600** waits for the patient to select the OK button **125** (FIG. **3B**) before proceeding.

FIGS. **30J** to **30M** illustrate that in an embodiment, the user navigates between the therapy, parameter changes and log information by selecting one of the dedicated inputs **43** illus-

US 7,815,595 B2

55

trated in FIG. 3B. FIG. 30J illustrates that the patient has selected the input 43 associated with the parameter changes information. The screen 40 in FIG. 30J now highlights the word "changes" instead of the word "therapy."

The parameter screen presents parameter information to the patient in a hierarchy format. First, as in FIG. 30J, the system 600 presents categories 625 of parameters, such as patient preferences, daily patient data, therapy parameters, nurse parameters and service parameters. The patient can scroll through the various categories 625 using the adjustment knob 122 of FIG. 3B, so that a desired category 625 is displayed in a highlighted display area 626. FIG. 30H illustrates that the patient preferences category 625 is currently displayed in the highlighted display area 626.

Once the user selects a highlighted category 625 by pressing the OK button 125 (FIG. 3B), a first door 628 slides open and presents the user with a list of the parameters 627 for the selected category 625 (e.g., the patient preferences category), as illustrated by the screen 40 of FIG. 30K. FIG. 30K illustrates that the patient preferences category 625 is displayed above the door 628, so that the patient knows which category 625 of parameters 627 is being displayed. At the same time, the highlighted display area 626 now displays one of a select group of the parameters 627 belonging to the patient preferences category 625.

The parameters 627 illustrated in FIG. 30K as belonging to the patient preferences category 625 include a display brightness percent, a speaker volume percent and a dialysate temperature in degree Celsius. Obviously, the patient preferences category 625 may include other parameters 627. The other categories 625 illustrated in FIG. 30J include different parameters 627 than those illustrated in FIG. 30K.

The patient can scroll through and select one of the parameters 627 for the patient preferences category 625 by rotating knob 122. In this manner, it should be appreciated that the signal knob 122 is used over and over again. This feature is in accordance with the goal of providing a simple system, wherein the patient only has to turn one knob instead of remembering which knob from a plurality of knobs applies to a particular feature. The knob 122 also enables the lettering to be bigger because the patient can scroll through to see additional parameter selections that are not displayed when the door 628 is initially displayed. That is, the functionality of the knob 122 provides freedom to the GUI 600 to not have to display all the possible parameters at once. It should be appreciated that this benefit also applies to the category selection screen of FIG. 30J, wherein each of the categories 625 does not have to be displayed simultaneously.

Once the patient selects one of the parameters of the patient preferences category, e.g., by pressing the OK button 125, a second door 630 slides open, wherein the display device 40 illustrates that the patient has selected the display brightness parameter 627 of the patient preferences category 625, which is still displayed by the first door 628 in FIG. 30L. The highlighted area 626 now displays one of the range of possible values 632 for the selected parameter 627 of the selected category.

In FIG. 30L display device 40 illustrates that the highlighted display area 626 currently shows a value 632 of eighty for the display brightness parameter 627 of the patient preferences category. Once again, the patient changes the value 632 of the selected parameter 627 by rotating the knob 122. When the patient selects a value 632 (by pressing the OK input 125 illustrated in FIG. 3B while the desired value is displayed) for the parameter of the chosen category, the GUI system 600 saves the value as indicated by the display device

56

40 in FIG. 30M. FIG. 30M illustrates that the system 600 provides a feedback message to the patient that the selected value has been saved.

The system 600 in an embodiment presents information and instructions to the operator through the various visual tools discussed above. In an alternative embodiment, in addition to the visual information and instructions 612, static images 616, animations 618, parameter information, etc., one, or more or all of the above disclosed methods of communication is presented audibly to the patient or operator through speakers 129 (FIG. 3B) and a sound card (not illustrated) that cooperate with the controller 30 of the system 10, 100.

The various programs that run on the main microprocessor can also include one or more programs that activate a certain sound file at a certain time during the therapy or upon a certain event initiated by the system 600, e.g., an alarm, or upon a patient or operator input. The sound files can contain the sound of a human voice or any other type of sound. The sound files walk the patient through the set-up portion of the therapy in an embodiment. The sound files can alert a patient who has made an inappropriate input into the GUI 600, etc. The system does not activate a sound during the cycles, e.g., while the patient sleeps, in an embodiment.

If the operator selects the dedicated input 43 corresponding to the log information (not illustrated), the GUI 600 displays a screen or screens that show therapy data. In an embodiment, the therapy data is presented in a number of operator selectable logs. One of the logs can be a default log that is displayed initially, wherein the operator can switch to another log via, e.g., the knob 122. The logs may pertain to the most recent therapy and/or can store data over a number of days and a number of therapies. The logs can store any type of operating parameter information such as cycle times, number of cycles, fluid volume delivered, fluid temperature information, fluid pressure information, concentration of dialysate constituents, any unusual or alarm type of events, etc.

It should be understood that various changes and modifications to the presently preferred embodiments described herein will be apparent to those skilled in the art. Such changes and modifications can be made without departing from the spirit and scope of the present invention and without diminishing its intended advantages. It is therefore intended that such changes and modifications be covered by the appended claims.

The invention is claimed as follows:

1. A peritoneal dialysis system comprising:

a hardware unit, the hardware unit including a pump actuator including a contact surface;

a disposable unit received by the hardware unit, the disposable unit having a moveable membrane operable with the pump actuator, the pump actuator moveable towards and away from the disposable unit; and

the hardware unit and disposable unit configured to hold a negative pressure between the pump actuator and the moveable membrane of the disposable unit, the negative pressure causing the moveable membrane to contact and conform to a shape of the contact surfaces, so as to follow the pump actuator as the actuator is moved away from the disposable unit, the pump actuator and membrane positioned relative to each other such that the membrane contacts and moves with the pump actuator as the actuator is moved into the disposable unit.

2. The peritoneal dialysis system of claim 1, the moveable membrane operable to draw dialysate into the disposable unit as the actuator is moved away from the disposable unit.

US 7,815,595 B2

57

3. The peritoneal dialysis system of claim 1, the moveable membrane operable to push dialysate out of the disposable unit as the actuator is moved into the disposable unit.

4. The peritoneal dialysis system of claim 1, wherein the disposable unit includes at least one of: (i) a disposable cassette attached to a plurality of tubes; and (ii) a rigid piece to which the flexible membrane is attached.

5. The peritoneal dialysis system of claim 1, wherein a pumping head of the pump actuator forms the contact surface.

6. The peritoneal dialysis system of claim 1, further including at least one of: (i) a heater, wherein the disposable unit includes a heating portion operable with the heater; and (ii) a valve actuator, wherein the disposable unit includes a valve portion operable with the valve actuator.

7. The peritoneal dialysis system of claim 1, wherein the pump actuator includes a stepper motor.

8. The peritoneal dialysis system of claim 7, wherein the stepper motor is one of: (i) coupled to a ball screw, the ball screw converting rotational motion of the stepper motor to a translational motion for the pump actuator; and (ii) a linear stepper motor.

9. The peritoneal dialysis system of claim 1, wherein the pump actuator includes at least one of a pump piston and a piston head.

10. The peritoneal dialysis system of claim 1, which includes an apparatus positioned and configured to aid in a seal that maintains the negative pressure between the pump actuator and the moveable membrane of the disposable unit.

11. A peritoneal dialysis system comprising:

a hardware unit, the hardware unit including a pump actuator;

a disposable unit received by the hardware unit, the disposable unit having a moveable membrane operable with the pump actuator, the pump actuator moveable towards and away from the disposable unit; and

a sealing diaphragm positioned so as to at least partially form a sealed area between the pump actuator and the moveable membrane of the disposable unit, the sealed area allowing a negative pressure to be applied between the pump actuator and the moveable membrane of the disposable unit, the negative pressure causing the moveable membrane to follow the pump actuator as the actuator is moved away from the disposable unit.

58

12. The peritoneal dialysis system of claim 11, wherein the sealing diaphragm is moveable with the pump actuator.

13. The peritoneal dialysis system of claim 11, wherein the sealing diaphragm includes a seal around a translating shaft of the pump actuator.

14. The peritoneal dialysis system of claim 11, wherein the negative pressure is supplied by a pneumatic source separate from the pump actuator.

15. A peritoneal dialysis system comprising:

an enclosure including a hardware unit that houses a pump actuator and a door having a recessed area capable of receiving at least a portion of the pump actuator, the recessed area forming at least a portion of a pump chamber;

a disposable cassette received between the hardware unit and the door; and

the cassette including a flexible membrane configured to be drawn into contact with the pump actuator by a negative pressure, the pump actuator operable to move the flexible membrane within the portion of the pump chamber formed by the recessed area of the door.

16. The peritoneal dialysis system of claim 15, wherein the hardware unit and the door of the enclosure hold the negative pressure between the pump actuator and the portion of the pump chamber formed by the recessed area of the door, the negative pressure causing the flexible membrane to follow the pump actuator as the actuator is moved away from the pump chamber.

17. The peritoneal dialysis system of claim 16, wherein the negative pressure is supplied by a pneumatic source separate from the pump actuator.

18. The peritoneal dialysis system of claim 15, wherein the door is hinged to the hardware unit.

19. The peritoneal dialysis system of claim 15, wherein the portion of the pump chamber of the door is a first portion, and wherein the hardware unit forms a second portion of the pump chamber.

20. The peritoneal dialysis system of claim 15, wherein the at least a portion of the pump chamber of the door is dome shaped.

* * * * *