



US006882877B2

(12) **United States Patent**  
**Bonutti**

(10) **Patent No.:** **US 6,882,877 B2**  
(45) **Date of Patent:** **Apr. 19, 2005**

(54) **MAGNETIC RESONANCE IMAGING  
SYSTEM AND METHOD**

(75) Inventor: **Peter M. Bonutti**, Effingham, IL (US)

(73) Assignee: **Bonutti Research, Inc.**, Effingham, IL  
(US)

(\*) Notice: Subject to any disclaimer, the term of this  
patent is extended or adjusted under 35  
U.S.C. 154(b) by 0 days.

(21) Appl. No.: **10/666,101**

(22) Filed: **Sep. 18, 2003**

(65) **Prior Publication Data**

US 2004/0133097 A1 Jul. 8, 2004

**Related U.S. Application Data**

(60) Continuation of application No. 09/118,665, filed on Jul. 17,  
1998, now Pat. No. 6,697,659, which is a division of  
application No. 08/455,074, filed on May 31, 1995, now Pat.  
No. 6,671,537, said application No. 08/455,074, is a division  
of application No. 08/221,848, filed on Apr. 1, 1994, now Pat.  
No. 5,577,503, said application No. 08/221,848, is a  
division of application No. 07/802,358, filed on Dec. 4,  
1991, now Pat. No. 5,349,956.

(51) **Int. Cl.**<sup>7</sup> ..... **A61B 5/055**

(52) **U.S. Cl.** ..... **600/410; 600/415; 600/422;**  
324/318

(58) **Field of Search** ..... 600/410, 415,  
600/422; 324/307, 309, 318, 322

(56) **References Cited**

**U.S. PATENT DOCUMENTS**

1,239,146 A	9/1917	Wantz
2,801,142 A	7/1957	Adams
2,972,505 A	2/1961	Weickgenannt
3,025,397 A	3/1962	Travis et al.
3,124,328 A	3/1964	Kortsch
3,521,876 A	7/1970	Smith
3,528,413 A	9/1970	Aydt
3,766,384 A	10/1973	Anderson

4,050,355 A	9/1977	Niskanen
4,136,858 A	1/1979	Petersen
4,232,681 A	11/1980	Tulaszewski
4,256,112 A	3/1981	Kopf et al.
4,291,229 A	9/1981	Patt
4,323,080 A	4/1982	Melhart
4,407,277 A	10/1983	Ellison
4,411,270 A *	10/1983	Damadian ..... 600/410
4,501,414 A	2/1985	Mason et al.
4,534,076 A	8/1985	Barge
4,534,358 A	8/1985	Young
4,562,588 A	12/1985	Ruf
4,616,814 A	10/1986	Harwood-Nash et al.
4,681,308 A	7/1987	Rice
4,717,133 A	1/1988	Walsh et al.

(Continued)

**FOREIGN PATENT DOCUMENTS**

DE	3140225	4/1983
DE	3609535	9/1987
DE	3732579	4/1989
EP	0067933	12/1982

**OTHER PUBLICATIONS**

Frank G. Shellock, PhD., et al, Article under the heading  
Technical Developments and Instrumentation, entitled  
"Patellofemoral Joint: Kinematic MR Imaging to Assess  
Tracking Abnormalities", published by Radiology, publica-  
tion date 1988, vol. 168, No. 2 pp. 551-553.


(Continued)

*Primary Examiner*—Ruth S. Smith  
(74) *Attorney, Agent, or Firm*—Fleit Kain Gibbons Gutman  
Bongini & Bianco P. L.; Paul D. Bianco

(57) **ABSTRACT**

The present invention relates to a magnetic resonance imag-  
ing system that includes a stationary electromagnet, a patient  
support for maintaining a patient in a standing or seated  
position, and an actuator for raising and lowering the patient  
relative to a magnetic field of the electromagnet such that the  
patient is located within the magnetic field.

**20 Claims, 16 Drawing Sheets**



## U.S. PATENT DOCUMENTS

4,791,372	A	12/1988	Kirk et al.	
4,827,496	A	5/1989	Cheney	
4,834,112	A	5/1989	Machek et al.	
4,875,485	A	* 10/1989	Matsutani	600/415
4,924,198	A	5/1990	Laskaris	
4,938,230	A	7/1990	Machek et al.	
4,968,937	A	11/1990	Akgun	
5,001,739	A	3/1991	Fischer	
5,007,912	A	4/1991	Albrektsson et al.	
5,008,624	A	4/1991	Yoshida	
5,078,140	A	1/1992	Kwoh	
5,153,546	A	10/1992	Laskaris	
5,154,178	A	* 10/1992	Shah	600/415
5,275,174	A	1/1994	Cook	
5,305,750	A	4/1994	Makita	
5,329,924	A	7/1994	Bonutti	
5,343,580	A	9/1994	Bonutti	
5,462,551	A	10/1995	Bailey et al.	
5,542,423	A	8/1996	Bonutti	
5,562,094	A	10/1996	Bonutti	
5,640,958	A	6/1997	Bonutti	
5,743,264	A	4/1998	Bonutti	
6,044,289	A	3/2000	Bonutti	
6,141,579	A	10/2000	Bonutti	
6,504,371	B1	1/2003	Damadian et al.	

## OTHER PUBLICATIONS

Manabu Minami, et al., Article entitled "MR Study of Normal Joint Function Using a Low Field Strength System", Journal of Computer Assisted Tomography, Nov./Dec. 1991 Raven Press, Ltd., New York, vol. 15, No. 6, pp. 1017-1023.

Frank G. Shellock, et al., Article entitled "Kinematic MRI of the Joints", Raven Press New York 1990, Chapter 10, pp. 553-590.

Keith R. Burnett, et al., Article entitled "Dynamic Display of the Temporomandibular Joint Meniscus by Using Fast-Scan MR Imaging", American Journal of Roentgenology: vol. 149, Nov. 1987, pp. 959-962.

Frank G. Shellock, et al., Article entitled "Patellofemoral Joint: Evaluation During Active Flexion With Ultrafast Spoiled Grass MR Imaging", Radiology, vol. 180, No. 2, Aug. 1991, pp. 581-585.

Joseph W. Carlson, et al., Article entitled "A Technique for MR Imaging of the Knee Under Large Flexing Angles", Magnetic Resonance Imaging, vol. 8, No. 4, 1990, pp. 407-410.

Frank G. Shellock, et al., Article entitled "Kinematic Magnetic Resonance Imaging for Evaluation of Patellar Tracking", The Physician and Sportsmedicine, vol. 17, No. 9, Sep. 89, pp. 99-108.

William F. Conway, MD, PhD, et al., Article entitled "Temporomandibular Joint Motion: Efficacy of Fast Low-Angle Shot MR Imaging", MS, Radiology, vol. 172, No. 3, Sep. 1989, pp. 821-826.


Frank G. Shellock, Patellofemoral Joint Abnormalities in Athletes: Evaluation by Kinematic Magnetic Resonance Imaging, Top Magn Reson Imaging, 1991;3 (4):71-95.

Frank G. Shellock et al., Patellar Tracking Abnormalities: Clinical Experience with Kinematic MR Imaging in 130 Patients, Radiology 1989; 172:799-804.

Pete Sullenberger, Letter to Solomon with attached flyers, Sep. 12, 1991.

Frank G. Shellock, Kinematic Magnetic Resonance Imaging of the Joints: Techniques and Clinical Application, Magnetic Resonance Quarterly vol. 7 No.2, pp. 104-135, 1991 Raven Press, Ltd., New York.

\* cited by examiner






Fig.1B








Fig.1C






Fig. 6




Fig. 7




Fig. 8













Fig. 9










Fig.19




Fig.20




Fig.21




Fig.22








Fig.23










Fig. 26A






Fig. 29




Fig.30




Fig.31




Fig.32




Fig.33






Fig.34




Fig.35




Fig.36




Fig.41




Fig.42




Fig.43




Fig.37




Fig.38




Fig.39




Fig.40

1

# MAGNETIC RESONANCE IMAGING SYSTEM AND METHOD

## CROSS-REFERENCE TO RELATED APPLICATION

This application is a continuation of application Ser. No. 09/118,665, filed Jul. 17, 1998, now U.S. Pat. No. 6,697,659. The aforementioned application Ser. No. 09/118,665 is itself a divisional of application Ser. No. 08/455,074, filed May 31, 1995, now U.S. Pat. No. 6,671,537. The aforementioned application Ser. No. 08/455,074 is a divisional of application Ser. No. 08/221,848, filed Apr. 1, 1994, now U.S. Pat. No. 5,577,503. The aforementioned application Ser. No. 08/221,848 is itself a divisional of application Ser. No. 07/802,358, filed Dec. 4, 1991, now U.S. Pat. No. 5,349,956.

## BACKGROUND OF THE INVENTION

The present invention relates to an apparatus and method for use in medical imaging. More particularly, the present invention relates to apparatus and method for positioning a patient and/or a secondary imaging coil inside a primary imaging coil.

In magnetic resonance imaging, a patient is placed inside a coil (the "primary" coil) which is large enough in diameter to receive the patient while he is lying prone on a table slidable into and out of the coil. A selected portion of the patient is then imaged by the use of electromagnetic radiation from the primary coil.

It is known to place smaller coils, called surface or volume coils, in close proximity to the specific part of the patient to be imaged, such as the neck, spine, or knee. These coils, referred to herein as secondary coils, are used to increase resolution by having a coil closer to the part to be imaged. It is essential to place the secondary coil in a particular orientation relative to the electromagnetic field generated by the primary coil.

Current imaging systems can only take images while a patient is in one particular position. One known device allows the patient to move his knee joint to different selected positions while the patient is in the primary coil. This device requires the patient to lie face down in the primary coil, which is extremely uncomfortable for the extended period of time required to image properly, especially in the close, almost claustrophobic confines of a primary MRI coil.

## SUMMARY OF THE INVENTION

The present invention is an apparatus and method for use in medical imaging. The present invention provides a system to simulate within an imaging coil normal movements of body parts such as joints, and to improve imaging of soft tissue and bony parts as compared to a static system in which images are taken of a joint in only one position.

In accordance with a first aspect of the present invention, there is provided controlled motion of an extremity, while in an imaging coil, either patient directed or operator directed. A joint or body part is moved into various positions in multiple planes within its range of motion while a series of images are taken of the joint in the different positions. These individual images may then be collated into a cine format to effectively show the joint in motion. Thus, the present invention allows for studying a joint in motion and also allows for studying a joint or other body part at any positions within its range of motion allowable within the confines of the primary coil.

In accordance with a second aspect of the present invention, a surface or volume coil (referred to herein as a

2

secondary coil) is coupled for movement with the joint or body part. The secondary coil is maintained in the proper spatial relationship with the primary coil's electromagnetic field. Keeping the secondary coil as close as possible to the joint or tissue being imaged, while moving the joint or body part, provides greatly enhanced resolution and more detail in the final image.

Thus, to illustrate these first two aspects of the invention in knee imaging, the knee is fixed by holding the upper and lower legs with cuffs and a secondary coil is placed around the knee itself. The knee is then imaged at 0 degrees by using the primary and secondary coils. The knee is then flexed (either by the patient or the operator), and the secondary coil moves with the knee. The knee is progressively moved through various positions within its range of motion as limited only by the size of the primary coil. Images are taken at each position. The images may then be collated and shown in sequence to visualize the movement of the knee joint, or may be studied individually to study the joint at each position.

Similar systems are available for other joints, the back, neck, etc. These systems all are preferably provided as mechanisms usable with existing imaging tables to reduce cost. Alternatively, some of these may be built into a new imaging table.

Coupling a surface coil for movement with the extremity provides the necessary detail in the images, even with a larger primary coil, which is not available with present systems. Accordingly, it is possible to use a larger diameter primary coil, allowing this increased range of movement, without the degradation in image quality which would be expected from the increased coil size. For example, the knee could be flexed through its entire range of motion to allow optimum imaging of the knee joint. This is currently impossible with the known small primary coils which only allow about 50 degrees of flexion.

In accordance with another aspect of the present invention, traction is applied to a joint being imaged, in order to load the joint. This can simulate normal loading of a joint. Distracting a joint can also allow a better view of the parts of the joint and thus an increased imaging benefit. It can also allow simulation of normal loading of a joint, such as when carrying a heavy object or performing an athletic or work-related task. This feature is not available with present imaging apparatus. Traction can also be applied to a joint being imaged when the joint is in various positions, to simulate normal loading of a joint within its range of motion. Again, this feature is not available with present imaging apparatus.

## BRIEF DESCRIPTION OF THE DRAWINGS

The foregoing and other features of the present invention will become apparent to one skilled in the art upon a consideration of the following description of the invention with reference to the accompanying drawings, wherein:

FIG. 1A is a perspective view of a magnetic resonance imaging installation including a patient support table constructed in accordance with the present invention;

FIG. 1B is a view of the table of FIG. 1A in another condition;

FIG. 1C is a view of the table of FIG. 1A in yet another condition;

FIG. 2 is an enlarged view of a back imaging platform of the table of FIG. 1A;

FIG. 3 illustrates the platform of FIG. 2 in a raised condition;

3

FIG. 4 illustrates the platform of FIG. 2 in a lowered condition;

FIG. 5 is a view similar to FIG. 2 showing a different actuating mechanism for the back imaging platform;

FIG. 6 illustrates the platform of FIG. 5 in a raised condition;

FIG. 7 illustrates the platform of FIG. 5 in a lowered condition;

FIG. 8 is an enlarged view of a knee imaging platform portion of the table of FIG. 1;

FIG. 9 illustrates the platform of FIG. 8 in a raised condition;

FIG. 10 is a view similar to FIG. 8 and showing a different actuating mechanism for the knee platform;

FIG. 11 illustrates the platform of FIG. 10 in a raised condition;

FIG. 12 is an enlarged view of a neck imaging platform portion of the table of FIG. 1A;

FIG. 13 illustrates the platform of FIG. 12 in a raised condition;

FIG. 14 illustrates the platform of FIG. 12 in a lowered condition;

FIG. 15 is a view similar to FIG. 12 illustrating a different actuating mechanism for the neck platform;

FIG. 16 illustrates the platform of FIG. 15 in a raised condition;

FIG. 17 illustrates the platform of FIG. 15 in a lowered condition;

FIG. 18 illustrates the platform of FIG. 15 with a foot rest attached for use in ankle imaging;

FIG. 19 illustrates the platform of FIG. 18 in a raised condition;

FIG. 20 illustrates the platform of FIG. 18 in a lowered condition;

FIG. 21 illustrates the platform of FIG. 18 with a different actuating mechanism;

FIG. 22 illustrates the platform of FIG. 21 in a raised condition;

FIG. 23 illustrates the platform of FIG. 21 in a lowered condition;

FIG. 24 is a top plan view of a shoulder positioning apparatus in accordance with the present invention shown attached to an imaging table with a shoulder coil;

FIG. 25 is a side view of the apparatus of FIG. 24;

FIG. 25A is a partial end view of the positioning apparatus of FIG. 25 taken along line 25A—25A of FIG. 24;

FIG. 26 is an enlarged perspective view of a portion of the positioning apparatus of FIG. 24;

FIG. 26A is a view similar to FIG. 26 showing an alternate indexing mechanism;

FIG. 27 is a top plan view of a head and neck positioning apparatus;

FIG. 28 is an end view of the apparatus of FIG. 27;

FIG. 29 is a perspective view of the apparatus of FIG. 27;

FIG. 30 is a top plan view of a wrist imaging apparatus embodying the present invention and including a hand cuff;

FIG. 31 illustrates the apparatus of FIG. 30 in a different condition;

FIG. 32 illustrates the apparatus of FIG. 30 with a different hand cuff;

FIG. 33 is an end view of the apparatus of FIG. 30;

4

FIG. 34 is an enlarged end view of the hand cuff of the apparatus of FIG. 30;

FIG. 35 is a side view of the hand cuff of FIG. 34;

FIG. 36 is a bottom plan view of the hand cuff of FIG. 34;

FIG. 37 is a top plan view of an independent patient directed knee positioning apparatus embodying the present invention;

FIG. 38 is a side view of the apparatus of FIG. 37;

FIG. 39 is a view similar to FIG. 38 with the apparatus in a raised condition;

FIG. 40 is a view similar to FIG. 39 with an optional distraction mechanism;

FIG. 41 is a schematic view showing the dimensions of a known primary MRI coil;

FIG. 42 is a schematic view showing the dimensions of a larger sized primary MRI coil embodying the present invention; and

FIG. 43 is a schematic view of a vertically extending primary MRI coil in accordance with the present invention.

#### DETAILED DESCRIPTION OF THE INVENTION

FIG. 1A illustrates a patient support table 10 for supporting a patient (not shown) during imaging inside a primary coil 12 of a magnetic resonance imaging installation. The table 10 is supported on tracks 14 and a floor support 16 for sliding longitudinal movement into and out of the coil 12.

The table 10 has an upper major side surface 20 extending between a head end 22 and a foot end 24. The table also has a right side 26 and a left side 28.

The table 10 includes a neck imaging platform indicated generally at 30. The neck imaging platform 30 includes a movable head panel 32 adjacent to a recess 34 for receiving a secondary imaging coil such as a cervical spine coil.

The table 10 includes a back imaging platform indicated generally at 40. The back imaging platform 40 includes a movable upper back panel 42 and a movable lower back panel 44. A movable center section 46 of the back imaging platform 40 includes a recessed panel 48 for receiving a secondary back imaging coil. The recessed panel 48 is located between a left side back panel 50 and a right side back panel 52.

The table 10 also includes a pair of knee imaging platforms 54 and 56. The left knee imaging platform 54 includes a movable upper left knee panel 58, a movable lower left knee panel 60 and, between them, a recessed panel 62 for receiving a left knee secondary imaging coil. Similarly, the right knee imaging platform 56 includes a movable upper right knee panel 64, a movable lower right knee panel 66, and a movable recessed panel 68 for receiving a right knee secondary imaging coil.

As can be seen in FIGS. 1A, 1B and 1C, the head panel 32 is movable between a plurality of positions relative to the upper major side surface 20 of the table 1A. In FIG. 1A, the head panel 32 is in a position level with the upper major side surface 20 of the table 10. In FIG. 1B, the head panel 32 is raised above the upper major side surface of the table 10. In FIG. 1C, the head panel 32 is lowered below the upper major side surface 20 of the table 10. With a patient's head on the head panel 32 and by moving the panel 32 in a manner as described below, a patient's cervical spine can be imaged in a variety of positions, by a coil placed on the recess panel 34.

Similarly, the back imaging platform 40 is movable between a plurality of positions relative to the upper major

5

side surface **20** of the table **10**, to image the back in varying positions. In FIG. 1A, the back imaging platform is level or flush with the upper major side surface **20** of the table **10**. In FIG. 1B, the back imaging platform **40** is raised up above the upper major side surface **20** of the table **10**, in order to hyper-extend the spine of a patient lying on the table **10**. In FIG. 1C, the back imaging platform **40** is lowered below the upper major side surface **20** of the table **10**, in order to flex the spine. Thus, by moving the back platform **40** between these various positions, in a manner to be described below, the back can be imaged in a plurality of different positions, rather than only in the one flat position possible with a flat table.

Similarly, the left and right knee imaging platforms **54** and **56**, respectively, are movable between a plurality of positions relative to the upper major side surface **20** of the table **10**. In FIGS. 1A and 1C, the platforms **54** and **56** are illustrated flush with the upper major side surface **20** of the table **10**. In FIG. 1B, the left knee imaging platform **54** is illustrated as raised up above the upper major side surface **20** of the table **10**. Both knee platforms **56** and **66** are each independently movable above or below the upper major side surface **20** of the table **10**. By thus moving a knee platform among these various positions, in a manner to be described below, a knee joint can be imaged in a plurality of positions, as opposed to the one single position available with a flat or nonmovable table.

FIGS. 2–4 illustrate in more detail the back imaging platform **40** and a mechanism for actuating same. These are exemplary of the other platforms and their actuating mechanisms. The fixed portion **70** of the table **10** includes a track **72** receiving mounting rollers **74** and **76** for the upper back panel **42** and lower back panel **44**, respectively. The upper back panel **42** is pivotally mounted at **78** to the center back section **46**. The lower back panel **44** is similarly pivotally mounted at **80** to the center back section **46**.

An inflatable bladder **82** extends between the center back section **46** and the lower panel **84** of the table **10**. The inflatable bladder **82** is supplied with fluid through a fluid supply line **86** extending along the table **10**. Fluid under pressure, preferably air, is supplied to the bladder **82** through the line **86** by means not shown such as a pump or a high pressure air line as is commonly found in hospitals, etc. Upon inflation of the bladder **82** from the condition shown in FIG. 2 to the condition shown in FIG. 3, the bladder **82** extends longitudinally, raising the center back section **46** of the table **10** upwardly from the major side surface **20** of the table **10**. The pivotal connections **74**, **76**, **78** and **80** allow the panels **42** and **44** to pivot upwardly, as illustrated in FIG. 3, sliding inwardly along the track **72**.

Similarly, upon the reduction of pressure in the bladder **82**, the bladder **82** compresses axially in length to the condition shown in FIG. 4, lowering the center back section **46** below the major side surface **20** of the table **10**. The state of inflation of the bladder **82** is selectively controllable to position the center back section **46**, relative to the upper major side surface **20** of the table **10**, at any position between the fully extended position illustrated in FIG. 3 and the fully flexed position illustrated in FIG. 4. The bladder **82** is preferably of a bellows-type construction for increased strength and controlled movement.

In FIGS. 2–4, a secondary coil **90** is illustrated as positioned on the center back section **46** of the table **10**. The secondary coil **90** may be any known imaging coil designed for imaging a portion of the spine of a patient. As the center back section **46** moves up and down, the secondary coil **90**

6

moves with the back section **46**. The secondary coil **90** is initially (FIG. 2) parallel to the upper major side surface **20** of the table **10**, and stays with the center back section **46** in that parallel orientation throughout the entire range of movement of the back section **46** and the coil **90**. It should be noted that any type of secondary coil either known or to be developed in the future—surface coil, volume coil, etc.—can be used with the present invention.

With a patient (not shown) lying on the table **10**, and a secondary coil **90** positioned as shown, the patient's back can be imaged in a plurality of orientations. If the table **10** is maintained in the position shown in FIG. 2, the patient's back is imaged in a flat position. If the table **10** is moved to a raised position, as illustrated in FIG. 3, the patient's back is imaged in an extended or hyper-extended condition. If the table **10** is moved to a lowered condition, as illustrated in FIG. 4, the patient's back is imaged in a flexed or hyper-flexed condition. Because the inflation state of the bladder **82** is selectively controllable and lockable at any state of inflation, the patient's back can be imaged when in any selected orientation within the full range of motion of the back imaging platform **40**. Meanwhile, the secondary coil **90** moves with the patient's back, always staying in close proximity thereto, to maintain the high resolution sought by use of a secondary coil. Further, the secondary coil **90** always maintains its orientation parallel to the upper side surface **20** of the table **10**, as is necessary for maximum resolution and clarity. Accordingly, it is seen that the present invention provides an apparatus for imaging the back of a patient, at any selected one or group of a plurality of orientations, while the patient is maintained in the primary coil **10** (FIG. 1) of the MRI imaging apparatus, and without any extra effort on the part of the patient.

FIGS. 5–7 illustrate an alternate actuating mechanism for the center back section **46** of the table **10**. The actuating mechanism includes a threaded rod **100** rotatably mounted in a block **102** fixed to the bottom panel **84** of the table **10**. The block **102** allows the rod **100** to rotate but prevents axial movement of the rod **100**.

The rod **100** includes a first threaded portion **104** threadedly received in a floating mounting block **106** (FIG. 6). The floating block **106** has a pin **108** secured thereto. The pin **108** is received in a slot **110** of an arm **112** fixed to the upper back panel **42**.

The rod **100** also includes a second threaded portion **114** which is of opposite hand from the first threaded portion **104**. The second threaded portion **114** extends through a floating mounting block **116** having a pin **118** secured thereto. The pin **118** is received in a slot **120** of an arm **122** fixed to the lower back panel **44**.

An end portion **124** of the rod **100** projects axially from the foot end **24** of the table **10**. A drive means indicated schematically at **126** is attached to the rod **100**. The drive means **126** may be a hand crank for rotating the rod **100** relative to the table **10**. The drive means **126** may also be an electric motor or fluid drive mechanism for rotating the rod **100**. The drive means **126** is selectively controllable to rotate the rod **100** to any given extent permitted by the actuating mechanism.

Upon actuation of the drive means **126**, the rod **100** rotates relative to the fixed mounting block **102** and the table **10**. Because the threaded rod portions **104** and **114** are of opposite hand, upon rotation of the rod **100** in one direction, the floating blocks **106** and **116** are moved inwardly toward each other as illustrated in FIG. 6; and upon rotation of the rod **100** in the opposite direction, the blocks **106** and **116** move axially outwardly away from each other, as illustrated in FIG. 7.

7

Upon movement of the blocks 106 and 116 toward each other as illustrated in FIG. 6, the pins 108 and 118 pull the arms 112 and 122, respectively, from the position shown in FIG. 5 to the position shown in FIG. 6. This causes the upper and lower back panels 42 and 44, respectively, to pivot and move inwardly along the track 72. This raises the center back section 46 of the table 10 upwardly away from the major side surface 20 of the table 10. The surface coil 90, as before, moves with the center back section 46 and maintains its alignment parallel to the upper major side surface 20 of the table 10.

Upon rotation of the rod 100 in the opposite direction, the blocks 106 and 116 (FIG. 7) move axially outwardly away from each other, thus causing the panels 42 and 44 to pivot to the position shown in FIG. 7. This drops the center back section 46 downwardly below the upper major side surface 20 of the table 10, taking with it the secondary coil 90.

Accordingly, it is seen from FIGS. 5-7 that an alternate mechanism for positioning the center back section 46 of the table 10, relative to the upper major side surface 20 of the table 10, is provided. With a patient lying on the upper major side surface 20 of the table 10, the patient's back may thus be imaged in any selected one of a plurality of positions between flexion and extension. Meanwhile, the secondary coil 90, if used, moves with the patient's back to maintain high resolution, while maintaining its planar orientation relative to the upper major side surface 20 of the table 10.

FIGS. 8 and 9 illustrate operation of the right knee imaging platform 56 of table 10. In this case, a knee secondary coil 130 is fixed by suitable means such as straps or VELCRO.RTM. to the central panel 68 of the imaging platform 56. The central panel 68 is pivotally mounted between the upper right knee panel 64 and the lower right knee panel 66. A threaded rod 132, having oppositely threaded portions 134 and 136, is rotatably mounted in a fixed mounting block 138 to block axial movement of the rod 132. Connection means 140 (similar to the block 106, pin 108, and arm 112) movably connects the upper right knee panel 64 to the threaded rod portion 134. Similar connection means 142 movably connects the lower right knee panel 66 to the rod threaded portion 136.

Upon rotation of the rod 132 by suitable drive means 144, in one direction, the connection means 140 and 142 cause the panels 64 and 66, respectively, to pivot and lift the central panel 68 upwardly away from the upper major side surface 20 of the table 10. The knee coil 130 moves with the panel 68 and stays in the correct planar orientation relative to the primary coil. The patient's knee (not shown) also moves upwardly away from the major side surface 20, into a different orientation than when the patient's knee is on the platform 56 when in the position shown in FIG. 8. Rotation of the rod 132 in the opposite direction causes the central panel 68 and knee coil 130 to drop below the upper major side surface 20 of the table 10, in a manner similar to that illustrated in FIG. 7 with the back section 46.

In the structure illustrated in FIGS. 10 and 11, the rod 132 and associated actuating mechanism are replaced by an inflatable bladder 150. The bladder 150 is supplied with fluid under pressure through suitable means (not shown). Upon inflation of the bladder 150 from the normal state illustrated in FIG. 10 to the extended state illustrated in FIG. 11, the panel 68 and the knee coil 130 are again raised above the upper major side surface 20 of the table 10. Upon the reduction of pressure in the bladder 150, the bladder 150 collapses axially to lower the panel 68 below the major side surface 20 of the table 10. Thus, the knee joint can also be imaged in a hyper-extended condition.

8

FIGS. 12-14 illustrate operation of the neck imaging platform 30 of the table 10. A cervical spine coil 152 is set in the recessed panel portion 34 of the table 10, below the upper major side surface 20. Attached to the head panel 32 is a pivot mechanism 154. A rod 156 has a threaded portion 158 extending through the mechanism 154. Upon rotation of the rod 156 by a suitable drive means 160, the actuating mechanism 154 causes the head panel 32 to pivot upwardly out of the plane of the major side surface 20 of the table 10, from the position shown in FIG. 12 to the position shown in FIG. 13. As shown in FIG. 14 the panel 32 can be lowered below the major side surface 20 of the table 10 by rotation of the rod 156 in the opposite direction.

With a patient's head lying on the panel 32 and the patient lying on the surface 20 of the table 10, movement of the panel 32 relative to the upper major side surface 20 of the table 10 causes flexion and extension of the patient's cervical spine. With the cervical spine coil 152 disposed about the cervical spine of the patient, the patient's cervical spine can be imaged in any selected one of a plurality of positions throughout the range of movement of the panel 32. Thus, rather than being limited to one image of the cervical spine while the patient is lying flat on an imaging table, the physician can obtain multiple images of the cervical spine at various positions throughout its range of motion. This is possible with any of the moving parts of the body which can be imaged. For example, movement can be measured and controlled in degrees—move a joint 5 degrees, image, move the joint 5 degrees further, image again, etc. Motion can also be measured in distances such as centimeters between positions.

FIGS. 15-17 illustrate an alternate mechanism for raising and lowering the panel 32. An inflatable bladder 162 is fixed between the head panel 32 and the bottom panel 84 of the table 10. When the bladder 162 is in its neutral condition, the panel 32 is flush with the upper major side surface 20 of the table 10. When the bladder 162 is inflated, the panel 32 is raised upwardly, out of the upper major side surface 20, to the elevated position illustrated in FIG. 16. When the bladder 162 is deflated, the panel 32 is lowered below the upper major side surface 20 to the depressed position illustrated in FIG. 17. Again, by controlling the pressure in the bladder 162, the MRI operator can fix the head panel at any given position within its range of motion, in order to image the cervical spine at a selected degree or flexion of extension.

FIGS. 18-20 illustrate the use of the neck imaging platform as modified for imaging an ankle of a patient (not shown). The head panel 32 is modified by the addition of a foot rest 170. The foot rest 170 is secured to the panel 32 by suitable means. The patient lies on the table 10 with, instead of his head at the head end 22, his feet toward that end. The bottom of the patient's foot is positioned against the foot rest 170, with the ankle over the recessed panel 34. An ankle imaging coil 172 is placed over the ankle. A strap 174 secures the patient's foot to the foot rest 170.

When the bladder 162 is in the neutral condition illustrated in FIG. 18, the patient's ankle is in a normal position and may be imaged. Upon further inflation of the bladder 162, the panel 32 raises upwardly away from the major side surface 20 of the table 10. The footrest 170 bends the patient's ankle and the ankle may then be imaged with the coil 172 in that bent condition. Upon the application of suction or lowering of pressure to the bladder 162 (FIG. 20), the panel 32 is pivoted down below the major side surface 20 of the table 10, thus bending the ankle in the opposite direction. The ankle may be imaged in that opposite direction with the secondary coil 172.

FIGS. 21–23 illustrate an alternate actuating mechanism for the ankle imaging platform of FIGS. 18–20. The actuating mechanism is like the actuating mechanism illustrated in FIGS. 12 and 13 for use of the panel 32 in cervical spine imaging. Upon rotation of the rod 156 in one direction or the other, the panel 32 and footrest 170 are pivoted either above the major side surface 20 of the table 10, or below the surface 20, to position the ankle for imaging at any selected position within its range of motion. It should be understood that a separate movable portion of the table 10 could be provided for use in ankle imaging, rather than using the neck imaging platform.

It should also be understood that a table in accordance with the present invention need not include every specific movable platform as shown herein. Rather, such a table may include only one movable platform, or any combination of various movable platforms. It should further be understood that suitable control means is provided for moving the several platforms, in a known manner, in order to provide repeatable movement of the various platforms through their respective ranges of motion, in order to provide repeatable imaging at known positions. The table can also be used, of course, for other types of imaging such as ultrasound or CAT scans. It should further be understood that any of the platforms may be provided as separate devices which can be placed atop a known imaging table, rather than being built into a new table as shown.

Accordingly, it is seen that the present invention provides an imaging table for positioning a body part so as to control the position or orientation of the body part. This positioning is independently controllable by the operator from a location external to the primary coil. This positioning requires no physical support effort by the patient during the time period of the imaging to maintain the selected position, as the table fully supports the weight of the body part connected therewith. Accordingly, a plurality of sequential images may be taken of a joint, for example, in differing positions, without undue effort on the part of the patient.

In another embodiment of the invention, FIGS. 24–26 illustrate an apparatus 200 for positioning a body part within a primary imaging coil. The apparatus 200 is mounted to an imaging table 202 which may be the imaging table 10 or may be a known imaging table. A known secondary imaging coil 204 is secured to the table 202 by suitable means. The coil is located in a position for imaging a particular body part. As illustrated in FIGS. 24 and 25, the coil 204 is positioned to image a shoulder of a patient who is lying on the table 202 with his head adjacent the end 206 of the table 202.

The apparatus 200 includes a support rod 210 extending longitudinally along the table 202 from a position over the table 202 (inside the primary coil) to a position off the end of the table 202 (outside the coil). The rod 210 has an inner end portion 212 to which is fixed an attachment member 214. The member 214 may be any suitable structure such as a cuff for attachment to a body part such as a forearm, for example, and may include means (such as the straps 215) for securing the cuff to the body part for movement therewith. The rod 210 also has an outward end portion 216 to which is attached a handle 218 for rotational and longitudinal movement of the rod 210 by a person other than the patient (not shown).

The rod 210 extends through and is positioned by an index mechanism 220, better seen in FIG. 26. The index mechanism 220 includes a base 222 having a first leg portion 224 and a second leg portion 226. The leg portion 224 is fixed to

the table 202. The leg portion 226 has an upper major side surface 228 to which are attached support blocks 230 and 232. The support block 230 has an opening 234 through which the rod 210 extends and is movable. The support block 232 has an opening 236, aligned with the opening 234, through which the rod 210 also extends and is movable. The blocks 230 and 232 support the rod 210, and thus the cuff 214. The block 232 also has a plurality of index openings 238. The index openings 238 are spaced regularly in a circle around the rod 210.

An index block 240 is disposed on the rod 210 outside the block 232. The rod 210 extends through an opening 246 in the index block 240. The index block 240 includes a split clamp portion 242 and a clamping bolt 244. When the split clamp 242 is loosened, the index block 240 is rotatable on and movable longitudinally on the rod 210. When the split clamp 242 is tightened, the block 240 is fixed for movement with the rod 210.

The index block 240 has an index pin opening 250 through which is extensible an index pin 252. The opening 250 is the same distance from the center of the opening 246, as the index openings 238 are from the center of the opening 236 in the block 232. Thus, the index pin opening 250 is alignable with any selected one of the index openings 238 on the support block 232. When the opening 250 is aligned with one of the index openings 238, the index pin 252 may be inserted through the index pin opening 250 and into the selected index opening 238, to block rotation of the index block 240 relative to the support block 232. If the index block 240 is clamped firmly to the rod 210, this blocks rotational movement of the rod 210 relative to the support block 232. Since the support block 232 is fixed to the table 202, this therefore blocks rotational movement of the rod 210 relative to the table 202, also thus fixing the cuff 214 in position. Further, when the index pin 252 is extended through the index pin opening 250 and into one of the index locations 238, the index assembly 220 blocks longitudinal movement of the rod 210 relative to the table 202. Thus, the cuff 214 is completely fixed in position relative to the table 202 and the coil 204.

It should be noted that other indexing mechanisms may be provided to replace the index pin opening 250 and index pin 252. For example, as shown in FIG. 26A, the index block 240 may have a spring loaded ball 254 on its radially outer surface facing the support block 232, which is selectively engageable at one of a plurality of ribbed index locations 256, thus functioning as a detent mechanism. This is suitable for a patient-directed operation. If the apparatus 200 is to be patient directed, the portion of the rod 210 extending outwardly past the index mechanism 220 may be omitted. The patient adjusts the index mechanism by moving the body part, thus moving the cuff and support rod. Other index constructions are equally feasible.

In operation of the positioner apparatus 200, the patient is first placed on the table 202 in a position as desired. The coil 204 is adjusted so as to properly image the body part in question. (It should be noted that use of a secondary coil such as the coil 204 is not essential to functioning or use of the apparatus 200.) The cuff 214 is then attached to a portion of the patient's body at a location selected to be able to move the body part to be imaged into a plurality of different positions. For example, if a shoulder joint is to be imaged, then the cuff 214 may be attached to the patient's forearm. Movement of the patient's forearm by means of the rod 210 will then cause the shoulder joint to move between a plurality of different positions. Similarly, if the patient's hip is to be imaged, the cuff 214 may be attached to the patient's

11

leg, for example, the lower leg. Movement of the cuff **214** will cause movement of the hip joint to a plurality of different positions in which it may be sequentially imaged.

The rod **210** as noted is longitudinally movable by pulling or pushing on the handle **218**. Thus, as the imaging operator moves the handle **218** longitudinally relative to the table **202**, the cuff **214** thus moves longitudinally also. The operator can therefore control the longitudinal position of the cuff **214**, and of its attached body part, from a location exterior to the primary coil.

The rod **210** is also rotatable, by means of the handle **218**. The operator rotates the handle **218** to position the cuff **214** and its attached body part in the desired orientation for imaging. This rotational position is then locked in by means of the index assembly **220**. It should be noted that any number, location, or sequence of index locations **238** may be provided. Those shown are illustrative only. In fact, an index assembly may be provided which can be locked in any rotational position within a full circle.

Many joints are movable in multiple degrees of freedom. The shoulder joint, for example, is movable in four degrees of freedom (or multiple planes of movement). In order to fully understand the joint anatomy, it is desirable to be able to image a joint in all these possible positions. Accordingly, the present invention provides for movement of a positioning apparatus such as the cuff **214** not merely rotationally and longitudinally, but also up and down and sideways.

Thus, as seen in FIGS. **24–26**, the apparatus **200** may be made movable up and down and also sideways relative to the table **202**. The index blocks **230** and **232** are movable up and down along rods **231** and **233**, respectively, which rods are fixed to the base block **222**. Thus, the support rod **210** and cuff **214** can be moved up and down to provide a third degree of movement in addition to the rotation and longitudinal movement available. Further, the index assembly **220** has a guide member **235** engaging in a slot **237**. Thus, the index assembly is movable sideways along the table **202** to carry the support rod **210** and the cuff **214** in a fourth degree of movement. With these multiple degrees of movement, in multiple planes, it is now possible to move a joint into almost any position to simulate natural joint movement, while within an imaging coil.

Another feature of the present invention is that traction can be applied to a joint being imaged, in order to distract the joint. For example, in the apparatus illustrated in FIGS. **24–26**, traction can be applied to a joint by pulling outwardly (to the right as viewed in FIG. **24**) on the rod **210**. Such force when applied to the rod **210** acts through the cuff **214** on the joint being imaged. Distracting a joint can allow a better view of the parts of the joint and thus an increased imaging benefit. This feature is not available with present imaging apparatus.

It should be noted that additional body part attachments are possible in order to better control movement and positioning. For example, extra cuffs or clamps, in addition to the one cuff shown in the drawings, may be attached to the body to more carefully and tightly control its movement and positioning. Further, it should be understood that other types of cuffs may be used, such as inflatable cuffs, etc. The cuffs should further be designed so that there is no plastic in contact with the skin. Such contact causes sweating and perspiration build up which causes imaging aberrations. Accordingly, a material is preferably provided against the skin to wick the perspiration away.

Accordingly, it is seen that the present invention provides an apparatus for longitudinally and rotationally positioning

12

a body part so as to control the position or orientation of a joint connected with the body part. This positioning is independently controllable by the operator from a location external to the primary coil. This positioning requires no physical support effort by the patient during the time period of the imaging, since the rod positioning apparatus fully supports the weight of the body part connected therewith. Nor does this adjustable positioner require any effort on the part of the patient to maintain the selected position, as the apparatus **200** performs that function also. A plurality of sequential images may be taken of a joint, for example, in differing positions, without undue effort on the part of the patient. (It should be noted that patient control of any of the positioning apparatus of the present invention is possible, as well as the described operator control.)

FIGS. **27–29** illustrate another body part positioner having two degrees of movement. An apparatus **260** includes a saddle **262** having upstanding side portions **264** and **266** joined by a bottom portion **268**. The saddle **262** is mounted on a base block **270**. The base block **270** is mounted on a panel **272** which may be the head panel **32** of the table illustrated in FIG. **1A**. The panel **272** has a plurality of index openings **274**. An index pin **276** (FIG. **27**) extends through a portion **278** of the base block **270** and is receivable in a selected one of the openings **274**. The base block **270** is pivotally mounted at **280** to the panel **272**. Thus, the base block **270**, with its attached saddle **262**, may be positioned at a selected one of a plurality of rotational positions relative to the panel **272**, as shown in phantom in FIG. **27**.

An index plate **282** is attached to the saddle **262**. The index plate **282** and saddle **262** are pivotally mounted at **284** to the base block **270**. The index plate **282** has a plurality of index openings **286** spaced in an arc about the pivot mounting **284**. A locator opening (not shown) is located behind the index plate **282**, in the base block **270**. The saddle **262**, with its attached index plate **282**, may be pivotally rotated about the mounting **284**, as shown in phantom in FIG. **28**, and secured in a position by insertion of an index pin (not shown) through the selected opening **286** into the locator opening in the base block **270**.

In operation of the positioning assembly **260**, the patient's head is secured in the saddle **262**. The saddle **262** and base block **270** are then swung around the pivot axis **280** and locked in a selected position with the index openings **274**. The saddle **262** is also rotated, with the index plate **282**, about the pivot axis **284** and locked in a selected position. The patient's head or cervical spine is then imaged. The apparatus **260** is then adjusted to a different condition, moving the patient's head or spine to a new position. The patient's head or spine is then imaged again.

Accordingly, it is seen that a patient's head or cervical spine, when the head is in the saddle **262**, can be selectively positioned in any one of a plurality of different orientations within two separate degrees of motion. Further, if the head panel **272** is pivotally mounted to the table **10**, the attached saddle **262** may also be moved up and down out of the plane of the table, thus moving the patient's head in the saddle **262** in yet a third degree of motion. The control of all these movements may be automated with a fluid drive or other means, may be made remotely controllable from a location outside the coil, or may be patient directed.

Accordingly, it is seen that the present invention also provides apparatus for positioning a body part of a patient for imaging in a plurality of different degrees of motion. For example, the patient's cervical spine may be imaged in a sequence of images by moving the saddle **262** in the desired



13

direction within the various degrees of motion and locking it in place at each selected position. There is no need for the patient to hold any selected position, as the positioning apparatus 260 does this for him. Accordingly, the imaging process is made significantly more stable and more comfortable for the patient.

Because coil support panels such as the recessed panels 34 and 48 (FIG. 1A) are located below the upper major side surface 20 of the table 10, a flat surface coil placed therein will not interfere with normal body positioning. Thus, it is seen that recessing the coils, itself, provides a significant benefit.

Several patient directed devices are illustrated in FIGS. 30-40. Such devices can be part of a new imaging table as described above, but can also be independent, that is, add-ons to an existing imaging table (as the apparatus 200 is an add-on to the table 202 in FIGS. 24-26). They are therefore less expensive and more widely usable.

These patient-directed devices can be end-mounted fixtures such as a modification of those shown in FIGS. 24-26. They can also be fixtures mounted to the upper surface of the table, in effect replacing the movable platforms of the table of FIG. 1A.

FIGS. 30-36 illustrate a patient directed wrist movement apparatus 300. The apparatus 300 includes a base 302 which may be secured to an imaging table with suitable means not shown. The base 302 supports an imaging coil 304. The patient's forearm is placed on a forearm cuff 306 secured to the base 302. The patient's hand is placed, thumb up as seen in FIG. 33, in a hand cuff 308. The hand cuff 308 is pivotally mounted at 310 to the base 302. A detent member 312 is located on a lower end portion 314 of the hand cuff 308. The detent member 312 is engageable with a ratchet-type member 316 on the base 302. Thus, the hand cuff 308 and the base 302 are releasably interlockable at a plurality of positions within their range of rotational movement.

To adjust the apparatus 300, the patient simply applies sufficient torque to release the interconnection between the hand cuff 308 and the base 302 and thereby flex or extend his wrist to the next desired position. The apparatus 300 then interlocks at this newly selected position for imaging by the coil 304.

A modification of the apparatus 300 is illustrated in FIG. 32 with an apparatus 320 having a different hand cuff 322. The hand cuff 322 is designed to have the hand lie flat rather than on edge, with the thumb to the side as viewed in FIG. 32. The wrist is again movable through its range of motion, this time being imaged in a position 90.degree. from that shown in FIG. 30.

FIGS. 37-40 illustrate a patient directed knee imaging apparatus 330. The apparatus is similar to the knee platform of the table 10 of FIG. 1A, but is instead designed to be patient directed (actuated) rather than technician or operator directed.

The apparatus 330 includes a base 332. The base 332 rests on the upper major side surface of the table. Suitable means may be provided to secure the base 332 to the table. A lower leg cuff 334 is releasably secured to the lower leg. The cuff 334 is attached at 336 to the base 332 to maintain the same focal point of imaging as the knee is flexed. An upper leg cuff 338 is releasably secured to the upper leg. The upper leg cuff is also attached at 340 to the base 332.

A knee imaging coil 342 is fixed for movement with a movable panel or portion 344 of the base 332. Attached to the patient's foot is a footrest 346 with a pawl member 348. The pawl member 348 engages a ratchet portion 350 on the base 332.

14

To adjust the device, the patient simply bends his knee to move his foot and thus the pawl member 348 along the ratchet portion 350 from one position to the next. The panel 344 moves in a manner as described above. The patient's foot is then held in that position firmly enough to allow for accurate imaging.

As the patient moves his foot and knee, the secondary coil 342 moves with the knee and generally stays in position relative to the knee. The secondary coil 342 is constrained for movement by the panel 344 so that it stays in the proper planar orientation relative to the primary imaging coil (not shown). As the patient moves his foot and knee, the cuffs 334 and 338 are constrained for proper movement by the mechanisms which attach them to the base 332 in order to maintain the same focal point of imaging as the knee is flexed. The coil 342 stays in close proximity to the knee, moves longitudinally and up and down as the knee moves, and maintains its proper planar orientation throughout its range of motion.

An optional addition to the apparatus 330 is a distraction member 352 (FIG. 40). As illustrated the distraction member 352 is an inflatable bladder having a bellows-type construction for support. The bladder 352 extends between the lower leg panel 354 and the lower leg cuff 334. When the bladder is inflated, it applies force to push the lower leg up to stress the knee outwardly. This can be done to check ligaments in the knee for damage or weakness. This feature of distraction or stressing of a joint can be applied in other joints, on other positioning apparatus, and to move body parts in a manner needed to simulate natural loading of a joint or to enable better imaging of the joint under various conditions.

Thus, it is seen that the present invention provides apparatus for use in imaging which can be attached to an existing table, allows the patient to direct the movement, controls the motion of the joint in a repeatable manner, moves a secondary coil with the joint, and holds the secondary coil in proper alignment. Of course, the movement can be operator controlled, also, by using, for example, a rod attached to the pawl device for moving the patient's foot. Similar constructions as modified can be used to provide for patient directed movement of other body parts and joints.

If a joint is small enough and/or limited enough in its range of motion that imaging the joint with one fixed secondary coil provides acceptable resolution, then it may not be necessary to move the secondary coil. For example, the wrist is a relatively small joint which, even when moved through its entire range of motion, does not take a large amount of space. Thus, the wrist is imaged using flat plates on either side of the wrist or a coil extending around the wrist.

An example of these features is the patient directed apparatus 300 (FIGS. 30-36) for moving the wrist through its range of motion within fixed coils. Similarly, the shoulder is a joint which does not move significantly through space when bent. Accordingly, it can usually be imaged successfully using a fixed coil. In this case, the present invention provides the fixture 200 (FIGS. 24-26) for moving the shoulder through its range of motion within the fixed coil.

However, some joints are large enough and/or move through space so that it is impossible to obtain optimum resolution with a fixed secondary coil through the entire range of motion of the joint. In this case, the present invention provides fixtures for moving the secondary coil with the joint. An example is the movable knee platforms of the table of FIG. 1A. Another example is the apparatus of FIGS. 37-40 for imaging the knee in a patient directed manner with a moving secondary coil.

15

It should be understood that the present invention contemplates the use of drive or actuating mechanisms other than those shown. For example, any one of the movable portions of the apparatus shown could be driven by a piston-cylinder device which is pneumatic or hydraulic. A pneumatic motor drive could be used, as well as an electric motor drive. Similarly, the pawl and ratchet or detent mechanisms illustrated in FIGS. 30–40 could be used in other configurations, as they are especially suitable for precise, repeatable incremental motion control.

In this regard, reproducibility of the movement is desirable so that the patient's progress over a period of time can be checked. Thus, indexing movement of the body part being imaged through degrees or distance is advantageous. Reproducibility, as provided by the present invention, is also useful in conducting clinical studies of groups of patients.

In accordance with the present invention, it is possible to use a larger diameter primary coil, allowing increased range of limb movement, without the degradation in image quality which would be expected from the increased coil size. This is possible because of total imaging available with the extensive use of secondary coils as described herein. For example, the knee could be flexed through its entire range of motion to allow optimum imaging of the knee joint. This is currently impossible with the known small primary coils which only allow about 50 degrees of flexion.

Thus, as illustrated in FIG. 41, a known primary MRI coil 360 with a table 362 has a height 361 from the table to the inside of the coil of 16". The table 362 has a width 363 of 19". As illustrated in FIG. 42, a replacement primary MRI coil 364 in accordance with the present invention, with a table 366, has a height 365 from the table to the inside of the coil of 21.5". The table has a width 367 of 24". With these dimensions and the moving secondary coils, substantially increased limb movement is possible, without degradation of image quality.

In a further embodiment of the present invention, an MRI primary coil is mounted to extend vertically rather than horizontally. Thus, as illustrated in FIG. 43, a primary MRI coil 368 extends vertically rather than horizontally. A patient may be placed in a standing or seated position on a support 370 for imaging in the coil 368. A ram 372 is operable to move the patient into and out of the coil 368. Positioning fixtures, etc. are mounted to a support member 374.

With the patient in a vertical or in a seated position, it is possible to simulate joint positionings and joint loadings which can not readily be simulated when the patient is lying down in a known horizontal imaging coil. For example, a weight or other tractive force can be attached to the arm to simulate shoulder joint loading experienced when carrying a heavy object. The knee can be imaged with the patient standing to see how the joint appears when loaded with body weight. The spine can be imaged when standing or seated to check for disc or vertebral problems which are experienced in normal life but which disappear when the patient lies down to be imaged in a known horizontal imaging coil. The possibilities for increased usefulness of the imaging methodology are manifold.

Any of the positioning apparatus disclosed herein are usable with or without a secondary coil. When used with a secondary coil, they provide the benefit of constraining movement of the secondary coil in a proper planar orientation relative to the primary coil, and also the benefit of keeping the secondary coil in close proximity to the body part being moved and imaged.

16

If the various apparatus of the present invention are used for magnetic resonance imaging, they must be made of non-ferromagnetic materials. Plastic is preferred for the table and the positioner, while brass is suitable for mechanical drive mechanisms. Fluid drive mechanisms are also highly suitable because they can be easily constructed using plastic components. Wood is also usable.

From the above description of the invention, those skilled in the art will perceive improvements, changes and modifications in the invention. Such improvements, changes and modifications within the skill of the art are intended to be covered by the appended claims.

What is claimed is:

1. A magnetic resonance imaging system comprising:

a stationary electromagnet;

a patient support located adjacent to the electromagnet, the support configured for maintaining a patient in a standing position; and

an actuator for raising and lowering the patient support and patient relative to a magnetic field of the electromagnet such that the patient is located within the magnetic field.

2. A magnetic resonance imaging system as defined in claim 1 further including at least one positioning fixture connected with the patient support for maintaining the patient in the standing position.

3. A magnetic resonance imaging system as defined in claim 2 further including at least one secondary coil positioned within the magnetic field of the stationary electromagnet.

4. A magnetic resonance imaging system comprising:

a stationary electromagnet having a longitudinal axis extending generally vertical;

a patient support located adjacent to the electromagnet for maintaining a patient in a seated position; and

an actuator for raising and lowering the patient support and patient relative to a magnetic field of the electromagnet such that the patient is located within the magnetic field.

5. A magnetic resonance imaging system as defined in claim 4 further including at least one positioning fixture connected with the patient support for maintaining the patient in the seated position.

6. A magnetic resonance imaging system as defined in claim 5 further including at least one secondary coil positioned within the magnetic field of the stationary electromagnet.

7. An apparatus for magnetic resonance imaging of a joint of a patient, the apparatus comprising:

a stationary electromagnet;

a patient support located adjacent the electromagnet, the support configured for maintaining a patient in a standing position, whereby a first force is applied to the joint;

at least one positioning fixture connected with the patient support for holding the joint of the patient; and

an actuator for raising and lowering the patient support and patient relative to a magnetic field of the electromagnet such that the joint of the patient is located within the magnetic field.

8. An apparatus as defined in claim 7 further including means for applying a second force to the joint.

9. An apparatus for magnetic resonance imaging of a joint of a patient, the apparatus comprising:

a stationary electromagnet having a longitudinal axis extending generally vertical;

17

a patient support locate adjacent to the electromagnet for maintaining a patient in a seated position, whereby a first force is applied to the joint;

at least one positioning fixture connected with the patient support for holding the joint of the patient; and

an actuator for raising and lowering the patient support and patient relative to a magnetic field of the electromagnet such that the joint of the patient is located within the magnetic field.

10. An apparatus as defined in claim 9 further including means for applying a second force to the joint.

11. An apparatus for magnetic resonance imaging of a spine of a patient, the apparatus comprising:

a stationary electromagnet;

a patient support located adjacent to the electromagnet, the support configured for maintaining a patient in a standing position, whereby a first force is applied to the spine;

at least one positioning fixture connected with the patient support for holding the spine of the patient; and

an actuator for raising and lowering the patient support and patient relative to a magnetic field of the electromagnet such that the spine of the patient is located within the magnetic field.

12. An apparatus as defined in claim 11 further including means for applying a second force to the spine.

13. An apparatus for magnetic resonance imaging of a spine of a patient, the apparatus comprising:

a stationary electromagnet having a longitudinal axis extending generally vertical;

a patient support located adjacent to the electromagnet for maintaining a patient in a seated position, whereby a first force is applied to the spine;

at least one positioning fixture connected with the patient support for holding the spine of the patient; and

an actuator for raising and lowering the patient support and patient relative to a magnetic field of the electro-

18

magnet such that the spine of the patient is located within the magnetic field.

14. An apparatus as defined in claim 13 further including means for applying a second force to the spine.

15. A method for magnetic resonance imaging, the method comprising the steps of:

positioning a patient against a patient support configured for maintaining the patient in a standing position;

moving the patient into a magnetic field of a stationary electromagnet; and

imaging the patient with the electromagnet.

16. A method as defined in claim 15 further including the step of using at least one positioning fixture to maintain the patient in a generally fixed position before imaging the patient with the electromagnet.

17. A method as defined in claim 15 wherein the step of imaging the patient with the electromagnet includes imaging the patient with the stationary electromagnet and a secondary coil.

18. A method for magnetic resonance imaging, the method comprising the steps of:

positioning a patient against a patient support such that the patient is maintained in a seated position;

moving the patient into a magnetic field of a stationary electromagnet having a longitudinal axis extending generally vertical; and

imaging the patient with the electromagnet.

19. A method as defined in claim 18 further including the step of using at least one positioning fixture to maintain the patient in a generally fixed position before imaging the patient with the electromagnet.

20. A method as defined in claim 18 wherein the step of imaging the patient with the electromagnet includes imaging the patient with the stationary electromagnet and a secondary coil.

\* \* \* \* \*

UNITED STATES PATENT AND TRADEMARK OFFICE  
**CERTIFICATE OF CORRECTION**

PATENT NO. : 6,882,877 B2  
DATED : April 19, 2005  
INVENTOR(S) : Peter M. Bonutti

Page 1 of 1

It is certified that error appears in the above-identified patent and that said Letters Patent is hereby corrected as shown below:

Column 16,

Line 53, the phrase "a patient support located adjacent the electromagnet," should read -- a patient support located adjacent -- to -- the electromagnet, --

Column 17,

Line 1, change "locate" to -- located --

Signed and Sealed this

Twenty-eighth Day of June, 2005

A handwritten signature in black ink on a light gray dotted background. The signature is written in a cursive style and reads "Jon W. Dudas".

JON W. DUDAS

*Director of the United States Patent and Trademark Office*